

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112125\
 Data File : VW032521.D
 Acq On : 21 Nov 2025 12:55
 Operator : SY/MD
 Sample : VW1121SBL01
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW1121SBL01

Quant Time: Nov 26 01:19:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 11 03:48:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	328508	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	615146	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	556656	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	251690	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.313	65	240876	56.874	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	113.740%
35) Dibromofluoromethane	7.898	113	201960	54.835	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	109.680%
50) Toluene-d8	10.325	98	745408	52.756	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	105.520%
62) 4-Bromofluorobenzene	12.617	95	261875	49.063	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	98.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112125\
Data File : VW032521.D
Acq On : 21 Nov 2025 12:55
Operator : SY/MD
Sample : VW1121SBL01
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW1121SBL01

Quant Time: Nov 26 01:19:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 11 03:48:21 2025
Response via : Initial Calibration

