

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW112219\
 Data File : VW014099.D
 Acq On : 22 Nov 2019 17:09
 Operator : SY/VA
 Sample : K5989-06RE
 Misc : 4.93G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0BJ7RE

Quant Time: Nov 23 00:04:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM111119S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 22 23:01:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	187144	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	83905	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	13141	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	82110	27.50	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	110.00%
7) Chloroethane-d5	2.89	69	82345	34.02	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	136.08%
10) 1,1-Dichloroethene-d2	4.02	63	121349	19.80	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	79.20%
20) 2-Butanone-d5	7.07	46	62399	71.94	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	143.88%#
24) Chloroform-d	7.65	84	206913	35.96	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	143.84%
26) 1,2-Dichloroethane-d4	8.31	65	89610	30.98	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	123.92%
29) Benzene-d6	8.27	84	390784	65.64	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	262.56%#
33) 1,2-Dichloropropane-d6	9.27	67	124881	71.61	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	286.44%#
37) Toluene-d8	10.32	98	220260	39.02	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	156.08%#
38) trans-1,3-Dichloropropene-	10.57	79	16577	22.20	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.80%
39) 2-Hexanone-d5	10.93	63	28318	85.63	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	171.26%#
48) 1,1,2,2-Tetrachloroethane-	12.69	84	59297	45.03	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	180.12%#
61) 1,2-Dichlorobenzene-d4	13.85	152	18621	35.27	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	141.08%#

Target Compounds

					Ovalue
13) Acetone	4.13	43	2786	3.156 ug/L	79
36) Methvlcvclohexane	9.33	83	2901	0.986 ug/L #	94
44) Toluene	10.39	91	3415	0.509 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW112219\
Data File : VW014099.D
Acq On : 22 Nov 2019 17:09
Operator : SY/VA
Sample : K5989-06RE
Misc : 4.93G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0BJ7RE

Quant Time: Nov 23 00:04:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM111119S.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 22 23:01:18 2019
Response via : Initial Calibration

