

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112221\
 Data File : VW020936.D
 Acq On : 22 Nov 2021 15:00
 Operator : SY/VA
 Sample : M4758-19
 Misc : 5.03g/5.0mL/MSVOA_W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP10

Quant Time: Nov 23 02:24:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 18 05:03:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.952	168	13229	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.842	114	19359	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.634	117	17413	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.554	152	9329	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.305	65	6198	50.579	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.160%
35) Dibromofluoromethane	7.885	113	5986	49.296	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.600%
50) Toluene-d8	10.323	98	21547	48.381	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.760%
62) 4-Bromofluorobenzene	12.615	95	7426	44.710	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.420%
Target Compounds						
16) Acetone	4.135	43	510	20.022	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112221\
Data File : VW020936.D
Acq On : 22 Nov 2021 15:00
Operator : SY/VA
Sample : M4758-19
Misc : 5.03g/5.0mL/MSVOA_W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP10

Quant Time: Nov 23 02:24:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111721S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 18 05:03:38 2021
Response via : Initial Calibration

