

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112221\
 Data File : VW020940.D
 Acq On : 22 Nov 2021 16:37
 Operator : SY/VA
 Sample : M4758-37
 Misc : 5.05g/5.0mL/MSVOA_W/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP19

Quant Time: Nov 23 02:25:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 18 05:03:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.945	168	88920	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.841	114	134537	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.628	117	121512	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.560	152	64653	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.305	65	43033	52.246	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	104.500%	
35) Dibromofluoromethane	7.884	113	42502	50.364	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	100.720%	
50) Toluene-d8	10.323	98	159746	51.613	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	103.220%	
62) 4-Bromofluorobenzene	12.615	95	56095	48.598	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	97.200%	
Target Compounds						
16) Acetone	4.129	43	8022	46.853	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112221\
Data File : VW020940.D
Acq On : 22 Nov 2021 16:37
Operator : SY/VA
Sample : M4758-37
Misc : 5.05g/5.0mL/MSVOA_W/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP19

Quant Time: Nov 23 02:25:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111721S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 18 05:03:38 2021
Response via : Initial Calibration

