

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112221\
 Data File : VW020943.D
 Acq On : 22 Nov 2021 17:49
 Operator : SY/VA
 Sample : M4814-01
 Misc : 5.09g/5.0mL/MSVOA_W/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 PL-1-112221

Quant Time: Nov 23 02:26:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 18 05:03:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.945	168	51654	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.841	114	81515	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.634	117	73205	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.554	152	33527	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.311	65	24194	50.565	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.140%
35) Dibromofluoromethane	7.884	113	23169	45.313	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.620%
50) Toluene-d8	10.323	98	82245	43.858	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	87.720%
62) 4-Bromofluorobenzene	12.615	95	27872	39.853	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	79.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112221\
Data File : VW020943.D
Acq On : 22 Nov 2021 17:49
Operator : SY/VA
Sample : M4814-01
Misc : 5.09g/5.0mL/MSVOA_W/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
PL-1-112221

Quant Time: Nov 23 02:26:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111721S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 18 05:03:38 2021
Response via : Initial Calibration

