

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112524\
 Data File : VW031104.D
 Acq On : 25 Nov 2024 10:50
 Operator : SY/MD
 Sample : VW1125SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK602

Quant Time: Nov 26 05:01:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM112024SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Nov 26 05:01:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	414705	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	393571	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	171158	25.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	125507	24.825	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	99.320%
7) Chloroethane-d5	2.899	69	101103	25.338	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.360%
11) 1,1-Dichloroethene-d2	4.021	65	55758	22.321	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	89.280%
21) 2-Butanone-d5	7.075	46	45062	38.712	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	77.420%
24) Chloroform-d	7.648	84	259528	23.038	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	92.160%
26) 1,2-Dichloroethane-d4	8.307	65	129431	23.515	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	94.080%
32) Benzene-d6	8.276	84	502607	22.550	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	90.200%
36) 1,2-Dichloropropane-d6	9.270	67	151569	23.063	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	92.240%
41) Toluene-d8	10.325	98	461762	22.741	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	90.960%
43) trans-1,3-Dichloroprop...	10.575	79	58233	21.284	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.120%
47) 2-Hexanone-d5	10.922	63	33874	35.863	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	71.720%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	105432	21.063	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	84.240%
66) 1,2-Dichlorobenzene-d4	13.848	152	153853	25.720	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	102.880%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112524\
Data File : VW031104.D
Acq On : 25 Nov 2024 10:50
Operator : SY/MD
Sample : VW1125SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK602

Quant Time: Nov 26 05:01:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM112024SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Nov 26 05:01:26 2024
Response via : Initial Calibration

