

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112625\
 Data File : VW032540.D
 Acq On : 26 Nov 2025 13:49
 Operator : SY/MD
 Sample : Q3719-03
 Misc : 4.45g/5mL/MSVOA_W/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-1125-11A

Quant Time: Nov 27 00:28:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 11 03:48:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	193034	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	464802	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.635	117	514695	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	246255	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	172035	69.127	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	138.260%
35) Dibromofluoromethane	7.904	113	149891	53.862	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	107.720%
50) Toluene-d8	10.325	98	515766	48.311	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	96.620%
62) 4-Bromofluorobenzene	12.617	95	227266	56.352	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	112.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112625\
Data File : VW032540.D
Acq On : 26 Nov 2025 13:49
Operator : SY/MD
Sample : Q3719-03
Misc : 4.45g/5mL/MSVOA_W/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-1125-11A

Quant Time: Nov 27 00:28:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 11 03:48:21 2025
Response via : Initial Calibration

