

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112625\
 Data File : VW032542.D
 Acq On : 26 Nov 2025 14:36
 Operator : SY/MD
 Sample : Q3604-04
 Misc : 5.00g/5mL/MSVOA_W/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 S-1

Quant Time: Nov 26 15:20:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 11 03:48:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	202457	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.849	114	487508	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	546383	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	253441	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	177410	67.969	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	135.940%
35) Dibromofluoromethane	7.898	113	155527	53.284	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	106.560%
50) Toluene-d8	10.325	98	518125	46.271	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	92.540%
62) 4-Bromofluorobenzene	12.617	95	227760	53.844	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	107.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112625\
Data File : VW032542.D
Acq On : 26 Nov 2025 14:36
Operator : SY/MD
Sample : Q3604-04
Misc : 5.00g/5mL/MSVOA_W/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
S-1

Quant Time: Nov 26 15:20:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 11 03:48:21 2025
Response via : Initial Calibration

