

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120424\
 Data File : VW031130.D
 Acq On : 04 Dec 2024 16:50
 Operator : SY/MD
 Sample : VW1204SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK400

Quant Time: Dec 05 02:40:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Dec 05 01:31:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	401699	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	375733	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.555	152	161631	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.362	65	139178	22.514	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.040%
7) Chloroethane-d5	2.899	69	115611	24.105	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	96.440%
11) 1,1-Dichloroethene-d2	4.015	65	65812	23.335	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	93.320%
21) 2-Butanone-d5	7.069	46	59421	54.963	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	109.920%
24) Chloroform-d	7.648	84	286068	24.315	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	97.240%
26) 1,2-Dichloroethane-d4	8.300	65	155151	26.799	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	107.200%
32) Benzene-d6	8.270	84	550568	23.841	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.360%
36) 1,2-Dichloropropane-d6	9.270	67	164822	24.366	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	97.480%
41) Toluene-d8	10.318	98	507753	24.364	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	97.440%
43) trans-1,3-Dichloroprop...	10.574	79	68622	25.197	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	100.800%
47) 2-Hexanone-d5	10.922	63	43003	54.260	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	108.520%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	120329	26.357	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	105.440%
66) 1,2-Dichlorobenzene-d4	13.848	152	162740	28.347	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	113.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120424\
Data File : VW031130.D
Acq On : 04 Dec 2024 16:50
Operator : SY/MD
Sample : VW1204SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK400

Quant Time: Dec 05 02:40:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120424SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Dec 05 01:31:20 2024
Response via : Initial Calibration

