

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120424\
 Data File : VW031135.D
 Acq On : 04 Dec 2024 19:21
 Operator : SY/MD
 Sample : P5066-05
 Misc : 5.21g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E28M4

Quant Time: Dec 05 02:42:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Dec 05 01:31:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	449342	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	354759	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	119125	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.362	65	138033	19.962	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.840%
7) Chloroethane-d5	2.899	69	114649	21.370	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.480%
11) 1,1-Dichloroethene-d2	4.021	65	62620	19.849	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	79.400%
21) 2-Butanone-d5	7.075	46	50324	41.613	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	83.220%
24) Chloroform-d	7.648	84	282165	21.440	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	85.760%
26) 1,2-Dichloroethane-d4	8.307	65	147144	22.722	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	90.880%
32) Benzene-d6	8.276	84	515997	23.665	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	94.640%
36) 1,2-Dichloropropane-d6	9.270	67	154187	24.141	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	96.560%
41) Toluene-d8	10.318	98	438632	22.291	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	89.160%
43) trans-1,3-Dichloroprop...	10.574	79	55343	21.523	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	86.080%
47) 2-Hexanone-d5	10.922	63	36462	48.727	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.460%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	98197	22.781	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	91.120%
66) 1,2-Dichlorobenzene-d4	13.848	152	94746	22.392	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.560%
Target Compounds						Qvalue
46) Tetrachloroethene	10.867	164	1692	0.345	ug/L	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120424\
Data File : VW031135.D
Acq On : 04 Dec 2024 19:21
Operator : SY/MD
Sample : P5066-05
Misc : 5.21g/10mL/MSVOA_W/SOIL/A
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E28M4

Quant Time: Dec 05 02:42:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120424SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Dec 05 01:31:20 2024
Response via : Initial Calibration

