

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120424\
 Data File : VW031125.D
 Acq On : 04 Dec 2024 11:03
 Operator : SY/MD
 Sample : VIBLK572
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK572

Quant Time: Dec 05 03:13:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Dec 05 01:31:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	417312	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	381884	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	170819	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	136738	21.292	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.160%
7) Chloroethane-d5	2.899	69	108866	21.850	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.400%
11) 1,1-Dichloroethene-d2	4.021	65	61226	20.896	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	83.600%
21) 2-Butanone-d5	7.075	46	52579	46.815	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	93.620%
24) Chloroform-d	7.648	84	255984	20.944	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	83.760%
26) 1,2-Dichloroethane-d4	8.307	65	137046	22.787	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.160%
32) Benzene-d6	8.276	84	516739	22.016	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	88.080%
36) 1,2-Dichloropropane-d6	9.270	67	149873	21.799	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	87.200%
41) Toluene-d8	10.319	98	466331	22.016	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	88.080%
43) trans-1,3-Dichloroprop...	10.575	79	63289	22.865	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.440%
47) 2-Hexanone-d5	10.922	63	37877	47.023	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.040%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	105956	22.835	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	91.360%
66) 1,2-Dichlorobenzene-d4	13.848	152	143574	23.663	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	94.640%
Target Compounds						
69) 1,3,5-Trichlorobenzene	14.623	180	11239	1.375	ug/L	99
70) 1,2,4-trichlorobenzene	15.129	180	11769	1.837	ug/L	97
71) Naphthalene	15.360	128	18436	1.786	ug/L #	91
72) 1,2,3-Trichlorobenzene	15.549	180	15420	2.812	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120424\
Data File : VW031125.D
Acq On : 04 Dec 2024 11:03
Operator : SY/MD
Sample : VIBLK572
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK572

Quant Time: Dec 05 03:13:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120424SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Dec 05 01:31:20 2024
Response via : Initial Calibration

