

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120624\
 Data File : VW031152.D
 Acq On : 06 Dec 2024 15:48
 Operator : SY/MD
 Sample : P5066-15
 Misc : 6.15g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E28N3

Quant Time: Dec 06 23:41:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 06 23:31:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	380727	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.629	117	317319	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	112779	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.356	65	124611	21.268	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.080%
7) Chloroethane-d5	2.899	69	93916	20.661	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.640%
11) 1,1-Dichloroethene-d2	4.015	65	60597	22.669	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	90.680%
21) 2-Butanone-d5	7.093	46	38005	37.090	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	74.180%
24) Chloroform-d	7.648	84	240722	21.587	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.360%
26) 1,2-Dichloroethane-d4	8.307	65	126493	23.053	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	92.200%
32) Benzene-d6	8.270	84	463963	23.789	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.160%
36) 1,2-Dichloropropane-d6	9.270	67	151376	26.497	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	106.000%
41) Toluene-d8	10.318	98	424497	24.118	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	96.480%
43) trans-1,3-Dichloroprop...	10.575	79	50854	22.111	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.440%
47) 2-Hexanone-d5	10.922	63	28425	42.469	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	84.940%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	83869	21.753	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	87.000%
66) 1,2-Dichlorobenzene-d4	13.848	152	95449	23.828	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	95.320%
Target Compounds					Qvalue	
14) Carbon disulfide	4.381	76	52793	2.977	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120624\
Data File : VW031152.D
Acq On : 06 Dec 2024 15:48
Operator : SY/MD
Sample : P5066-15
Misc : 6.15g/10mL/MSVOA_W/SOIL/A
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E28N3

Quant Time: Dec 06 23:41:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120424SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 06 23:31:46 2024
Response via : Initial Calibration

