

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120624\
 Data File : VW031150.D
 Acq On : 06 Dec 2024 15:00
 Operator : SY/MD
 Sample : P5066-17
 Misc : 5.61g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E28N5

Quant Time: Dec 06 23:40:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 06 23:31:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	213748	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	153265	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	36120	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	78802	23.957	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	95.840%	
7) Chloroethane-d5	2.899	69	56121	21.991	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	87.960%	
11) 1,1-Dichloroethene-d2	4.021	65	34669	23.101	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	92.400%	
21) 2-Butanone-d5	7.106	46	13770	23.937	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	47.880%	
24) Chloroform-d	7.648	84	138803	22.172	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	88.680%	
26) 1,2-Dichloroethane-d4	8.307	65	79198	25.709	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	102.840%	
32) Benzene-d6	8.270	84	237978	25.263	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	101.040%	
36) 1,2-Dichloropropane-d6	9.270	67	82153	29.773	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	119.080%	
41) Toluene-d8	10.319	98	187517	22.058	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	88.240%	
43) trans-1,3-Dichloroprop...	10.575	79	20426	18.387	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	73.560%	
47) 2-Hexanone-d5	10.922	63	9493	29.365	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	58.720%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	52325	28.098	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	112.400%	
66) 1,2-Dichlorobenzene-d4	13.848	152	26869	20.943	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	83.760%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120624\
Data File : VW031150.D
Acq On : 06 Dec 2024 15:00
Operator : SY/MD
Sample : P5066-17
Misc : 5.61g/10mL/MSVOA_W/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E28N5

Quant Time: Dec 06 23:40:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120424SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 06 23:31:46 2024
Response via : Initial Calibration

