

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120821\
 Data File : VW021221.D
 Acq On : 08 Dec 2021 11:42
 Operator : SY/VA
 Sample : VIBLK513
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK513

Quant Time: Dec 08 23:00:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Dec 08 22:58:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	98117	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	93745	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	49246	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	31725	23.207	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.840%
7) Chloroethane-d5	2.885	69	23782	25.988	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	103.960%
11) 1,1-Dichloroethene-d2	4.019	63	35390	16.205	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	64.840%
21) 2-Butanone-d5	7.086	46	10779	34.569	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	69.140%
24) Chloroform-d	7.647	84	52072	21.192	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	84.760%
26) 1,2-Dichloroethane-d4	8.305	65	28222	20.994	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	83.960%
32) Benzene-d6	8.275	84	103696	21.634	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	86.520%
36) 1,2-Dichloropropane-d6	9.274	67	28472	21.375	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	85.480%
41) Toluene-d8	10.323	98	101093	20.569	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	82.280%
43) trans-1,3-Dichloroprop...	10.579	79	10533	17.414	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	69.640%
47) 2-Hexanone-d5	10.926	63	9032	35.440	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	70.880%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	22574	18.318	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	73.280%
66) 1,2-Dichlorobenzene-d4	13.853	152	40053	22.700	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	90.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120821\
Data File : VW021221.D
Acq On : 08 Dec 2021 11:42
Operator : SY/VA
Sample : VIBLK513
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK513

Quant Time: Dec 08 23:00:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Dec 08 22:58:53 2021
Response via : Initial Calibration

