

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120821\
 Data File : VW021216.D
 Acq On : 08 Dec 2021 09:37
 Operator : SY/VA
 Sample : VW1208SBL01
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK418

Quant Time: Dec 08 22:59:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Dec 08 22:58:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	111582	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	102470	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	54850	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	36417	23.425	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.680%
7) Chloroethane-d5	2.885	69	29135	27.995	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	112.000%
11) 1,1-Dichloroethene-d2	4.025	63	41730	16.802	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	67.200%
21) 2-Butanone-d5	7.080	46	13240	37.337	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	74.680%
24) Chloroform-d	7.653	84	60586	21.681	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.720%
26) 1,2-Dichloroethane-d4	8.311	65	33491	21.907	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.640%
32) Benzene-d6	8.281	84	121772	23.242	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	92.960%
36) 1,2-Dichloropropane-d6	9.275	67	32796	22.525	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	90.080%
41) Toluene-d8	10.323	98	119399	22.225	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	88.880%
43) trans-1,3-Dichloroprop...	10.579	79	12877	19.477	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	77.920%
47) 2-Hexanone-d5	10.927	63	10520	37.764	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	75.520%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	27091	20.112	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	80.440%
66) 1,2-Dichlorobenzene-d4	13.853	152	46019	23.416	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	93.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120821\
Data File : VW021216.D
Acq On : 08 Dec 2021 09:37
Operator : SY/VA
Sample : VW1208SBL01
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK418

Quant Time: Dec 08 22:59:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Dec 08 22:58:53 2021
Response via : Initial Calibration

