

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120821\
 Data File : VW021226.D
 Acq On : 08 Dec 2021 13:42
 Operator : SY/VA
 Sample : VIBLK515
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK515

Quant Time: Dec 08 23:01:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Dec 08 22:58:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	93848	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	95482	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	55016	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	39349	30.094	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 120.360%	
7) Chloroethane-d5	2.885	69	27001	30.848	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 123.400%	
11) 1,1-Dichloroethene-d2	4.019	63	43769	20.954	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 83.800%	
21) 2-Butanone-d5	7.074	46	15680	52.574	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 105.140%	
24) Chloroform-d	7.653	84	65786	27.991	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 111.960%	
26) 1,2-Dichloroethane-d4	8.305	65	37171	28.909	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 115.640%	
32) Benzene-d6	8.275	84	132108	27.060	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 108.240%	
36) 1,2-Dichloropropane-d6	9.274	67	36691	27.044	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 108.160%	
41) Toluene-d8	10.323	98	131150	26.199	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 104.800%	
43) trans-1,3-Dichloroprop...	10.579	79	14697	23.857	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 95.440%	
47) 2-Hexanone-d5	10.927	63	12626	48.641	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 97.280%	
56) 1,1,2,2-Tetrachloroeth...	12.688	84	30607	24.385	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 97.560%	
66) 1,2-Dichlorobenzene-d4	13.853	152	53104	26.940	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 107.760%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120821\
Data File : VW021226.D
Acq On : 08 Dec 2021 13:42
Operator : SY/VA
Sample : VIBLK515
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK515

Quant Time: Dec 08 23:01:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Dec 08 22:58:53 2021
Response via : Initial Calibration

