

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120821\
 Data File : VW021232.D
 Acq On : 08 Dec 2021 16:07
 Operator : SY/VA
 Sample : VIBLK518
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK518

Quant Time: Dec 08 23:03:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Dec 08 22:58:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	110494	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	104575	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	53577	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	35679	23.176	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.720%
7) Chloroethane-d5	2.879	69	26104	25.330	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.320%
11) 1,1-Dichloroethene-d2	4.013	63	37138	15.101	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.400%
21) 2-Butanone-d5	7.098	46	11065	31.511	ug/L	0.02
Spiked Amount	50.000	Range	20 - 135	Recovery	=	63.020%
24) Chloroform-d	7.647	84	59120	21.365	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	85.440%
26) 1,2-Dichloroethane-d4	8.305	65	32134	21.227	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	84.920%
32) Benzene-d6	8.275	84	114537	21.421	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	85.680%
36) 1,2-Dichloropropane-d6	9.274	67	31592	21.261	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	85.040%
41) Toluene-d8	10.323	98	113641	20.727	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	82.920%
43) trans-1,3-Dichloroprop...	10.579	79	12072	17.892	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	71.560%
47) 2-Hexanone-d5	10.927	63	9122	32.086	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	64.180%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	24368	17.726	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	70.920%
66) 1,2-Dichlorobenzene-d4	13.853	152	42608	22.196	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	88.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120821\
Data File : VW021232.D
Acq On : 08 Dec 2021 16:07
Operator : SY/VA
Sample : VIBLK518
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK518

Quant Time: Dec 08 23:03:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Dec 08 22:58:53 2021
Response via : Initial Calibration

