

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW121018\
 Data File : VW007358.D
 Acq On : 10 Dec 2018 15:27
 Operator : SY/AP
 Sample : J6297-01
 Misc : 6.38G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 F9E36

Quant Time: Dec 11 02:26:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM120718S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 11 01:18:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	120578	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	97098	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	44180	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	20417	26.00	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	104.00%
7) Chloroethane-d5	2.90	69	15438	23.21	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.84%
10) 1,1-Dichloroethene-d2	4.01	63	50099	18.49	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	73.96%
20) 2-Butanone-d5	7.07	46	23288	47.04	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.08%
24) Chloroform-d	7.65	84	68842	24.21	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	96.84%
26) 1,2-Dichloroethane-d4	8.31	65	42446	26.22	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	104.88%
29) Benzene-d6	8.27	84	131768	24.99	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	99.96%
33) 1,2-Dichloropropane-d6	9.27	67	38137	25.34	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.36%
37) Toluene-d8	10.32	98	125864	25.46	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	101.84%
38) trans-1,3-Dichloropropene-	10.58	79	19522	25.37	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	101.48%
39) 2-Hexanone-d5	10.93	63	18550	51.21	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	102.42%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	33612	24.51	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	98.04%
61) 1,2-Dichlorobenzene-d4	13.86	152	39261	25.26	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	101.04%

Target Compounds					Ovalue
13) Acetone	4.12	43	3352	5.823 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW121018\
Data File : VW007358.D
Acq On : 10 Dec 2018 15:27
Operator : SY/AP
Sample : J6297-01
Misc : 6.38G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
F9E36

Quant Time: Dec 11 02:26:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM120718S.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 11 01:18:15 2018
Response via : Initial Calibration

