

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121024\
 Data File : VW031196.D
 Acq On : 10 Dec 2024 17:06
 Operator : SY/MD
 Sample : P5155-04
 Misc : 5.28g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Dec 11 00:32:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Dec 11 00:26:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	575944	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.629	117	454023	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	123881	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	99088	11.180	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	44.720%
7) Chloroethane-d5	2.893	69	88273	12.837	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	51.360%
11) 1,1-Dichloroethene-d2	4.015	65	43598	10.782	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	43.120%#
21) 2-Butanone-d5	7.088	46	54873	35.400	ug/L	-0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	70.800%
24) Chloroform-d	7.648	84	256902	15.230	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	60.920%
26) 1,2-Dichloroethane-d4	8.374	65	324	0.039	ug/L	0.07
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.160%#
32) Benzene-d6	8.270	84	410436	14.708	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	58.840%
36) 1,2-Dichloropropane-d6	9.270	67	153585	18.789	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	75.160%
41) Toluene-d8	10.319	98	335477	13.321	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	53.280%
43) trans-1,3-Dichloroprop...	10.575	79	51649	15.695	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	62.760%
47) 2-Hexanone-d5	10.922	63	40809	42.613	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.220%
56) 1,1,2,2-Tetrachloroeth...	12.684	84	110569	20.043	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	80.160%
66) 1,2-Dichlorobenzene-d4	13.849	152	68677	15.608	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	62.440%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121024\
Data File : VW031196.D
Acq On : 10 Dec 2024 17:06
Operator : SY/MD
Sample : P5155-04
Misc : 5.28g/10mL/MSVOA_W/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Dec 11 00:32:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120424SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Dec 11 00:26:57 2024
Response via : Initial Calibration

