

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121024\
 Data File : VW031199.D
 Acq On : 10 Dec 2024 18:18
 Operator : SY/MD
 Sample : P5155-01
 Misc : 6.90g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E28Q6

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 12/11/2024
 Supervised By :Semsettin Yesilyurt 12/11/2024

Quant Time: Dec 11 00:33:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Dec 11 00:26:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	322406	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.629	117	225559	25.000	ug/L	# 0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	49786	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	90670	18.275	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	73.080%		
7) Chloroethane-d5	2.899	69	79547	20.665	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	82.680%		
11) 1,1-Dichloroethene-d2	4.015	65	37793	16.696	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	66.800%		
21) 2-Butanone-d5	7.094	46	37667m	43.410	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	86.820%		
24) Chloroform-d	7.648	84	210907	22.335	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	89.360%		
26) 1,2-Dichloroethane-d4	8.301	65	128490	27.653	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	110.600%		
32) Benzene-d6	8.270	84	326991	23.587	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	94.360%		
36) 1,2-Dichloropropane-d6	9.270	67	127722	31.452	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	125.800%#		
41) Toluene-d8	10.319	98	232796	18.607	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	74.440%		
43) trans-1,3-Dichloroprop...	10.575	79	33525	20.506	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	82.040%		
47) 2-Hexanone-d5	10.922	63	23423	49.232	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	98.460%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	86408	31.529	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	126.120%#		
66) 1,2-Dichlorobenzene-d4	13.848	152	36561	20.675	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	82.720%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121024\
Data File : VW031199.D
Acq On : 10 Dec 2024 18:18
Operator : SY/MD
Sample : P5155-01
Misc : 6.90g/10mL/MSVOA_W/SOIL/A
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E28Q6

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 12/11/2024
Supervised By :Semsettin Yesilyurt 12/11/2024

Quant Time: Dec 11 00:33:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120424SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Dec 11 00:26:57 2024
Response via : Initial Calibration

