

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121124\
 Data File : VW031210.D
 Acq On : 11 Dec 2024 13:41
 Operator : SY/MD
 Sample : P5155-04
 Misc : 5.08g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Dec 12 03:03:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Dec 12 03:01:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	299823	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.629	117	292116	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	120643	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	69468	15.056	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	60.240%
7) Chloroethane-d5	2.899	69	61041	17.052	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	68.200%
11) 1,1-Dichloroethene-d2	4.015	65	31465	14.947	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	59.800%
21) 2-Butanone-d5	7.093	46	41885	51.907	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	103.820%
24) Chloroform-d	7.648	84	171479	19.527	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	78.120%
26) 1,2-Dichloroethane-d4	8.307	65	100586	23.278	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	93.120%
32) Benzene-d6	8.276	84	274337	15.280	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	61.120%
36) 1,2-Dichloropropane-d6	9.270	67	101481	19.296	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	77.200%
41) Toluene-d8	10.318	98	228888	14.127	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	56.520%
43) trans-1,3-Dichloroprop...	10.575	79	38049	17.970	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	71.880%
47) 2-Hexanone-d5	10.922	63	27927	45.324	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	90.640%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	87648	24.694	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	98.760%
66) 1,2-Dichlorobenzene-d4	13.848	152	83817	19.560	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	78.240%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121124\
Data File : VW031210.D
Acq On : 11 Dec 2024 13:41
Operator : SY/MD
Sample : P5155-04
Misc : 5.08g/10mL/MSVOA_W/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Dec 12 03:03:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120424SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Dec 12 03:01:22 2024
Response via : Initial Calibration

