

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121124\
 Data File : VW031207.D
 Acq On : 11 Dec 2024 12:30
 Operator : SY/MD
 Sample : P5155-01
 Misc : 8.34g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E28Q6

Manual Integrations
 APPROVED

Reviewed By :Romaben Patel 12/12/2024
 Supervised By :Mahesh Dadoda 12/13/2024

Quant Time: Dec 12 03:02:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Dec 12 03:01:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	303013m	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	280020	25.000	ug/L	# 0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	117104	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	96070	20.602	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	82.400%		
7) Chloroethane-d5	2.899	69	85264	23.568	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	94.280%		
11) 1,1-Dichloroethene-d2	4.015	65	44794	21.055	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	84.200%		
21) 2-Butanone-d5	7.094	46	42920	52.629	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	105.260%		
24) Chloroform-d	7.642	84	229260	25.833	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	103.320%		
26) 1,2-Dichloroethane-d4	8.301	65	127342m	29.160	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	116.640%		
32) Benzene-d6	8.270	84	378483	21.991	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	87.960%		
36) 1,2-Dichloropropane-d6	9.270	67	140105	27.791	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	111.160%		
41) Toluene-d8	10.319	98	324264	20.877	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	83.520%		
43) trans-1,3-Dichloroprop...	10.575	79	51717	25.481	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	101.920%		
47) 2-Hexanone-d5	10.922	63	30094	50.951	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	101.900%		
56) 1,1,2,2-Tetrachloroeth...	12.684	84	102154	30.025	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	120.080%#		
66) 1,2-Dichlorobenzene-d4	13.842	152	109105	26.231	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	104.920%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121124\
Data File : VW031207.D
Acq On : 11 Dec 2024 12:30
Operator : SY/MD
Sample : P5155-01
Misc : 8.34g/10mL/MSVOA_W/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E28Q6

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 12/12/2024
Supervised By :Mahesh Dadoda 12/13/2024

Quant Time: Dec 12 03:02:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120424SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Dec 12 03:01:22 2024
Response via : Initial Calibration

