

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121124\
 Data File : VW031212.D
 Acq On : 11 Dec 2024 14:29
 Operator : SY/MD
 Sample : P5155-12
 Misc : 4.50g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E28T7

Quant Time: Dec 12 03:04:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Dec 12 03:01:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	393743	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.629	117	270492	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.550	152	56433	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	87949	14.515	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	58.040%
7) Chloroethane-d5	2.893	69	76661	16.307	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	65.240%
11) 1,1-Dichloroethene-d2	4.021	65	36723	13.284	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	53.120%
21) 2-Butanone-d5	7.087	46	41165	38.846	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	77.700%
24) Chloroform-d	7.648	84	195055	16.914	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	67.640%
26) 1,2-Dichloroethane-d4	8.307	65	114853	20.240	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	80.960%
32) Benzene-d6	8.276	84	307848	18.517	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	74.080%
36) 1,2-Dichloropropane-d6	9.270	67	115375	23.692	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.760%
41) Toluene-d8	10.319	98	222934	14.859	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	59.440%
43) trans-1,3-Dichloroprop...	10.575	79	35942	18.332	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	73.320%
47) 2-Hexanone-d5	10.922	63	27520	48.234	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	96.460%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	84315	25.654	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	102.600%
66) 1,2-Dichlorobenzene-d4	13.848	152	39063	19.488	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	77.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121124\
Data File : VW031212.D
Acq On : 11 Dec 2024 14:29
Operator : SY/MD
Sample : P5155-12
Misc : 4.50g/10mL/MSVOA_W/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E28T7

Quant Time: Dec 12 03:04:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120424SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Dec 12 03:01:22 2024
Response via : Initial Calibration

