

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121225\
 Data File : VW032641.D
 Acq On : 12 Dec 2025 08:26
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 12 23:37:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W121025S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 11 01:56:28 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.310	0.263	15.2	99	0.00
3 P	Chloromethane	0.548	0.435	20.6	98	0.00
4 C	Vinyl Chloride	0.654	0.601	8.1#	106	0.00
5 T	Bromomethane	0.435	0.397	8.7	105	0.00
6 T	Chloroethane	0.422	0.398	5.7	107	0.00
7 T	Trichlorofluoromethane	0.572	0.553	3.3	107	0.00
8 T	Diethyl Ether	0.383	0.352	8.1	107	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.573	0.534	6.8	107	0.00
10 T	Methyl Iodide	0.798	0.755	5.4	105	0.00
11 T	Tert butyl alcohol	0.052	0.045	13.5	112	0.00
12 CM	1,1-Dichloroethene	0.607	0.566	6.8#	106	0.00
13 T	Acrolein	0.087	0.080	8.0	88	0.00
14 T	Allyl chloride	1.128	1.044	7.4	103	0.00
15 T	Acrylonitrile	0.214	0.205	4.2	114	0.00
16 T	Acetone	0.262	0.224	14.5	114	0.00
17 T	Carbon Disulfide	1.764	1.632	7.5	105	0.00
18 T	Methyl Acetate	0.511	0.495	3.1	114	0.00
19 T	Methyl tert-butyl Ether	1.205	1.174	2.6	107	0.00
20 T	Methylene Chloride	0.800	0.678	15.3	105	0.01
21 T	trans-1,2-Dichloroethene	0.663	0.629	5.1	108	0.00
22 T	Diisopropyl ether	2.285	2.213	3.2	105	0.00
23 T	Vinyl Acetate	1.545	1.497	3.1	108	0.00
24 P	1,1-Dichloroethane	1.307	1.219	6.7	105	0.00
25 T	2-Butanone	0.308	0.286	7.1	114	0.00
26 T	2,2-Dichloropropane	0.734	0.695	5.3	104	0.00
27 T	cis-1,2-Dichloroethene	0.774	0.734	5.2	104	0.00
28 T	Bromochloromethane	0.519	0.553	-6.6	106	0.00
29 T	Tetrahydrofuran	0.185	0.183	1.1	118	0.00
30 C	Chloroform	1.289	1.223	5.1#	106	0.00
31 T	Cyclohexane	1.204	1.094	9.1	105	0.00
32 T	1,1,1-Trichloroethane	0.948	0.910	4.0	106	0.00
33 S	1,2-Dichloroethane-d4	0.749	0.680	9.2	111	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.338	0.319	5.6	111	0.00
36 T	1,1-Dichloropropene	0.515	0.515	0.0	106	0.00
37 T	Ethyl Acetate	0.353	0.370	-4.8	116	0.00
38 T	Carbon Tetrachloride	0.476	0.472	0.8	105	0.00
39 T	Methylcyclohexane	0.647	0.638	1.4	101	0.00
40 TM	Benzene	1.562	1.531	2.0	104	0.00
41 T	Methacrylonitrile	0.188	0.192	-2.1	109	0.00
42 TM	1,2-Dichloroethane	0.497	0.499	-0.4	108	0.00
43 T	Isopropyl Acetate	0.625	0.626	-0.2	111	0.00
44 TM	Trichloroethene	0.357	0.353	1.1	106	0.00
45 C	1,2-Dichloropropane	0.395	0.400	-1.3#	107	0.00
46 T	Dibromomethane	0.232	0.235	-1.3	110	0.00
47 T	Bromodichloromethane	0.539	0.529	1.9	104	0.00
48 T	Methyl methacrylate	0.295	0.318	-7.8	121	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121225\
 Data File : VW032641.D
 Acq On : 12 Dec 2025 08:26
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 12 23:37:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W121025S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 11 01:56:28 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	119	0.00
50 S	Toluene-d8	1.283	1.212	5.5	107	0.00
51 T	4-Methyl-2-Pentanone	0.349	0.368	-5.4	119	0.00
52 CM	Toluene	0.961	0.983	-2.3#	107	0.00
53 T	t-1,3-Dichloropropene	0.531	0.541	-1.9	106	0.00
54 T	cis-1,3-Dichloropropene	0.612	0.619	-1.1	106	0.00
55 T	1,1,2-Trichloroethane	0.311	0.315	-1.3	111	0.00
56 T	Ethyl methacrylate	0.448	0.464	-3.6	110	0.00
57 T	1,3-Dichloropropane	0.558	0.570	-2.2	109	0.00
58 T	2-Chloroethyl Vinyl ether	0.209	0.199	4.8	142	0.00
59 T	2-Hexanone	0.253	0.264	-4.3	118	0.00
60 T	Dibromochloromethane	0.346	0.361	-4.3	109	0.00
61 T	1,2-Dibromoethane	0.300	0.302	-0.7	106	0.00
62 S	4-Bromofluorobenzene	0.471	0.453	3.8	112	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.321	0.303	5.6	108	0.00
65 PM	Chlorobenzene	1.142	1.152	-0.9	110	0.00
66 T	1,1,1,2-Tetrachloroethane	0.357	0.353	1.1	103	0.00
67 C	Ethyl Benzene	1.973	2.031	-2.9#	107	0.00
68 T	m/p-Xylenes	0.757	0.755	0.3	104	0.00
69 T	o-Xylene	0.698	0.711	-1.9	105	0.00
70 T	Styrene	1.218	1.231	-1.1	105	0.00
71 P	Bromoform	0.207	0.212	-2.4	115	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
73 T	Isopropylbenzene	3.803	3.831	-0.7	101	0.00
74 T	N-amyl acetate	1.303	1.337	-2.6	110	0.00
75 P	1,1,2,2-Tetrachloroethane	0.916	0.956	-4.4	115	0.00
76 T	1,2,3-Trichloropropane	0.671	0.695	-3.6	114	0.00
77 T	Bromobenzene	0.873	0.897	-2.7	112	0.00
78 T	n-propylbenzene	4.858	5.043	-3.8	106	0.00
79 T	2-Chlorotoluene	2.851	2.964	-4.0	107	0.00
80 T	1,3,5-Trimethylbenzene	3.194	3.333	-4.4	107	0.00
81 T	trans-1,4-Dichloro-2-butene	0.301	0.316	-5.0	114	0.00
82 T	4-Chlorotoluene	3.028	3.119	-3.0	109	0.00
83 T	tert-Butylbenzene	2.721	2.788	-2.5	104	0.00
84 T	1,2,4-Trimethylbenzene	3.222	3.376	-4.8	106	0.00
85 T	sec-Butylbenzene	4.166	4.254	-2.1	105	0.00
86 T	p-Isopropyltoluene	3.389	3.429	-1.2	104	0.00
87 T	1,3-Dichlorobenzene	1.786	1.853	-3.8	109	0.00
88 T	1,4-Dichlorobenzene	1.779	1.788	-0.5	106	0.00
89 T	n-Butylbenzene	3.405	3.509	-3.1	107	0.00
90 T	Hexachloroethane	0.633	0.625	1.3	102	0.00
91 T	1,2-Dichlorobenzene	1.612	1.667	-3.4	110	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.163	0.159	2.5	114	0.00
93 T	1,2,4-Trichlorobenzene	0.958	1.007	-5.1	111	0.00
94 T	Hexachlorobutadiene	0.453	0.451	0.4	103	0.00
95 T	Naphthalene	2.305	2.422	-5.1	113	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121225\
Data File : VW032641.D
Acq On : 12 Dec 2025 08:26
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Dec 12 23:37:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W121025S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 11 01:56:28 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.905	0.904	0.1	105	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6