

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW121418\
 Data File : VW007441.D
 Acq On : 12 Dec 2018 14:34
 Operator : SY/AP
 Sample : J6238-08
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK01

Quant Time: Dec 13 01:11:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM120718S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 13 00:38:14 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	108892	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	102112	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	49357	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	19430	27.39	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	109.56%
7) Chloroethane-d5	2.89	69	14420	24.01	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	96.04%
10) 1,1-Dichloroethene-d2	4.01	63	50253	20.54	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	82.16%
20) 2-Butanone-d5	7.07	46	25261	56.50	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	113.00%
24) Chloroform-d	7.65	84	67737	26.37	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	105.48%
26) 1,2-Dichloroethane-d4	8.31	65	42041	28.76	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	115.04%
29) Benzene-d6	8.27	84	132873	23.96	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.84%
33) 1,2-Dichloropropane-d6	9.27	67	37558	23.73	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.92%
37) Toluene-d8	10.32	98	129725	24.95	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	99.80%
38) trans-1,3-Dichloropropene-	10.58	79	19930	24.63	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	98.52%
39) 2-Hexanone-d5	10.93	63	20856	54.74	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	109.48%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	40352	27.98	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	111.92%
61) 1,2-Dichlorobenzene-d4	13.86	152	50098	28.85	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	115.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW121418\
Data File : VW007441.D
Acq On : 12 Dec 2018 14:34
Operator : SY/AP
Sample : J6238-08
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK01

Quant Time: Dec 13 01:11:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM120718S.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 13 00:38:14 2018
Response via : Initial Calibration

