

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121523\
 Data File : VW027644.D
 Acq On : 15 Dec 2023 19:10
 Operator : SY/MD
 Sample : 05834-01
 Misc : 4.25g/10mL/MSVOA_W/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGQZ3

Quant Time: Dec 16 01:39:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121523SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Dec 16 00:05:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	610893	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	545132	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	233941	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	219864	25.413	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.640%
7) Chloroethane-d5	2.899	69	178233	26.459	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	105.840%
11) 1,1-Dichloroethene-d2	4.021	65	85665	24.723	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	98.880%
21) 2-Butanone-d5	7.081	46	55541	37.700	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	75.400%
24) Chloroform-d	7.648	84	378800	26.059	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	104.240%
26) 1,2-Dichloroethane-d4	8.307	65	177822	26.053	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	104.200%
32) Benzene-d6	8.270	84	827659	25.978	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	103.920%
36) 1,2-Dichloropropane-d6	9.270	67	240290	26.622	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	106.480%
41) Toluene-d8	10.319	98	730409	24.710	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	98.840%
43) trans-1,3-Dichloroprop...	10.575	79	77259	22.108	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.440%
47) 2-Hexanone-d5	10.922	63	49712	37.794	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	75.580%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	157026	23.005	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	92.000%
66) 1,2-Dichlorobenzene-d4	13.849	152	204203	24.385	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	97.560%
Target Compounds						Qvalue
46) Tetrachloroethene	10.861	164	3431	0.499	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121523\
Data File : VW027644.D
Acq On : 15 Dec 2023 19:10
Operator : SY/MD
Sample : 05834-01
Misc : 4.25g/10mL/MSVOA_W/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGQZ3

Quant Time: Dec 16 01:39:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121523SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Dec 16 00:05:51 2023
Response via : Initial Calibration

