

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121523\
 Data File : VW027697.D
 Acq On : 16 Dec 2023 17:22
 Operator : SY/MD
 Sample : 05788-24
 Misc : 5.10g/10mL/MSVOA_W/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DCLS6

Quant Time: Dec 17 00:20:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121523SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sun Dec 17 00:15:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	608010	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	546279	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	257536	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	159575	18.532	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	74.120%
7) Chloroethane-d5	2.893	69	129049	19.248	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.000%
11) 1,1-Dichloroethene-d2	4.015	65	64028	18.566	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	74.280%
21) 2-Butanone-d5	7.087	46	51778	35.312	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	70.620%
24) Chloroform-d	7.648	84	279331	19.307	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	77.240%
26) 1,2-Dichloroethane-d4	8.301	65	131565	19.368	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	77.480%
32) Benzene-d6	8.270	84	615925	19.292	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	77.160%
36) 1,2-Dichloropropane-d6	9.270	67	175126	19.362	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	77.440%
41) Toluene-d8	10.319	98	560759	18.931	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	75.720%
43) trans-1,3-Dichloroprop...	10.575	79	64410	18.393	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	73.560%
47) 2-Hexanone-d5	10.922	63	50670	38.441	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	76.880%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	123302	18.026	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	72.120%
66) 1,2-Dichlorobenzene-d4	13.848	152	175918	19.083	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	76.320%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121523\
Data File : VW027697.D
Acq On : 16 Dec 2023 17:22
Operator : SY/MD
Sample : 05788-24
Misc : 5.10g/10mL/MSVOA_W/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DCLS6

Quant Time: Dec 17 00:20:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121523SMA.M
Quant Title : SFAM01.0
QLast Update : Sun Dec 17 00:15:03 2023
Response via : Initial Calibration

