

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121624\
 Data File : VW031264.D
 Acq On : 16 Dec 2024 12:11
 Operator : SY/MD
 Sample : P5232-16
 Misc : 5.10g/10mL/MSVOA_W/SOILB
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E28T0

Quant Time: Dec 17 00:35:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Dec 17 00:33:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	950596	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	593989	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.550	152	166176	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	289324	20.228	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.920%
7) Chloroethane-d5	2.899	69	215103	20.331	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.320%
11) 1,1-Dichloroethene-d2	4.021	65	128749	18.977	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	75.920%
21) 2-Butanone-d5	7.094	46	58359	21.666	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	43.340%
24) Chloroform-d	7.648	84	483408	18.560	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	74.240%
26) 1,2-Dichloroethane-d4	8.307	65	268196	20.504	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	82.000%
32) Benzene-d6	8.270	84	933842	24.529	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	98.120%
36) 1,2-Dichloropropane-d6	9.270	67	280422	25.273	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.080%
41) Toluene-d8	10.319	98	717523	20.259	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	81.040%
43) trans-1,3-Dichloroprop...	10.575	79	88166	17.233	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	68.920%
47) 2-Hexanone-d5	10.922	63	51478	34.499	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	69.000%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	149432	21.450	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	85.800%
66) 1,2-Dichlorobenzene-d4	13.849	152	104896	17.003	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	68.000%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121624\
Data File : VW031264.D
Acq On : 16 Dec 2024 12:11
Operator : SY/MD
Sample : P5232-16
Misc : 5.10g/10mL/MSVOA_W/SOILB
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E28T0

Quant Time: Dec 17 00:35:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Dec 17 00:33:32 2024
Response via : Initial Calibration

