

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121624\
 Data File : VW031272.D
 Acq On : 16 Dec 2024 15:23
 Operator : SY/MD
 Sample : P5232-17RE
 Misc : 4.32g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E28T1RE

Quant Time: Dec 17 00:37:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Dec 17 00:33:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	775027	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	408742	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.555	152	76651	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.356	65	247755	21.246	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.000%
7) Chloroethane-d5	2.893	69	208676	24.191	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	96.760%
11) 1,1-Dichloroethene-d2	4.014	65	119808	21.659	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	86.640%
21) 2-Butanone-d5	7.099	46	38862	17.696	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	35.400%
24) Chloroform-d	7.648	84	496206	23.367	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.480%
26) 1,2-Dichloroethane-d4	8.300	65	260277	24.406	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	97.640%
32) Benzene-d6	8.270	84	924667	35.295	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	141.200%#
36) 1,2-Dichloropropane-d6	9.270	67	294500	38.571	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	154.280%#
41) Toluene-d8	10.318	98	639706	26.248	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	105.000%
43) trans-1,3-Dichloroprop...	10.574	79	73758	20.950	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	83.800%
47) 2-Hexanone-d5	10.922	63	31417	30.597	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	61.200%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	159067	33.181	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	132.720%#
66) 1,2-Dichlorobenzene-d4	13.848	152	63663	22.372	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.480%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121624\
Data File : VW031272.D
Acq On : 16 Dec 2024 15:23
Operator : SY/MD
Sample : P5232-17RE
Misc : 4.32g/10mL/MSVOA_W/SOIL/B
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E28T1RE

Quant Time: Dec 17 00:37:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Dec 17 00:33:32 2024
Response via : Initial Calibration

