

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122023\
 Data File : VW027807.D
 Acq On : 20 Dec 2023 13:56
 Operator : SY/MD
 Sample : 05869-05
 Misc : 4.29g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0B36

Quant Time: Dec 20 23:16:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121523SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sun Dec 17 01:25:09 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	622321	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.623	117	561269	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	277208	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	169991	19.288	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.160%
7) Chloroethane-d5	2.899	69	133625	19.473	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.880%
11) 1,1-Dichloroethene-d2	4.021	65	71610	20.287	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	81.160%
21) 2-Butanone-d5	7.088	46	74723	49.789	ug/L	-0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	99.580%
24) Chloroform-d	7.648	84	338992	22.892	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	91.560%
26) 1,2-Dichloroethane-d4	8.307	65	159836	22.988	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.960%
32) Benzene-d6	8.276	84	712648	21.725	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	86.920%
36) 1,2-Dichloropropane-d6	9.276	67	204876	22.046	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	88.200%
41) Toluene-d8	10.319	98	646795	21.252	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.000%
43) trans-1,3-Dichloroprop...	10.575	79	79037	21.967	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	87.880%
47) 2-Hexanone-d5	10.922	63	74738	55.187	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	110.380%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	172000	24.474	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	97.880%
66) 1,2-Dichlorobenzene-d4	13.848	152	216759	21.844	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	87.360%
Target Compounds						Qvalue
13) Acetone	4.167	43	7620	5.593	ug/L	99
15) Methyl Acetate	4.692	43	2071	0.841	ug/L #	89
51) Chlorobenzene	11.654	112	8908	0.398	ug/L	96
65) 1,4-Dichlorobenzene	13.574	146	48617	2.863	ug/L	98
70) 1,2,4-trichlorobenzene	15.123	180	11059	1.145	ug/L	90
72) 1,2,3-Trichlorobenzene	15.549	180	3793	0.470	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122023\
Data File : VW027807.D
Acq On : 20 Dec 2023 13:56
Operator : SY/MD
Sample : 05869-05
Misc : 4.29g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0B36

Quant Time: Dec 20 23:16:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121523SMA.M
Quant Title : SFAM01.0
QLast Update : Sun Dec 17 01:25:09 2023
Response via : Initial Calibration

