




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122024\
Data File : VW031441.D
Acq On : 20 Dec 2024 21:27
Operator : SY/MD
Sample : P5333-20
Misc : 11.70g/10mL/MSVOA_W/SOIL/A
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E2923

Manual IntegrationsAPPROVED

Quant Time: Dec 20 23:59:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 20 23:42:37 2024
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 12/23/2024
Supervised By :Mahesh Dadoda 12/23/2024



TIC: VW031441.D\data.ms

(43) trans-1,3-Dichloropropene-d4 (S)

10.581min (+ 0.006) 0.93 ug/L

response 1094

Ion	Exp%	Act%
79.00	100.00	100.00
81.00	31.90	29.89
42.00	46.20	49.91
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122024\
Data File : VW031441.D
Acq On : 20 Dec 2024 21:27
Operator : SY/MD
Sample : P5333-20
Misc : 11.70g/10mL/MSVOA_W/SOIL/A
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E2923

Manual IntegrationsAPPROVED

Quant Time: Dec 20 23:59:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 20 23:42:37 2024
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 12/23/2024
Supervised By :Mahesh Dadoda 12/23/2024



TIC: VW031441.D\data.ms

(43) trans-1,3-Dichloropropene-d4 (S)

10.581min (+ 0.006) 1.13 ug/L m

response 1324

Ion	Exp%	Act%
79.00	100.00	100.00
81.00	31.90	24.70
42.00	46.20	41.24
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\Vw122024\
 Data File : VW031441.D
 Acq On : 20 Dec 2024 21:27
 Operator : SY/MD
 Sample : P5333-20
 Misc : 11.70g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E2923

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 12/23/2024
 Supervised By :Mahesh Dadoda 12/23/2024

Quant Time: Dec 20 23:59:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 20 23:42:37 2024
 Response via : Initial Calibration

Compound	R.T.	Q	Ion	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene	8.837	114		314625	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117		136283	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152		24683	25.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3	2.363	65		153804	32.490	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150		Recovery	=	129.960%	
7) Chloroethane-d5	2.893	69		113165	32.316	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150		Recovery	=	129.280%	
11) 1,1-Dichloroethene-d2	4.021	65		53842	23.978	ug/L	0.00
Spiked Amount 25.000	Range 45	- 110		Recovery	=	95.920%	
21) 2-Butanone-d5	0.000	46		0d	0.000	ug/L	
Spiked Amount 50.000	Range 20	- 135		Recovery	=	0.000%#	
24) Chloroform-d	7.642	84		167555	19.437	ug/L	0.00
Spiked Amount 25.000	Range 40	- 150		Recovery	=	77.760%	
26) 1,2-Dichloroethane-d4	8.307	65		104522	24.143	ug/L	0.00
Spiked Amount 25.000	Range 70	- 130		Recovery	=	96.560%	
32) Benzene-d6	8.270	84		293588	33.611	ug/L	0.00
Spiked Amount 25.000	Range 20	- 135		Recovery	=	134.440%	
36) 1,2-Dichloropropane-d6	9.270	67		90403	35.511	ug/L	0.00
Spiked Amount 25.000	Range 70	- 120		Recovery	=	142.040%#	
41) Toluene-d8	10.319	98		172118	21.181	ug/L	0.00
Spiked Amount 25.000	Range 30	- 130		Recovery	=	84.720%	
43) trans-1,3-Dichloroprop...	10.581	79		1324m	1.128	ug/L	0.00
Spiked Amount 25.000	Range 30	- 135		Recovery	=	4.520%#	
47) 2-Hexanone-d5	0.000	63		0d	0.000	ug/L	
Spiked Amount 50.000	Range 20	- 135		Recovery	=	0.000%#	
56) 1,1,2,2-Tetrachloroeth...	12.690	84		46015	28.788	ug/L	0.00
Spiked Amount 25.000	Range 45	- 120		Recovery	=	115.160%	
66) 1,2-Dichlorobenzene-d4	13.848	152		15676	17.107	ug/L	0.00
Spiked Amount 25.000	Range 75	- 120		Recovery	=	68.440%#	
Target Compounds							
13) Acetone	4.179	43		9448	7.610	ug/L #	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122024\
Data File : VW031441.D
Acq On : 20 Dec 2024 21:27
Operator : SY/MD
Sample : P5333-20
Misc : 11.70g/10mL/MSVOA_W/SOIL/A
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E2923

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 12/23/2024
Supervised By :Mahesh Dadoda 12/23/2024

Quant Time: Dec 20 23:59:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 20 23:42:37 2024
Response via : Initial Calibration

