

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122124\
 Data File : VW031474.D
 Acq On : 21 Dec 2024 20:56
 Operator : SY/MD
 Sample : P5333-18
 Misc : 5.71g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Dec 23 00:21:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Dec 23 00:16:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	867004	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	645937	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	211458	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	207605	15.914	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	63.640%
7) Chloroethane-d5	2.899	69	167958	17.405	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	69.640%
11) 1,1-Dichloroethene-d2	4.015	65	90620	14.645	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	58.560%
21) 2-Butanone-d5	7.106	46	45740	18.618	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	37.240%
24) Chloroform-d	7.648	84	345446	14.542	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	58.160%
26) 1,2-Dichloroethane-d4	8.544	65	86	0.007	ug/L	0.24
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.040%#
32) Benzene-d6	8.270	84	725826	17.532	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	70.120%
36) 1,2-Dichloropropane-d6	9.270	67	208475	17.278	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	69.120%#
41) Toluene-d8	10.319	98	637994	16.565	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	66.240%
43) trans-1,3-Dichloroprop...	10.575	79	58358	10.489	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	41.960%
47) 2-Hexanone-d5	10.922	63	38106	23.483	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	46.960%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	117923	15.565	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	62.280%
66) 1,2-Dichlorobenzene-d4	13.848	152	118334	15.074	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	60.280%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122124\
Data File : VW031474.D
Acq On : 21 Dec 2024 20:56
Operator : SY/MD
Sample : P5333-18
Misc : 5.71g/10mL/MSVOA_W/SOIL/B
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Dec 23 00:21:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Dec 23 00:16:39 2024
Response via : Initial Calibration

