

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW122218\
 Data File : VW007787.D
 Acq On : 21 Dec 2018 11:16
 Operator : SY/AP
 Sample : J6389-12
 Misc : 6.01G/5ML/MSVOA_W/SOIL
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 P001-SS014-1824-01

Quant Time: Dec 21 12:28:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 20 06:30:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	62538	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	105643	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	68959	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	20134	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	35070	51.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.32%	
35) Dibromofluoromethane	7.88	113	29018	45.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.06%	
50) Toluene-d8	10.32	98	80666	33.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	66.40%	
62) 4-Bromofluorobenzene	12.62	95	21747	23.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	46.34%	
Target Compounds						
16) Acetone	4.12	43	23749	185.024	ug/l	96
25) 2-Butanone	7.18	43	2558	17.014	ug/l #	77
51) 4-Methyl-2-Pentanone	10.21	43	984	2.953	ug/l	100
52) Toluene	10.39	92	6188	3.390	ug/l	96
67) Ethyl Benzene	11.73	91	2722	1.063	ug/l	98
68) m/p-Xylenes	11.84	106	5464	5.500	ug/l	99
69) o-Xylene	12.17	106	1971	2.116	ug/l	98
84) 1,2,4-Trimethylbenzene	13.26	105	2093	1.580	ug/l	97
95) Naphthalene	15.37	128	1842	2.315	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW122218\
Data File : VW007787.D
Acq On : 21 Dec 2018 11:16
Operator : SY/AP
Sample : J6389-12
Misc : 6.01G/5ML/MSVOA_W/SOIL
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
P001-SS014-1824-01

Quant Time: Dec 21 12:28:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 20 06:30:23 2018
Response via : Initial Calibration

