


Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW122218\
Data File : VW007745.D
Acq On : 20 Dec 2018 14:03
Operator : SY/AP
Sample : BFB
Misc : 5.00G/5ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
Title : SW846 8260
Last Update : Thu Dec 20 06:30:23 2018



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.8	11183	PASS
75	95	30	60	52.7	33117	PASS
95	95	100	100	100.0	62829	PASS
96	95	5	9	6.4	4044	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	76.7	48168	PASS
175	174	5	9	8.5	4084	PASS
176	174	95	101	97.0	46720	PASS
177	176	5	9	6.7	3151	PASS