

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122221\
 Data File : VW021569.D
 Acq On : 22 Dec 2021 17:41
 Operator : SY/VA
 Sample : VIBLK562
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK562

Quant Time: Dec 23 00:37:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Dec 23 00:34:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	121574	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	122307	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	70924	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	36908	16.669	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	66.680%
7) Chloroethane-d5	2.879	69	25957	17.733	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	70.920%
11) 1,1-Dichloroethene-d2	4.013	63	50320	17.047	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	68.200%
21) 2-Butanone-d5	7.080	46	20367	65.255	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	130.520%
24) Chloroform-d	7.647	84	82298	24.291	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	97.160%
26) 1,2-Dichloroethane-d4	8.305	65	48625	26.981	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	107.920%
32) Benzene-d6	8.275	84	160395	23.925	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.680%
36) 1,2-Dichloropropane-d6	9.268	67	45816	25.482	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.920%
41) Toluene-d8	10.323	98	155204	22.658	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	90.640%
43) trans-1,3-Dichloroprop...	10.579	79	19716	23.816	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	95.280%
47) 2-Hexanone-d5	10.921	63	17447	65.623	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	131.240%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	41107	26.596	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	106.400%
66) 1,2-Dichlorobenzene-d4	13.853	152	68845	25.046	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	100.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122221\
Data File : VW021569.D
Acq On : 22 Dec 2021 17:41
Operator : SY/VA
Sample : VIBLK562
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK562

Quant Time: Dec 23 00:37:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Dec 23 00:34:33 2021
Response via : Initial Calibration

