

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122221\
 Data File : VW021589.D
 Acq On : 23 Dec 2021 02:30
 Operator : SY/VA
 Sample : M5213-13
 Misc : 6.34g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGLE5

Quant Time: Dec 23 03:41:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Dec 23 02:55:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	143704	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	139779	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	77640	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	82625	31.570	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery = 126.280%			
7) Chloroethane-d5	2.879	69	60109	34.741	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery = 138.960%			
11) 1,1-Dichloroethene-d2	4.013	63	111643	31.996	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery = 128.000%#			
21) 2-Butanone-d5	7.086	46	30807	83.505	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery = 167.000%#			
24) Chloroform-d	7.653	84	182399	45.546	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery = 182.200%#			
26) 1,2-Dichloroethane-d4	8.305	65	96038	45.084	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery = 180.320%#			
32) Benzene-d6	8.275	84	350746	45.779	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery = 183.120%#			
36) 1,2-Dichloropropane-d6	9.275	67	96459	46.943	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery = 187.760%#			
41) Toluene-d8	10.323	98	357362	45.650	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery = 182.600%#			
43) trans-1,3-Dichloroprop...	10.579	79	41048	43.386	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery = 173.560%#			
47) 2-Hexanone-d5	10.921	63	26552	87.386	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery = 174.780%#			
56) 1,1,2,2-Tetrachloroeth...	12.689	84	70808	40.086	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery = 160.360%#			
66) 1,2-Dichlorobenzene-d4	13.853	152	125238	41.621	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery = 166.480%#			
Target Compounds						Qvalue
46) Tetrachloroethene	10.866	164	5048	2.486	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122221\
Data File : VW021589.D
Acq On : 23 Dec 2021 02:30
Operator : SY/VA
Sample : M5213-13
Misc : 6.34g/10.0mL/MSVOA_W/SOIL
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGLE5

Quant Time: Dec 23 03:41:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Dec 23 02:55:43 2021
Response via : Initial Calibration

