

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122221\
 Data File : VW021563.D
 Acq On : 22 Dec 2021 15:17
 Operator : SY/VA
 Sample : VIBLK559
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK559

Quant Time: Dec 23 00:36:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Dec 23 00:34:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	102076	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	103624	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	60883	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	33668	18.110	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	72.440%
7) Chloroethane-d5	2.879	69	23894	19.442	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.760%
11) 1,1-Dichloroethene-d2	4.007	63	45955	18.542	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	74.160%
21) 2-Butanone-d5	7.086	46	20569	78.491	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	156.980%#
24) Chloroform-d	7.647	84	72863	25.614	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	102.440%
26) 1,2-Dichloroethane-d4	8.305	65	43128	28.502	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	114.000%
32) Benzene-d6	8.275	84	137819	24.264	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	97.040%
36) 1,2-Dichloropropane-d6	9.274	67	40332	26.477	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	105.920%
41) Toluene-d8	10.323	98	138658	23.892	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	95.560%
43) trans-1,3-Dichloroprop...	10.573	79	17543	25.011	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	100.040%
47) 2-Hexanone-d5	10.920	63	17476	77.583	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	155.160%#
56) 1,1,2,2-Tetrachloroeth...	12.688	84	39009	29.789	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	119.160%
66) 1,2-Dichlorobenzene-d4	13.853	152	58810	24.924	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	99.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122221\
Data File : VW021563.D
Acq On : 22 Dec 2021 15:17
Operator : SY/VA
Sample : VIBLK559
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK559

Quant Time: Dec 23 00:36:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Dec 23 00:34:33 2021
Response via : Initial Calibration

