

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122221\
 Data File : VW021599.D
 Acq On : 23 Dec 2021 08:32
 Operator : SY/VA
 Sample : VW1222SBL03
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK428

Quant Time: Dec 24 00:30:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 24 00:28:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	98686	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	96077	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	55791	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	40543	22.557	ug/L	0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.240%
7) Chloroethane-d5	2.885	69	27098	22.806	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	91.240%
11) 1,1-Dichloroethene-d2	4.019	63	47422	19.791	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	79.160%
21) 2-Butanone-d5	7.080	46	16009	63.189	ug/L	-0.02
Spiked Amount	50.000	Range	20 - 135	Recovery	=	126.380%
24) Chloroform-d	7.647	84	66156	24.055	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	96.240%
26) 1,2-Dichloroethane-d4	8.305	65	38631	26.407	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	105.640%
32) Benzene-d6	8.275	84	129341	24.560	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	98.240%
36) 1,2-Dichloropropane-d6	9.274	67	35410	25.071	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	100.280%
41) Toluene-d8	10.323	98	131160	24.376	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	97.520%
43) trans-1,3-Dichloroprop...	10.579	79	15423	23.716	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	94.880%
47) 2-Hexanone-d5	10.927	63	13025	62.365	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	124.740%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	30448	25.078	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	100.320%
66) 1,2-Dichlorobenzene-d4	13.853	152	48147	22.267	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.080%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122221\
Data File : VW021599.D
Acq On : 23 Dec 2021 08:32
Operator : SY/VA
Sample : VW1222SBL03
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK428

Quant Time: Dec 24 00:30:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 24 00:28:54 2021
Response via : Initial Calibration

