

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122222\
 Data File : VW025173.D
 Acq On : 22 Dec 2022 23:59
 Operator : SY/MD
 Sample : N6193-02
 Misc : 5.06g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGSM1

Quant Time: Dec 23 03:36:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120822SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 23 02:08:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	340008	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	286589	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	132444	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	54284	13.850	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	55.400%
7) Chloroethane-d5	2.885	69	38931	15.383	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	61.520%
11) 1,1-Dichloroethene-d2	4.019	63	126739	14.760	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	59.040%
21) 2-Butanone-d5	7.074	46	40354	41.010	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	82.020%
24) Chloroform-d	7.647	84	207472	23.639	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.560%
26) 1,2-Dichloroethane-d4	8.305	65	117209	26.334	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	105.320%
32) Benzene-d6	8.274	84	387906	23.080	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	92.320%
36) 1,2-Dichloropropane-d6	9.274	67	108943	22.277	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	89.120%
41) Toluene-d8	10.323	98	351599	23.073	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	92.280%
43) trans-1,3-Dichloroprop...	10.579	79	47951	23.212	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	92.840%
47) 2-Hexanone-d5	10.920	63	33170	46.230	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	92.460%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	80468	22.282	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	89.120%
66) 1,2-Dichlorobenzene-d4	13.847	152	111187	24.768	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	99.080%
Target Compounds						Qvalue
16) Methylene chloride	4.921	84	17526	3.298	ug/L	90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122222\
Data File : VW025173.D
Acq On : 22 Dec 2022 23:59
Operator : SY/MD
Sample : N6193-02
Misc : 5.06g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGSM1


Quant Time: Dec 23 03:36:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120822SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 23 02:08:31 2022
Response via : Initial Calibration





#16
Methylene chloride
Concen: 3.298 ug/L
RT: 4.921 min Scan# 49
Delta R.T. 0.006 min
Lab File: VW025173.D
Acq: 22 Dec 2022 23:59

Instrument :
MSVOA_W
ClientSampleId :
BGSM1



Tgt Ion: 84 Resp: 17526
Ion Ratio Lower Upper
84 100
86 61.5 45.0 83.6
49 107.5 86.3 160.3

