

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122624\
 Data File : VW031535.D
 Acq On : 26 Dec 2024 17:47
 Operator : SY/MD
 Sample : VIBLK585
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK585

Quant Time: Dec 27 00:27:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 27 00:21:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	521218	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	466454	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.550	152	230072	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	153881	19.622	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.480%
7) Chloroethane-d5	2.893	69	131607	22.686	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.760%
11) 1,1-Dichloroethene-d2	4.015	65	72367	19.454	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	77.800%
21) 2-Butanone-d5	7.100	46	86681	58.691	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	117.380%
24) Chloroform-d	7.648	84	346473	24.261	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	97.040%
26) 1,2-Dichloroethane-d4	8.301	65	192153	26.792	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	107.160%
32) Benzene-d6	8.270	84	673297	22.521	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	90.080%
36) 1,2-Dichloropropane-d6	9.270	67	207976	23.869	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	95.480%
41) Toluene-d8	10.319	98	608921	21.893	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.560%
43) trans-1,3-Dichloroprop...	10.575	79	79173	19.706	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.840%
47) 2-Hexanone-d5	10.922	63	69099	58.969	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	117.940%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	166451	30.425	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	121.720%#
66) 1,2-Dichlorobenzene-d4	13.849	152	214774	25.146	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	100.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122624\
Data File : VW031535.D
Acq On : 26 Dec 2024 17:47
Operator : SY/MD
Sample : VIBLK585
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK585

Quant Time: Dec 27 00:27:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 27 00:21:46 2024
Response via : Initial Calibration

