

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122724\
 Data File : VW031550.D
 Acq On : 27 Dec 2024 14:30
 Operator : SY/MD
 Sample : P5378-06
 Misc : 5.01g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Dec 27 22:23:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 27 22:18:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	521529	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	465333	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	232971	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.356	65	146455	18.664	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	74.640%
7) Chloroethane-d5	2.899	69	131264	22.614	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.440%
11) 1,1-Dichloroethene-d2	4.015	65	66551	17.879	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	71.520%
21) 2-Butanone-d5	7.100	46	75793	51.288	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	102.580%
24) Chloroform-d	7.648	84	320518	22.430	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.720%
26) 1,2-Dichloroethane-d4	8.307	65	176350	24.574	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	98.280%
32) Benzene-d6	8.270	84	623775	20.914	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	83.640%
36) 1,2-Dichloropropane-d6	9.270	67	195043	22.438	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	89.760%
41) Toluene-d8	10.318	98	571933	20.613	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	82.440%
43) trans-1,3-Dichloroprop...	10.575	79	71992	17.962	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	71.840%
47) 2-Hexanone-d5	10.922	63	58743	50.252	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	100.500%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	148663	27.239	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	108.960%
66) 1,2-Dichlorobenzene-d4	13.848	152	202456	23.408	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	93.640%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122724\
Data File : VW031550.D
Acq On : 27 Dec 2024 14:30
Operator : SY/MD
Sample : P5378-06
Misc : 5.01g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Dec 27 22:23:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 27 22:18:12 2024
Response via : Initial Calibration

