

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122921\
 Data File : VW021689.D
 Acq On : 29 Dec 2021 14:26
 Operator : SY/VA
 Sample : VW1229SBL01
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK432

Quant Time: Dec 30 01:19:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM122821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Dec 30 01:18:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	84393	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	83705	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	46989	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	42075	25.286	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.160%
7) Chloroethane-d5	2.879	69	28195	25.158	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	100.640%
11) 1,1-Dichloroethene-d2	4.019	63	39438	16.196	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	64.800%
21) 2-Butanone-d5	7.086	46	10688	43.904	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	87.800%
24) Chloroform-d	7.647	84	59292	22.491	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.960%
26) 1,2-Dichloroethane-d4	8.305	65	31186	22.403	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	89.600%
32) Benzene-d6	8.275	84	113788	22.894	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	91.560%
36) 1,2-Dichloropropane-d6	9.274	67	31071	23.045	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	92.160%
41) Toluene-d8	10.323	98	110948	22.404	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	89.600%
43) trans-1,3-Dichloroprop...	10.579	79	14085	21.468	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.880%
47) 2-Hexanone-d5	10.927	63	8695	42.522	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.040%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	24385	21.937	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	87.760%
66) 1,2-Dichlorobenzene-d4	13.853	152	40772	22.690	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	90.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122921\
Data File : VW021689.D
Acq On : 29 Dec 2021 14:26
Operator : SY/VA
Sample : VW1229SBL01
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK432

Quant Time: Dec 30 01:19:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM122821SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Dec 30 01:18:20 2021
Response via : Initial Calibration

