

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122921\
 Data File : VW021697.D
 Acq On : 29 Dec 2021 18:25
 Operator : SY/VA
 Sample : VIBLK576
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK576

Quant Time: Dec 30 01:20:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM122821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Dec 30 01:18:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	79640	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	82608	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	45034	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	42658	27.167	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 108.680%	
7) Chloroethane-d5	2.879	69	28860	27.288	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 109.160%	
11) 1,1-Dichloroethene-d2	4.013	63	41087	17.880	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 71.520%	
21) 2-Butanone-d5	7.080	46	12470	54.281	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 108.560%	
24) Chloroform-d	7.653	84	59681	23.990	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 95.960%	
26) 1,2-Dichloroethane-d4	8.305	65	32074	24.416	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 97.680%	
32) Benzene-d6	8.275	84	114861	23.417	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 93.680%	
36) 1,2-Dichloropropane-d6	9.274	67	31455	23.639	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 94.560%	
41) Toluene-d8	10.323	98	114063	23.339	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 93.360%	
43) trans-1,3-Dichloroprop...	10.579	79	14620	22.580	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 90.320%	
47) 2-Hexanone-d5	10.927	63	10173	50.411	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 100.820%	
56) 1,1,2,2-Tetrachloroeth...	12.688	84	27568	25.130	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 100.520%	
66) 1,2-Dichlorobenzene-d4	13.853	152	42591	24.732	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 98.920%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122921\
Data File : VW021697.D
Acq On : 29 Dec 2021 18:25
Operator : SY/VA
Sample : VIBLK576
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK576

Quant Time: Dec 30 01:20:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM122821SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Dec 30 01:18:20 2021
Response via : Initial Calibration

