


Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF010319\
 Data File : VF061262.D
 Acq On : 3 Jan 2019 12:12
 Operator : VA/AP
 Sample : BFB
 Misc : 5.00µ/5mL/MSVOA-F/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F010319S.M
 Title : SW846 8260
 Last Update : Fri Jan 04 03:27:18 2019



AutoFind: Scans 1092, 1093, 1094; Background Corrected with Scan 1087

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.1	24980	PASS
75	95	30	60	46.6	57840	PASS
95	95	100	100	100.0	124229	PASS
96	95	5	9	6.8	8478	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	63.3	78688	PASS
175	174	5	9	7.8	6105	PASS
176	174	95	101	97.7	76917	PASS
177	176	5	9	6.4	4907	PASS