

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF010819\
 Data File : VF061285.D
 Acq On : 8 Jan 2019 14:33
 Operator : VA/AP
 Sample : VSTDIC010
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDIC010

Quant Time: Jan 09 06:05:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F010819S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 15:56:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.88	168	214274	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.60	114	358772	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.79	117	331602	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.56	152	174711	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	4.86	65	23857	8.77	ug/l	0.00
Spiked Amount			Recovery	=	17.54%	
35) Dibromofluoromethane	4.12	113	26154	9.48	ug/l	0.00
Spiked Amount			Recovery	=	18.96%	
50) Toluene-d8	7.57	98	80740	9.86	ug/l	0.00
Spiked Amount			Recovery	=	19.72%	
62) 4-Bromofluorobenzene	11.41	95	35859	9.88	ug/l	0.00
Spiked Amount			Recovery	=	19.76%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	0.96	85	41361	10.053	ug/l	93
3) Chloromethane	1.09	50	37338	11.846	ug/l	90
4) Vinyl Chloride	1.14	62	33395	12.232	ug/l	95
5) Bromomethane	1.30	94	24421	13.084	ug/l	100
6) Chloroethane	1.37	64	15724	12.709	ug/l	91
7) Trichlorofluoromethane	1.46	101	55215	11.239	ug/l	99
8) Diethyl Ether	1.64	74	7351	10.488	ug/l	90
9) 1,1,2-Trichlorotrifluoroet	1.82	101	30412	12.390	ug/l	92
10) Methyl Iodide	1.86	142	46339	11.685	ug/l #	97
11) Tert butyl alcohol	2.58	59	4415	38.467	ug/l	93
12) 1,1-Dichloroethene	1.76	96	22560	10.934	ug/l	96
13) Acrolein	2.00	56	6401	59.585	ug/l	96
14) Allyl chloride	2.10	41	41328	14.489	ug/l	95
15) Acrylonitrile	2.96	53	15929	51.914	ug/l #	89
16) Acetone	2.24	43	31169	57.220	ug/l	93
17) Carbon Disulfide	1.78	76	80965	12.887	ug/l	96
18) Methyl Acetate	2.35	43	9822	8.148	ug/l	99
19) Methyl tert-butyl Ether	2.44	73	43968	10.223	ug/l	96
20) Methylene Chloride	2.19	84	25215	9.974	ug/l	94
21) trans-1,2-Dichloroethene	2.32	96	25178	11.603	ug/l	99
22) Diisopropyl ether	2.80	45	77902	10.862	ug/l	96
23) Vinyl Acetate	3.20	43	100886	30.714	ug/l	98
24) 1,1-Dichloroethane	2.88	63	46452	11.047	ug/l	97
25) 2-Butanone	4.34	43	47532	46.484	ug/l	94
26) 2,2-Dichloropropane	3.61	77	32693	13.501	ug/l	99
27) cis-1,2-Dichloroethene	3.48	96	32634	10.234	ug/l	93
28) Bromochloromethane	3.73	49	20921	10.510	ug/l #	97
29) Tetrahydrofuran	4.09	42	19017	47.833	ug/l #	91
30) Chloroform	3.87	83	59801	10.167	ug/l	91
31) Cyclohexane	3.71	56	50124	11.246	ug/l	99
32) 1,1,1-Trichloroethane	4.10	97	50219	13.145	ug/l	91
36) 1,1-Dichloropropene	4.29	75	43775	10.360	ug/l	93
37) Ethyl Acetate	4.11	43	15839	9.388	ug/l #	80
38) Carbon Tetrachloride	4.01	117	44006	13.009	ug/l	98

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF010819\
 Data File : VF061285.D
 Acq On : 8 Jan 2019 14:33
 Operator : VA/AP
 Sample : VSTDIC010
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDIC010

Quant Time: Jan 09 06:05:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F010819S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 15:56:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	5.47	83	55873	12.370	ug/l	97
40) Benzene	4.65	78	105261	10.977	ug/l	99
41) Methacrylonitrile	4.75	41	8813	9.187	ug/l #	86
42) 1,2-Dichloroethane	4.95	62	31488	9.814	ug/l	97
43) Isopropyl Acetate	6.96	43	18851	8.374	ug/l #	86
44) Trichloroethene	5.51	130	31022	10.393	ug/l	94
45) 1,2-Dichloropropane	6.24	63	24917	10.177	ug/l	97
46) Dibromomethane	6.10	93	15830	9.826	ug/l	96
47) Bromodichloromethane	6.39	83	37496	9.144	ug/l	88
48) Methyl methacrylate	6.73	41	12543	8.605	ug/l	96
49) 1,4-Dioxane	6.72	88	1914	189.505	ug/l #	72
51) 4-Methyl-2-Pentanone	8.27	43	79578	50.898	ug/l	98
52) Toluene	7.64	92	67813	10.830	ug/l	94
53) t-1,3-Dichloropropene	8.29	75	30696	9.182	ug/l	99
54) cis-1,3-Dichloropropene	7.31	75	37942	8.930	ug/l	97
55) 1,1,2-Trichloroethane	8.50	97	17584	9.765	ug/l	93
56) Ethyl methacrylate	8.63	69	19031	9.563	ug/l	98
57) 1,3-Dichloropropane	8.87	76	31545	9.319	ug/l	96
58) 2-Chloroethyl Vinyl ether	7.31	63	3230	13.365	ug/l #	90
59) 2-Hexanone	9.50	43	53945	47.659	ug/l	98
60) Dibromochloromethane	8.73	129	22969	8.980	ug/l	85
61) 1,2-Dibromoethane	9.00	107	17928	9.225	ug/l	97
64) Tetrachloroethene	8.16	164	25278	9.947	ug/l	90
65) Chlorobenzene	9.81	112	66653	9.885	ug/l	99
66) 1,1,1,2-Tetrachloroethane	9.94	131	27410	9.899	ug/l	94
67) Ethyl Benzene	9.91	91	141259	11.396	ug/l	93
68) m/p-Xylenes	10.14	106	95397	21.089	ug/l	98
69) o-Xylene	10.72	106	51005	10.423	ug/l	86
70) Styrene	10.80	104	68567	10.179	ug/l	96
71) Bromoform	10.78	173	10323	8.007	ug/l	88
73) Isopropylbenzene	11.13	105	160848	10.838	ug/l	98
74) N-amyl acetate	11.38	43	34528	9.215	ug/l	95
75) 1,1,2,2-Tetrachloroethane	11.70	83	24425	9.422	ug/l	98
76) 1,2,3-Trichloropropane	11.80	75	17804	9.252	ug/l	99
77) Bromobenzene	11.50	156	31120	9.989	ug/l	91
78) n-propylbenzene	11.60	91	195634	11.059	ug/l	96
79) 2-Chlorotoluene	11.73	91	106908	10.465	ug/l	95
80) 1,3,5-Trimethylbenzene	11.83	105	129771	10.831	ug/l	94
81) trans-1,4-Dichloro-2-buten	11.88	75	10552	11.964	ug/l #	88
82) 4-Chlorotoluene	11.91	91	110083	10.630	ug/l	96
83) tert-Butylbenzene	12.13	119	129830	10.510	ug/l	93
84) 1,2,4-Trimethylbenzene	12.21	105	130301	10.925	ug/l	96
85) sec-Butylbenzene	12.31	105	184031	10.983	ug/l	96
86) p-Isopropyltoluene	12.47	119	147727	10.700	ug/l	95
87) 1,3-Dichlorobenzene	12.48	146	63061	10.490	ug/l	97
88) 1,4-Dichlorobenzene	12.57	146	57470	9.798	ug/l	98
89) n-Butylbenzene	12.85	91	151667	10.692	ug/l	96
90) Hexachloroethane	12.92	117	31271	9.181	ug/l	95
91) 1,2-Dichlorobenzene	12.95	146	60744	10.460	ug/l	95
92) 1,2-Dibromo-3-Chloropropan	13.65	75	3421	7.344	ug/l	96


Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF010819\
Data File : VF061285.D
Acq On : 8 Jan 2019 14:33
Operator : VA/AP
Sample : VSTDICC010
Misc : 5.00g/5mL/MSVOA-F/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
VSTDICC010

Quant Time: Jan 09 06:05:58 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F010819S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 15:56:30 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	14.21	180	38664	9.508	ug/l	94
94) Hexachlorobutadiene	14.20	225	24533	9.469	ug/l	98
95) Naphthalene	14.46	128	69433	9.135	ug/l	99
96) 1,2,3-Trichlorobenzene	14.61	180	36237	9.611	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.88 min Scan# 431

Delta R.T. 0.01 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC010

Tot Ion:168 Resp: 214274

Ion Ratio Lower Upper

168 100

99 67.5 52.7 79.1



Abundance Ion 168.00 (167.70 to 168.70): V

Ion 99.00 (98.70 to 99.70): VF0

4.88

80000



60000

40000

20000

0

Time--> 4.70 4.80 4.90 5.00 5.10



#2

Dichlorodifluoromethane

Concen: 10.053 ug/l

RT: 0.96 min Scan# 33

Delta R.T. -0.00 min

Lab File: VF061285.D


Acq: 8 Jan 2019 14:33

Tgt Ion: 85 Resp: 41361

Ion Ratio Lower Upper

85 100

87 27.1 15.4 46.2



Abundance Ion 85.00 (84.70 to 85.70): VF0

Ion 87.00 (86.70 to 87.70): VF0

0.96


15000


10000

5000

0

Time--> 0.90 1.00 1.10 1.20





#3

Chloromethane

Concen: 11.846 ug/l

RT: 1.09 min Scan# 46

Delta R.T. 0.01 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC010

Tot Ion: 50 Resp: 37338

Ion Ratio Lower Upper




50 100

52 29.5 28.2 42.2



Abundance Ion 50.00 (49.70 to 50.70): VF0

Ion 52.00 (51.70 to 52.70): VF0



#4

Vinyl Chloride

Concen: 12.232 ug/l

RT: 1.14 min Scan# 51

Delta R.T. 0.01 min

Lab File: VF061285.D


Acq: 8 Jan 2019 14:33

Tgt Ion: 62 Resp: 33395

Ion Ratio Lower Upper



62 100


64 28.7 25.4 38.0



Abundance Ion 62.00 (61.70 to 62.70): VF0

Ion 64.00 (63.70 to 64.70): VF0





#5

Bromomethane

Concen: 13.084 ug/l

RT: 1.30 min Scan# 68

Delta R.T. 0.02 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC010

Tot Ion: 94 Resp: 24421

Ion Ratio Lower Upper




94 100

96 87.9 70.2 105.4



Abundance Ion 94.00 (93.70 to 94.70): VF0

Ion 96.00 (95.70 to 96.70): VF0



#6

Chloroethane

Concen: 12.709 ug/l

RT: 1.37 min Scan# 75

Delta R.T. 0.01 min

Lab File: VF061285.D


Acq: 8 Jan 2019 14:33

Tgt Ion: 64 Resp: 15724

Ion Ratio Lower Upper



64 100


66 34.2 23.4 35.0



Abundance Ion 64.00 (63.70 to 64.70): VF0

Ion 66.00 (65.70 to 66.70): VF0





#7

Trichlorofluoromethane

Concen: 11.239 ug/l

RT: 1.46 min Scan# 84

Delta R.T. 0.01 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC010

Tot Ion:101 Resp: 55215

Ion Ratio Lower Upper




101 100

103 62.9 49.9 74.9



Abundance Ion 101.00 (100.70 to 101.70): V

Ion 103.00 (102.70 to 103.70): V



#8

Diethyl Ether

Concen: 10.488 ug/l

RT: 1.64 min Scan# 102

Delta R.T. 0.01 min

Lab File: VF061285.D


Acq: 8 Jan 2019 14:33

Tgt Ion: 74 Resp: 7351

Ion Ratio Lower Upper



74 100


45 135.1 62.1 186.3



Abundance Ion 74.00 (73.70 to 74.70): VF0

Ion 45.00 (44.70 to 45.70): VF0





#9

1,1,2-Trichlorotrifluoroethane

Concen: 12.390 ug/l

RT: 1.82 min Scan# 120

Delta R.T. 0.01 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC010


Tot Ion:101 Resp: 30412

Ion Ratio Lower Upper

101 100

85 55.4 38.4 57.6




151 59.3 51.4 77.2



Abundance Ion 101.00 (100.70 to 101.70): V

Ion 85.00 (84.70 to 85.70): VF0

Ion 151.00 (150.70 to 151.70): V



#10

Methyl Iodide

Concen: 11.685 ug/l

RT: 1.86 min Scan# 124

Delta R.T. 0.01 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33


Tgt Ion:142 Resp: 46339

Ion Ratio Lower Upper

142 100

127 55.1 44.3 66.5



141 12.1 13.2 19.8#




Abundance Ion 142.00 (141.70 to 142.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 141.00 (140.70 to 141.70): V









#11

Tert butyl alcohol
 Concen: 38.467 ug/l
 RT: 2.58 min Scan# 197
 Delta R.T. 0.01 min
 Lab File: VF061285.D
 Acq: 8 Jan 2019 14:33

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDIC010




Tot Ion: 59 Resp: 4415
 Ion Ratio Lower Upper
 59 100
 57 15.8 15.1 22.7




#12

1,1-Dichloroethene
 Concen: 10.934 ug/l
 RT: 1.76 min Scan# 114
 Delta R.T. 0.01 min
 Lab File: VF061285.D
 Acq: 8 Jan 2019 14:33

Tgt Ion: 96 Resp: 22560
 Ion Ratio Lower Upper
 96 100
 61 209.1 171.8 257.8
 98 70.9 54.2 81.4





#13

Acrolein

Concen: 59.585 ug/l

RT: 2.00 min Scan# 139

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :





VSTDIC010

Tot Ion: 56 Resp: 6401

Ion Ratio Lower Upper

56 100

55 70.5 59.4 89.0



#14

Allyl chloride

Concen: 14.489 ug/l

RT: 2.10 min Scan# 149

Delta R.T. 0.01 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33




Tgt Ion: 41 Resp: 41328


Ion Ratio Lower Upper

41 100

39 100.0 82.6 123.8

76 28.5 26.9 40.3





#15

Acrylonitrile

Concen: 51.914 ug/l

RT: 2.96 min Scan# 236

Delta R.T. 0.02 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC010


Tot Ion: 53 Resp: 15929

Ion Ratio Lower Upper

53 100

52 85.6 74.6 111.8




51 36.2 37.8 56.6#



Abundance Ion 53.00 (52.70 to 53.70): VF0

Ion 52.00 (51.70 to 52.70): VF0

Ion 51.00 (50.70 to 51.70): VF0



#16

Acetone

Concen: 57.220 ug/l

RT: 2.24 min Scan# 163

Delta R.T. 0.01 min

Lab File: VF061285.D


Acq: 8 Jan 2019 14:33

Tgt Ion: 43 Resp: 31169

Ion Ratio Lower Upper



43 100


58 17.7 16.6 25.0



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0





#17

Carbon Disulfide

Concen: 12.887 ug/l

RT: 1.78 min Scan# 116

Delta R.T. 0.01 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC010

Tot Ion: 76 Resp: 80965

Ion Ratio Lower Upper




76 100

78 8.3 7.8 11.6



Abundance Ion 76.00 (75.70 to 76.70): VF0

Ion 78.00 (77.70 to 78.70): VF0



#18

Methyl Acetate

Concen: 8.148 ug/l

RT: 2.35 min Scan# 174

Delta R.T. 0.01 min

Lab File: VF061285.D


Acq: 8 Jan 2019 14:33

Tgt Ion: 43 Resp: 9822

Ion Ratio Lower Upper



43 100


74 14.9 11.6 17.4



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 74.00 (73.70 to 74.70): VF0









#19
Methvl tert-butvl Ether
Concen: 10.223 ug/l
RT: 2.44 min Scan# 183
Delta R.T. 0.01 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC010


Tot Ion: 73 Resp: 43968
Ion Ratio Lower Upper
73 100
57 21.1 15.5 23.3



#20
Methylene Chloride
Concen: 9.974 ug/l
RT: 2.19 min Scan# 158
Delta R.T. 0.01 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33

Tgt Ion: 84 Resp: 25215
Ion Ratio Lower Upper
84 100
49 152.9 116.6 175.0
51 46.4 37.7 56.5
86 68.9 49.8 74.8





#21

trans-1,2-Dichloroethene

Concen: 11.603 ug/l

RT: 2.32 min Scan# 171

Delta R.T. 0.01 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC010


Tot Ion: 96 Resp: 25178

Ion Ratio Lower Upper

96 100

61 157.1 127.0 190.4




98 65.5 51.7 77.5



Abundance Ion 96.00 (95.70 to 96.70): VF0

Ion 61.00 (60.70 to 61.70): VF0

Ion 98.00 (97.70 to 98.70): VF0



#22

Diisopropyl ether

Concen: 10.862 ug/l

RT: 2.80 min Scan# 220

Delta R.T. 0.01 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Tgt Ion: 45 Resp: 77902


Ion Ratio Lower Upper

45 100

43 49.6 37.4 56.2

87 18.4 14.2 21.2

59 12.0 8.2 12.2






Abundance Ion 45.00 (44.70 to 45.70): VF0

Ion 43.00 (42.70 to 43.70): VF0

Ion 87.00 (86.70 to 87.70): VF0

Ion 59.00 (58.70 to 59.70): VF0





#23

Vinyl Acetate

Concen: 30.714 ug/l

RT: 3.20 min Scan# 260

Delta R.T. 0.01 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :





VSTDIC010

Tot Ion: 43 Resp: 100886

Ion Ratio Lower Upper

43 100

86 7.5 6.5 9.7



#24

1,1-Dichloroethane

Concen: 11.047 ug/l

RT: 2.88 min Scan# 228

Delta R.T. 0.01 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33




Tgt Ion: 63 Resp: 46452


Ion Ratio Lower Upper

63 100

98 4.4 2.5 7.6

100 2.2 1.7 5.0





#25

2-Butanone

Concen: 46.484 ug/l

RT: 4.34 min Scan# 376

Delta R.T. 0.01 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :





VSTDIC010

Tot Ion: 43 Resp: 47532

Ion Ratio Lower Upper

43 100

72 17.8 16.5 24.7



#26

2,2-Dichloropropane

Concen: 13.501 ug/l

RT: 3.61 min Scan# 302

Delta R.T. 0.01 min

Lab File: VF061285.D




Acq: 8 Jan 2019 14:33


Tgt Ion: 77 Resp: 32693

Ion Ratio Lower Upper

77 100

97 17.6 8.6 25.9









#27

cis-1,2-Dichloroethene
Concen: 10.234 ug/l
RT: 3.48 min Scan# 289
Delta R.T. 0.01 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33

Instrument :
MSVOA_F
ClientSampleId :
VSTDIC010




Tot Ion	Ratio	Lower	Upper
96	100		
61	134.8	0.0	288.2
98	61.6	0.0	132.2




#28

Bromochloromethane
Concen: 10.510 ug/l
RT: 3.73 min Scan# 314
Delta R.T. 0.01 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33

Tgt Ion	Ratio	Lower	Upper
49	100		
129	1.9	0.0	0.0#
130	75.9	58.6	88.0









#29
Tetrahydrofuran
Concen: 47.833 ug/l
RT: 4.09 min Scan# 351
Delta R.T. 0.02 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC010


Tot Ion: 42 Resp: 19017
Ion Ratio Lower Upper
42 100
72 31.1 32.3 48.5#
71 35.5 29.5 44.3



#30
Chloroform
Concen: 10.167 ug/l
RT: 3.87 min Scan# 328
Delta R.T. -0.00 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33

Tgt Ion: 83 Resp: 59801
Ion Ratio Lower Upper
83 100
85 61.1 54.4 81.6





#31

Cyclohexane

Concen: 11.246 ug/l

RT: 3.71 min Scan# 312

Delta R.T. 0.01 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC010


Tot Ion: 56 Resp: 50124

Ion Ratio Lower Upper

56 100

69 31.3 24.7 37.1




84 90.9 72.2 108.4



Abundance Ion 56.00 (55.70 to 56.70): VF0

Ion 69.00 (68.70 to 69.70): VF0

Ion 84.00 (83.70 to 84.70): VF0



#32

1,1,1-Trichloroethane

Concen: 13.145 ug/l

RT: 4.10 min Scan# 352

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33


Tgt Ion: 97 Resp: 50219

Ion Ratio Lower Upper

97 100

99 57.2 52.7 79.1



61 46.8 40.8 61.2




Abundance Ion 97.00 (96.70 to 97.70): VF0

Ion 99.00 (98.70 to 99.70): VF0

Ion 61.00 (60.70 to 61.70): VF0









#33
 1,2-Dichloroethane-d4
 Concen: 8.771 ug/l
 RT: 4.86 min Scan# 429
 Delta R.T. 0.01 min
 Lab File: VF061285.D
 Acq: 8 Jan 2019 14:33




Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDIC010


Tot Ion: 65 Resp: 23857
 Ion Ratio Lower Upper
 65 100
 67 43.4 0.0 98.2



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 5.60 min Scan# 504
 Delta R.T. -0.00 min
 Lab File: VF061285.D
 Acq: 8 Jan 2019 14:33

Tgt Ion: 114 Resp: 358772
 Ion Ratio Lower Upper
 114 100
 63 21.5 0.0 42.8
 88 19.2 0.0 40.0





#35

Dibromofluoromethane

Concen: 9.475 ug/l

RT: 4.12 min Scan# 354

Delta R.T. 0.01 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC010


Tot Ion:113 Resp: 26154

Ion Ratio Lower Upper

113 100

111 106.3 78.0 117.0




192 17.4 10.6 16.0#



Abundance Ion 113.00 (112.70 to 113.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 192.00 (191.70 to 192.70): V



#36

1,1-Dichloropropene

Concen: 10.360 ug/l

RT: 4.29 min Scan# 371

Delta R.T. 0.01 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33


Tgt Ion: 75 Resp: 43775

Ion Ratio Lower Upper

75 100

110 28.9 16.6 49.8



77 28.3 25.1 37.7




Abundance Ion 75.00 (74.70 to 75.70): VF0

Ion 110.00 (109.70 to 110.70): V

Ion 77.00 (76.70 to 77.70): VF0





#37

Ethyl Acetate

Concen: 9.388 ug/l

RT: 4.11 min Scan# 353

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC010





Tot Ion: 43 Resp: 15839

Ion Ratio Lower Upper

43 100

61 166.9 113.3 169.9

70 7.8 9.4 14.0#



#38

Carbon Tetrachloride

Concen: 13.009 ug/l

RT: 4.01 min Scan# 342

Delta R.T. 0.01 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33




Tgt Ion: 117 Resp: 44006


Ion Ratio Lower Upper

117 100

119 98.0 78.5 117.7

121 28.0 25.3 37.9









#39
Methvlcvclohexane
Concen: 12.370 ug/l
RT: 5.47 min Scan# 490
Delta R.T. -0.00 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC010


Tot Ion: 83 Resp: 55873
Ion Ratio Lower Upper
83 100
55 87.0 66.4 99.6
98 49.6 39.1 58.7



#40
Benzene
Concen: 10.977 ug/l
RT: 4.65 min Scan# 407
Delta R.T. 0.01 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33

Tgt Ion: 78 Resp: 105261
Ion Ratio Lower Upper
78 100
77 24.5 19.1 28.7





#41

Methacrylonitrile

Concen: 9.187 ug/l

RT: 4.75 min Scan# 417

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC010

Tot Ion: 41 Resp: 8813





Ion Ratio Lower Upper

41 100

39 68.0 52.7 79.1

67 48.0 41.4 62.2

52 68.7 33.0 49.4#



#42

1,2-Dichloroethane

Concen: 9.814 ug/l

RT: 4.95 min Scan# 438

Delta R.T. 0.01 min

Lab File: VF061285.D




Acq: 8 Jan 2019 14:33


Tgt Ion: 62 Resp: 31488

Ion Ratio Lower Upper

62 100

98 9.5 0.0 16.8





#43

Isopropyl Acetate

Concen: 8.374 ug/l

RT: 6.96 min Scan# 642

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC010





Tot Ion: 43 Resp: 18851

Ion Ratio Lower Upper

43 100

61 37.8 24.3 36.5#

87 0.0 0.2 0.4#



#44

Trichloroethene

Concen: 10.393 ug/l

RT: 5.51 min Scan# 495

Delta R.T. -0.00 min

Lab File: VF061285.D




Acq: 8 Jan 2019 14:33


Tgt Ion: 130 Resp: 31022

Ion Ratio Lower Upper

130 100

95 98.3 0.0 184.4





#45

1,2-Dichloropropane

Concen: 10.177 ug/l

RT: 6.24 min Scan# 569

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC010

Tot Ion: 63 Resp: 24917

Ion Ratio Lower Upper




63 100

65 29.9 25.1 37.7



Abundance Ion 63.00 (62.70 to 63.70): VF0

Ion 65.00 (64.70 to 65.70): VF0



#46

Dibromomethane

Concen: 9.826 ug/l

RT: 6.10 min Scan# 554

Delta R.T. 0.01 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33


Tgt Ion: 93 Resp: 15830

Ion Ratio Lower Upper

93 100

95 92.2 68.8 103.2



174 78.8 62.9 94.3




Abundance Ion 93.00 (92.70 to 93.70): VF0

Ion 95.00 (94.70 to 95.70): VF0

Ion 174.00 (173.70 to 174.70): V





#47

Bromodichloromethane

Concen: 9.144 ug/l

RT: 6.39 min Scan# 584

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC010


Tot Ion: 83 Resp: 37496

Ion Ratio Lower Upper




83 100

85 73.4 50.4 75.6

127 8.7 6.4 9.6



Abundance Ion 83.00 (82.70 to 83.70): VF0
Ion 85.00 (84.70 to 85.70): VF0
Ion 127.00 (126.70 to 127.70): V



#48

Methyl methacrylate

Concen: 8.605 ug/l

RT: 6.73 min Scan# 618

Delta R.T. 0.01 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33


Tgt Ion: 41 Resp: 12543

Ion Ratio Lower Upper



41 100


69 61.2 51.5 77.3

39 52.3 44.1 66.1



Abundance Ion 41.00 (40.70 to 41.70): VF0
Ion 69.00 (68.70 to 69.70): VF0
Ion 39.00 (38.70 to 39.70): VF0









#49
1,4-Dioxane
Concen: 189.505 ug/l
RT: 6.72 min Scan# 617
Delta R.T. 0.01 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC010


Tot Ion: 88 Resp: 1914
Ion Ratio Lower Upper
88 100
43 57.4 38.5 57.7
58 40.4 57.0 85.4#



#50
Toluene-d8
Concen: 9.857 ug/l
RT: 7.57 min Scan# 703
Delta R.T. -0.00 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33

Tgt Ion: 98 Resp: 80740
Ion Ratio Lower Upper
98 100
100 72.8 55.6 83.4





#51

4-Methyl-2-Pentanone

Concen: 50.898 ug/l

RT: 8.27 min Scan# 774

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC010

Tot Ion: 43 Resp: 79578

Ion Ratio Lower Upper

43 100

58 35.9 29.6 44.4



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0

8.27

40000

30000



20000

10000

0

Time-->

8.20 8.30 8.40



#52

Toluene

Concen: 10.830 ug/l

RT: 7.64 min Scan# 710

Delta R.T. -0.00 min

Lab File: VF061285.D


Acq: 8 Jan 2019 14:33

Tgt Ion: 92 Resp: 67813

Ion Ratio Lower Upper

92 100

91 170.0 129.2 193.8



Abundance Ion 92.00 (91.70 to 92.70): VF0

Ion 91.00 (90.70 to 91.70): VF0

7.64

50000

40000

30000


20000


10000

0

Time-->

7.50 7.60 7.70 7.80





#53

t-1,3-Dichloropropene

Concen: 9.182 ug/l

RT: 8.29 min Scan# 776

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :





VSTDIC010

Tot Ion: 75 Resp: 30696

Ion Ratio Lower Upper

75 100

77 30.2 24.7 37.1



#54

cis-1,3-Dichloropropene

Concen: 8.930 ug/l

RT: 7.31 min Scan# 677

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33




Tgt Ion: 75 Resp: 37942


Ion Ratio Lower Upper

75 100

77 34.2 27.3 40.9

39 60.8 46.3 69.5





#55

1,1,2-Trichloroethane

Concen: 9.765 ug/l

RT: 8.50 min Scan# 798

Delta R.T. 0.01 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC010

Tot Ion: 97 Resp: 17584


Ion Ratio Lower Upper

97 100

83 79.2 67.8 101.6

85 64.0 45.1 67.7

99 68.3 51.0 76.6






Abundance Ion 97.00 (96.70 to 97.70): VF0

Ion 82.95 (82.65 to 83.65): VF0

Ion 84.95 (84.65 to 85.65): VF0

Ion 98.95 (98.65 to 99.65): VF0



#56

Ethyl methacrylate

Concen: 9.563 ug/l

RT: 8.63 min Scan# 811

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33


Tgt Ion: 69 Resp: 19031

Ion Ratio Lower Upper

69 100

41 72.2 59.0 88.4



39 46.4 35.3 52.9




Abundance Ion 69.00 (68.70 to 69.70): VF0

Ion 41.00 (40.70 to 41.70): VF0

Ion 39.00 (38.70 to 39.70): VF0





#57

1,3-Dichloropropane

Concen: 9.319 ug/l

RT: 8.87 min Scan# 835

Delta R.T. 0.01 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC010

Tot Ion: 76 Resp: 31545

Ion Ratio Lower Upper

76 100

78 31.1 26.8 40.2



Abundance Ion 76.00 (75.70 to 76.70): VF0

Ion 78.00 (77.70 to 78.70): VF0

15000



10000

5000

0

8.80 8.87 8.90 9.00

Time-->



#58

2-Chloroethyl Vinyl ether

Concen: 13.365 ug/l

RT: 7.31 min Scan# 677

Delta R.T. 0.01 min

Lab File: VF061285.D


Acq: 8 Jan 2019 14:33

Tgt Ion: 63 Resp: 3230

Ion Ratio Lower Upper

63 100

106 25.3 16.5 24.7#



Abundance Ion 63.00 (62.70 to 63.70): VF0

Ion 106.00 (105.70 to 106.70): V

1500


1000


500

0

7.25 7.30 7.35

Time-->









#59
2-Hexanone
Concen: 47.659 ug/l
RT: 9.50 min Scan# 899
Delta R.T. -0.00 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC010


Tot Ion: 43 Resp: 53945
Ion Ratio Lower Upper
43 100
58 49.6 25.6 76.8



#60
Dibromochloromethane
Concen: 8.980 ug/l
RT: 8.73 min Scan# 821
Delta R.T. 0.01 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33

Tgt Ion: 129 Resp: 22969
Ion Ratio Lower Upper
129 100
127 65.7 39.4 118.1





#61

1,2-Dibromoethane

Concen: 9.225 ug/l

RT: 9.00 min Scan# 848

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC010

Tot Ion:107 Resp: 17928

Ion Ratio Lower Upper




107 100

109 97.1 75.1 112.7



Abundance Ion 107.00 (106.70 to 107.70): V

Ion 109.00 (108.70 to 109.70): V



#62

4-Bromofluorobenzene

Concen: 9.880 ug/l

RT: 11.41 min Scan# 1093

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33


Tgt Ion: 95 Resp: 35859

Ion Ratio Lower Upper

95 100

174 55.4 0.0 114.2



176 58.7 0.0 117.8




Abundance Ion 95.00 (94.70 to 95.70): VF0

Ion 174.00 (173.70 to 174.70): V

Ion 176.00 (175.70 to 176.70): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 9.79 min Scan# 928
Delta R.T. -0.00 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33

Instrument :
MSVOA_F
ClientSampleId :
VSTDIC010

Tot Ion:117 Resp: 331602
Ion Ratio Lower Upper
117 100
82 56.9 45.3 67.9
119 31.5 25.4 38.0




Abundance Ion 117.00 (116.70 to 117.70): V
Ion 82.00 (81.70 to 82.70): VF0
Ion 119.00 (118.70 to 119.70): V






#64
Tetrachloroethene
Concen: 9.947 ug/l
RT: 8.16 min Scan# 763
Delta R.T. -0.00 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33

Tgt Ion:164 Resp: 25278
Ion Ratio Lower Upper
164 100
166 110.8 100.1 150.1
129 92.7 80.1 120.1
131 85.3 76.7 115.1



Abundance Ion 163.85 (163.55 to 164.55): V
Ion 165.80 (165.50 to 166.50): V
Ion 128.90 (128.60 to 129.60): V
Ion 130.90 (130.60 to 131.60): V





#65

Chlorobenzene

Concen: 9.885 ug/l

RT: 9.81 min Scan# 930

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC010

Tot Ion:112 Resp: 66653

Ion Ratio Lower Upper




112 100

114 33.0 25.8 38.8



Abundance Ion 112.00 (111.70 to 112.70): V

Ion 114.00 (113.70 to 114.70): V



#66

1,1,1,2-Tetrachloroethane

Concen: 9.899 ug/l

RT: 9.94 min Scan# 944

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33


Tgt Ion:131 Resp: 27410

Ion Ratio Lower Upper

131 100

133 91.6 49.9 149.6



119 69.7 33.9 101.6




Abundance Ion 131.00 (130.70 to 131.70): V

Ion 133.00 (132.70 to 133.70): V

Ion 119.00 (118.70 to 119.70): V









#67
Ethyl Benzene
Concen: 11.396 ug/l
RT: 9.91 min Scan# 941
Delta R.T. 0.01 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC010


Tot Ion: 91 Resp: 141259
Ion Ratio Lower Upper
91 100
106 30.5 21.4 32.2



#68
m/p-Xylenes
Concen: 21.089 ug/l
RT: 10.14 min Scan# 964
Delta R.T. -0.00 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33

Tgt Ion: 106 Resp: 95397
Ion Ratio Lower Upper
106 100
91 209.5 165.0 247.6









#69
o-Xylene
Concen: 10.423 ug/l
RT: 10.72 min Scan# 1023
Delta R.T. -0.00 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC010


Tot Ion:106 Resp: 51005
Ion Ratio Lower Upper
106 100
91 239.3 108.1 324.4



#70
Styrene
Concen: 10.179 ug/l
RT: 10.80 min Scan# 1031
Delta R.T. 0.01 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33

Tgt Ion:104 Resp: 68567
Ion Ratio Lower Upper
104 100
78 43.7 38.0 57.0
103 43.8 36.7 55.1









#71
Bromoform
Concen: 8.007 ug/l
RT: 10.78 min Scan# 1029
Delta R.T. -0.00 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC010


Tot Ion:173 Resp: 10323
Ion Ratio Lower Upper
173 100
175 56.1 24.1 72.2
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.56 min Scan# 1209
Delta R.T. -0.00 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33

Tgt Ion:152 Resp: 174711
Ion Ratio Lower Upper
152 100
115 58.8 29.6 88.9
150 153.5 0.0 315.4





#73

Isopropylbenzene

Concen: 10.838 ug/l

RT: 11.13 min Scan# 1064

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :





VSTDIC010

Tot Ion:105 Resp: 160848

Ion Ratio Lower Upper

105 100

120 24.7 12.9 38.7



#74

N-ethyl acetate

Concen: 9.215 ug/l

RT: 11.38 min Scan# 1090

Delta R.T. 0.01 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Tgt Ion: 43 Resp: 34528




Ion Ratio Lower Upper


43 100

70 39.3 27.4 41.0

55 19.6 14.7 22.1

61 23.6 19.0 28.4





#75

1,1,2,2-Tetrachloroethane

Concen: 9.422 ug/l

RT: 11.70 min Scan# 1122

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC010


Tot Ion: 83 Resp: 24425

Ion Ratio Lower Upper

83 100

131 7.0 3.9 11.7




85 64.2 31.6 94.7



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 131.00 (130.70 to 131.70): VF0

Ion 85.00 (84.70 to 85.70): VF0



#76

1,2,3-Trichloropropane

Concen: 9.252 ug/l

RT: 11.80 min Scan# 1132

Delta R.T. -0.00 min

Lab File: VF061285.D


Acq: 8 Jan 2019 14:33

Tgt Ion: 75 Resp: 17804

Ion Ratio Lower Upper



75 100


77 34.5 17.5 52.5



Abundance Ion 75.00 (74.70 to 75.70): VF0

Ion 77.00 (76.70 to 77.70): VF0





#77

Bromobenzene

Concen: 9.989 ug/l

RT: 11.50 min Scan# 1102

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC010


Tot Ion:156 Resp: 31120

Ion Ratio Lower Upper

156 100

77 187.1 84.3 252.9




158 92.6 47.1 141.3



Abundance Ion 156.00 (155.70 to 156.70): V

Ion 77.00 (76.70 to 77.70): VF0

Ion 158.00 (157.70 to 158.70): V



#78

n-propylbenzene

Concen: 11.059 ug/l

RT: 11.60 min Scan# 1112

Delta R.T. -0.00 min

Lab File: VF061285.D


Acq: 8 Jan 2019 14:33

Tgt Ion: 91 Resp: 195634

Ion Ratio Lower Upper



91 100


120 19.3 10.7 32.1



Abundance Ion 91.00 (90.70 to 91.70): VF0

Ion 120.00 (119.70 to 120.70): V









#79
2-Chlorotoluene
Concen: 10.465 ug/l
RT: 11.73 min Scan# 1125
Delta R.T. -0.00 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC010


Tot Ion: 91 Resp: 106908
Ion Ratio Lower Upper
91 100
126 30.4 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 10.831 ug/l
RT: 11.83 min Scan# 1135
Delta R.T. -0.00 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33

Tgt Ion: 105 Resp: 129771
Ion Ratio Lower Upper
105 100
120 46.0 25.2 75.6





#81

trans-1,4-Dichloro-2-butene

Concen: 11.964 ug/l

RT: 11.88 min Scan# 1140

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC010





Tot Ion: 75 Resp: 10552

Ion Ratio Lower Upper

75 100

53 70.7 62.2 93.2

89 38.2 40.2 60.4#



#82

4-Chlorotoluene

Concen: 10.630 ug/l

RT: 11.91 min Scan# 1143

Delta R.T. -0.00 min

Lab File: VF061285.D




Acq: 8 Jan 2019 14:33


Tgt Ion: 91 Resp: 110083

Ion Ratio Lower Upper

91 100

126 30.9 16.4 49.4









#83
tert-Butylbenzene
Concen: 10.510 ug/l
RT: 12.13 min Scan# 1166
Delta R.T. -0.00 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33

Instrument :
MSVOA_F
ClientSampleId :
VSTDIC010

Tot Ion:119 Resp: 129830
Ion Ratio Lower Upper
119 100
91 68.1 30.8 92.4
134 22.5 11.9 35.9




Abundance Ion 119.00 (118.70 to 119.70): V
Ion 91.00 (90.70 to 91.70): VF0
Ion 134.00 (133.70 to 134.70): V






#84
1,2,4-Trimethylbenzene
Concen: 10.925 ug/l
RT: 12.21 min Scan# 1174
Delta R.T. -0.00 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33

Tgt Ion:105 Resp: 130301
Ion Ratio Lower Upper
105 100
120 46.9 24.7 74.1



Abundance Ion 105.00 (104.70 to 105.70): V
Ion 120.00 (119.70 to 120.70): V





#85

sec-Butylbenzene

Concen: 10.983 ug/l

RT: 12.31 min Scan# 1184

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :





VSTDIC010

Tot Ion:105 Resp: 184031

Ion Ratio Lower Upper

105 100

134 17.5 9.7 28.9



#86

p-Isopropyltoluene

Concen: 10.700 ug/l

RT: 12.47 min Scan# 1200

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33




Tgt Ion:119 Resp: 147727


Ion Ratio Lower Upper

119 100

134 23.1 13.0 39.0

91 22.6 12.2 36.4





#87

1,3-Dichlorobenzene

Concen: 10.490 ug/l

RT: 12.48 min Scan# 1201

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC010





Tot Ion:146 Resp: 63061

Ion Ratio Lower Upper

146 100

111 42.1 21.1 63.3

148 62.6 33.1 99.5



#88

1,4-Dichlorobenzene

Concen: 9.798 ug/l

RT: 12.57 min Scan# 1210

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33




Tgt Ion:146 Resp: 57470


Ion Ratio Lower Upper

146 100

111 42.7 21.4 64.2

148 63.5 33.0 98.9









#89
n-Butylbenzene
Concen: 10.692 ug/l
RT: 12.85 min Scan# 1239
Delta R.T. -0.00 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC010


Tot Ion: 91 Resp: 151667
Ion Ratio Lower Upper
91 100
92 60.9 32.1 96.3
134 24.1 13.4 40.1



#90
Hexachloroethane
Concen: 9.181 ug/l
RT: 12.92 min Scan# 1246
Delta R.T. -0.00 min
Lab File: VF061285.D
Acq: 8 Jan 2019 14:33

Tgt Ion: 117 Resp: 31271
Ion Ratio Lower Upper
117 100
201 52.0 27.9 83.7









#91
 1,2-Dichlorobenzene
 Concen: 10.460 ug/l
 RT: 12.95 min Scan# 1249
 Delta R.T. -0.00 min
 Lab File: VF061285.D
 Acq: 8 Jan 2019 14:33




Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDIC010

Tot Ion:146 Resp: 60744
 Ion Ratio Lower Upper
 146 100
 111 40.2 21.4 64.3
 148 62.3 33.1 99.2



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 7.344 ug/l
 RT: 13.65 min Scan# 1320
 Delta R.T. -0.00 min
 Lab File: VF061285.D
 Acq: 8 Jan 2019 14:33

Tgt Ion: 75 Resp: 3421
 Ion Ratio Lower Upper
 75 100
 155 99.0 51.3 154.0
 157 117.0 60.7 182.0









#93
 1,2,4-Trichlorobenzene
 Concen: 9.508 ug/l
 RT: 14.21 min Scan# 1376
 Delta R.T. -0.00 min
 Lab File: VF061285.D
 Acq: 8 Jan 2019 14:33

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDIC010

Tot Ion:180 Resp: 38664
 Ion Ratio Lower Upper
 180 100
 182 100.9 47.0 141.2
 145 36.6 19.7 59.1




Abundance Ion 180.00 (179.70 to 180.70): V
 Ion 182.00 (181.70 to 182.70): V
 Ion 145.00 (144.70 to 145.70): V






#94
 Hexachlorobutadiene
 Concen: 9.469 ug/l
 RT: 14.20 min Scan# 1375
 Delta R.T. -0.00 min
 Lab File: VF061285.D
 Acq: 8 Jan 2019 14:33

Tgt Ion:225 Resp: 24533
 Ion Ratio Lower Upper
 225 100
 223 66.9 31.8 95.4
 227 62.4 31.3 93.9



Abundance Ion 225.00 (224.70 to 225.70): V
 Ion 223.00 (222.70 to 223.70): V
 Ion 227.00 (226.70 to 227.70): V





#95

Naphthalene

Concen: 9.135 ug/l

RT: 14.46 min Scan# 1402

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC010


Tot Ion:128 Resp: 69433

Ion Ratio Lower Upper

128 100

127 11.8 9.4 14.0




129 11.4 8.8 13.2



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V



#96

1,2,3-Trichlorobenzene

Concen: 9.611 ug/l

RT: 14.61 min Scan# 1417

Delta R.T. -0.00 min

Lab File: VF061285.D

Acq: 8 Jan 2019 14:33


Tgt Ion:180 Resp: 36237

Ion Ratio Lower Upper

180 100

182 88.5 46.5 139.3

145 35.7 17.5 52.6



Abundance Ion 180.00 (179.70 to 180.70): V

Ion 182.00 (181.70 to 182.70): V

Ion 145.00 (144.70 to 145.70): V

