

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF010819\
 Data File : VF061289.D
 Acq On : 8 Jan 2019 16:27
 Operator : VA/AP
 Sample : VSTDICC100
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDICC100

Quant Time: Jan 09 06:08:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F010819S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 15:56:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.89	168	186238	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	5.61	114	328018	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.79	117	307713	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.56	152	159767	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	4.86	65	215833	91.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	182.60%	
35) Dibromofluoromethane	4.13	113	240340	95.24	ug/l	0.01
Spiked Amount	50.000		Recovery	=	190.48%	
50) Toluene-d8	7.57	98	720325	96.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	192.38%	
62) 4-Bromofluorobenzene	11.42	95	298590	89.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	179.98%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	0.96	85	328726	91.928	ug/l	94
3) Chloromethane	1.08	50	293272	107.056	ug/l	96
4) Vinyl Chloride	1.14	62	263389	111.002	ug/l	99
5) Bromomethane	1.30	94	186238	114.801	ug/l	96
6) Chloroethane	1.37	64	128157	119.177	ug/l	97
7) Trichlorofluoromethane	1.47	101	462019	108.202	ug/l	98
8) Diethyl Ether	1.63	74	68518	112.478	ug/l	89
9) 1,1,2-Trichlorotrifluoroet	1.81	101	227653	106.712	ug/l	97
10) Methyl Iodide	1.85	142	378855	109.914	ug/l	97
11) Tert butyl alcohol	2.57	59	45063	451.735	ug/l	95
12) 1,1-Dichloroethene	1.76	96	186556	104.031	ug/l	89
13) Acrolein	2.01	56	42977	460.281	ug/l	94
14) Allyl chloride	2.10	41	352381	142.136	ug/l	98
15) Acrylonitrile	2.95	53	154057	577.665	ug/l	98
16) Acetone	2.23	43	213008	449.909	ug/l	94
17) Carbon Disulfide	1.77	76	600048	109.887	ug/l	99
18) Methyl Acetate	2.34	43	93940	89.663	ug/l	100
19) Methyl tert-butyl Ether	2.43	73	382502	102.328	ug/l	98
20) Methylene Chloride	2.19	84	183451	83.491	ug/l	99
21) trans-1,2-Dichloroethene	2.31	96	200697	106.408	ug/l	95
22) Diisopropyl ether	2.80	45	679261	108.963	ug/l	98
23) Vinyl Acetate	3.19	43	1205579	422.288	ug/l	100
24) 1,1-Dichloroethane	2.88	63	398479	109.026	ug/l	99
25) 2-Butanone	4.35	43	372355	418.967	ug/l	95
26) 2,2-Dichloropropane	3.62	77	265895	126.332	ug/l	98
27) cis-1,2-Dichloroethene	3.49	96	285081	102.861	ug/l	93
28) Bromochloromethane	3.73	49	156022	90.176	ug/l	92
29) Tetrahydrofuran	4.08	42	154992	448.537	ug/l	99
30) Chloroform	3.87	83	490409	95.926	ug/l	99
31) Cyclohexane	3.71	56	393728	101.637	ug/l	98
32) 1,1,1-Trichloroethane	4.11	97	404160	121.712	ug/l	97
36) 1,1-Dichloropropene	4.30	75	359196	92.979	ug/l	99
37) Ethyl Acetate	4.12	43	132727	86.048	ug/l #	94
38) Carbon Tetrachloride	4.01	117	416013	134.507	ug/l	97

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_F\DATA\VF010819\
 Data File : VF061289.D
 Acq On : 8 Jan 2019 16:27
 Operator : VA/AP
 Sample : VSTDICC100
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDICC100

Quant Time: Jan 09 06:08:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F010819S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 15:56:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	5.47	83	446417	108.102	ug/l	98
40) Benzene	4.64	78	812226	92.640	ug/l	98
41) Methacrylonitrile	4.75	41	76391	87.097	ug/l	92
42) 1,2-Dichloroethane	4.96	62	280767	95.711	ug/l	99
43) Isopropyl Acetate	6.97	43	199524	96.945	ug/l #	97
44) Trichloroethene	5.52	130	240027	87.951	ug/l	95
45) 1,2-Dichloropropane	6.25	63	233438	104.287	ug/l	98
46) Dibromomethane	6.10	93	147175	99.922	ug/l	92
47) Bromodichloromethane	6.40	83	360435	96.134	ug/l	92
48) Methyl methacrylate	6.72	41	130261	97.741	ug/l	97
49) 1,4-Dioxane	6.71	88	14519	1572.300	ug/l	90
51) 4-Methyl-2-Pentanone	8.27	43	697449	487.915	ug/l	97
52) Toluene	7.64	92	532183	92.960	ug/l	100
53) t-1,3-Dichloropropene	8.29	75	279903	91.578	ug/l	96
54) cis-1,3-Dichloropropene	7.32	75	368616	94.890	ug/l	99
55) 1,1,2-Trichloroethane	8.50	97	159921	97.137	ug/l	97
56) Ethyl methacrylate	8.63	69	197090	108.322	ug/l	97
57) 1,3-Dichloropropane	8.86	76	284780	92.020	ug/l	99
58) 2-Chloroethyl Vinyl ether	7.31	63	39173	177.289	ug/l #	88
59) 2-Hexanone	9.50	43	500205	483.356	ug/l	98
60) Dibromochloromethane	8.73	129	227178	97.146	ug/l	98
61) 1,2-Dibromoethane	9.00	107	175172	98.588	ug/l	93
64) Tetrachloroethene	8.16	164	218763	92.768	ug/l	96
65) Chlorobenzene	9.81	112	597479	95.488	ug/l	97
66) 1,1,1,2-Tetrachloroethane	9.95	131	257223	100.102	ug/l	95
67) Ethyl Benzene	9.91	91	1118625	97.252	ug/l	93
68) m/p-Xylenes	10.15	106	780250	185.877	ug/l	91
69) o-Xylene	10.73	106	440856	97.085	ug/l	96
70) Styrene	10.80	104	578353	92.520	ug/l	99
71) Bromoform	10.79	173	110311	92.209	ug/l #	99
73) Isopropylbenzene	11.13	105	1213009	89.381	ug/l	97
74) N-amyl acetate	11.38	43	308124	89.928	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.70	83	218227	92.060	ug/l	98
76) 1,2,3-Trichloropropane	11.80	75	156273	88.805	ug/l	96
77) Bromobenzene	11.51	156	263435	92.466	ug/l	95
78) n-propylbenzene	11.61	91	1404858	86.847	ug/l	95
79) 2-Chlorotoluene	11.73	91	831330	88.986	ug/l	96
80) 1,3,5-Trimethylbenzene	11.83	105	968509	88.399	ug/l	95
81) trans-1,4-Dichloro-2-buten	11.88	75	93648	116.111	ug/l	96
82) 4-Chlorotoluene	11.91	91	848228	89.565	ug/l	97
83) tert-Butylbenzene	12.14	119	956442	84.668	ug/l	96
84) 1,2,4-Trimethylbenzene	12.22	105	996428	91.357	ug/l	98
85) sec-Butylbenzene	12.32	105	1373527	89.638	ug/l	95
86) p-Isopropyltoluene	12.46	119	1107888	87.751	ug/l	98
87) 1,3-Dichlorobenzene	12.48	146	479293	87.184	ug/l	99
88) 1,4-Dichlorobenzene	12.57	146	502603	93.701	ug/l	98
89) n-Butylbenzene	12.85	91	1150093	88.661	ug/l	96
90) Hexachloroethane	12.93	117	281707	90.440	ug/l	88
91) 1,2-Dichlorobenzene	12.95	146	473871	89.230	ug/l	97
92) 1,2-Dibromo-3-Chloropropan	13.65	75	43875	102.995	ug/l	72


Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF010819\
Data File : VF061289.D
Acq On : 8 Jan 2019 16:27
Operator : VA/AP
Sample : VSTDICC100
Misc : 5.00g/5mL/MSVOA-F/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
VSTDICC100

Quant Time: Jan 09 06:08:23 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F010819S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 15:56:30 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	14.21	180	329896	88.712	ug/l	95
94) Hexachlorobutadiene	14.20	225	214263	90.433	ug/l	97
95) Naphthalene	14.47	128	711713	102.390	ug/l	99
96) 1,2,3-Trichlorobenzene	14.61	180	313615	90.957	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.89 min Scan# 432

Delta R.T. 0.01 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC100

Tot Ion:168 Resp: 186238

Ion Ratio Lower Upper

168 100

99 67.8 52.7 79.1



Abundance Ion 168.00 (167.70 to 168.70): VF0

Ion 99.00 (98.70 to 99.70): VF0

4.89



60000

40000

20000

0

Time--> 4.70 4.80 4.90 5.00 5.10



#2

Dichlorodifluoromethane

Concen: 91.928 ug/l

RT: 0.96 min Scan# 34

Delta R.T. 0.00 min

Lab File: VF061289.D


Acq: 8 Jan 2019 16:27

Tgt Ion: 85 Resp: 328726

Ion Ratio Lower Upper

85 100

87 33.9 15.4 46.2



Abundance Ion 85.00 (84.70 to 85.70): VF0

Ion 87.00 (86.70 to 87.70): VF0

0.96

120000

100000

80000


60000


40000

20000

0

Time--> 0.80 0.90 1.00 1.10 1.20





#3

Chloromethane

Concen: 107.056 ug/l

RT: 1.08 min Scan# 46

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC100

Tot Ion: 50 Resp: 293272

Ion Ratio Lower Upper

50 100

52 32.7 28.2 42.2



Abundance Ion 50.00 (49.70 to 50.70): VF0

Ion 52.00 (51.70 to 52.70): VF0

100000

80000

60000

40000

20000

0

1.08



1.00

1.10

1.20

1.30

Time-->



#4

Vinyl Chloride

Concen: 111.002 ug/l

RT: 1.14 min Scan# 52

Delta R.T. 0.01 min

Lab File: VF061289.D


Acq: 8 Jan 2019 16:27

Tgt Ion: 62 Resp: 263389

Ion Ratio Lower Upper

62 100

64 32.5 25.4 38.0



Abundance Ion 62.00 (61.70 to 62.70): VF0

Ion 64.00 (63.70 to 64.70): VF0

80000

60000

40000

20000

0

1.14


1.00


1.10

1.20

1.30

Time-->





#5

Bromomethane

Concen: 114.801 ug/l

RT: 1.30 min Scan# 68

Delta R.T. 0.01 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC100

Tot Ion: 94 Resp: 186238

Ion Ratio Lower Upper




94 100

96 91.4 70.2 105.4



Abundance Ion 94.00 (93.70 to 94.70): VF0

Ion 96.00 (95.70 to 96.70): VF0



#6

Chloroethane

Concen: 119.177 ug/l

RT: 1.37 min Scan# 75

Delta R.T. 0.00 min

Lab File: VF061289.D


Acq: 8 Jan 2019 16:27

Tgt Ion: 64 Resp: 128157

Ion Ratio Lower Upper



64 100


66 30.7 23.4 35.0



Abundance Ion 64.00 (63.70 to 64.70): VF0

Ion 66.00 (65.70 to 66.70): VF0









#7

Trichlorofluoromethane
 Concen: 108.202 ug/l
 RT: 1.47 min Scan# 85
 Delta R.T. 0.01 min
 Lab File: VF061289.D
 Acq: 8 Jan 2019 16:27

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDIC100




Tot Ion:101 Resp: 462019
 Ion Ratio Lower Upper
 101 100
 103 63.9 49.9 74.9




#8

Diethyl Ether
 Concen: 112.478 ug/l
 RT: 1.63 min Scan# 102
 Delta R.T. 0.00 min
 Lab File: VF061289.D
 Acq: 8 Jan 2019 16:27

Tgt Ion: 74 Resp: 68518
 Ion Ratio Lower Upper
 74 100
 45 136.9 62.1 186.3





#9

1,1,2-Trichlorotrifluoroethane

Concen: 106.712 ug/l

RT: 1.81 min Scan# 120

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC100


Tot Ion:101 Resp: 227653

Ion Ratio Lower Upper

101 100

85 50.1 38.4 57.6




151 67.2 51.4 77.2



Abundance Ion 101.00 (100.70 to 101.70): V

Ion 85.00 (84.70 to 85.70): VF0

Ion 151.00 (150.70 to 151.70): V



#10

Methyl Iodide

Concen: 109.914 ug/l

RT: 1.85 min Scan# 124

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27


Tgt Ion:142 Resp: 378855

Ion Ratio Lower Upper

142 100

127 57.9 44.3 66.5



141 16.0 13.2 19.8




Abundance Ion 142.00 (141.70 to 142.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 141.00 (140.70 to 141.70): V









#11

Tert butyl alcohol
Concen: 451.735 ug/l
RT: 2.57 min Scan# 197
Delta R.T. 0.00 min
Lab File: VF061289.D
Acq: 8 Jan 2019 16:27

Instrument :
MSVOA_F
ClientSampleId :
VSTDIC100




Tot Ion: 59 Resp: 45063
Ion Ratio Lower Upper
59 100
57 16.7 15.1 22.7




#12

1,1-Dichloroethene
Concen: 104.031 ug/l
RT: 1.76 min Scan# 115
Delta R.T. 0.01 min
Lab File: VF061289.D
Acq: 8 Jan 2019 16:27

Tgt Ion: 96 Resp: 186556
Ion Ratio Lower Upper
96 100
61 194.2 171.8 257.8
98 65.6 54.2 81.4





#13

Acrolein

Concen: 460.281 ug/l

RT: 2.01 min Scan# 140

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :





VSTDIC100

Tot Ion: 56 Resp: 42977

Ion Ratio Lower Upper

56 100

55 69.2 59.4 89.0



#14

Allyl chloride

Concen: 142.136 ug/l

RT: 2.10 min Scan# 149

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27




Tgt Ion: 41 Resp: 352381


Ion Ratio Lower Upper

41 100

39 101.4 82.6 123.8

76 31.4 26.9 40.3





#15

Acrylonitrile

Concen: 577.665 ug/l

RT: 2.95 min Scan# 235

Delta R.T. 0.00 min


Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :
MSVOA_F
ClientSampleId :
VSTDIC100

Tot Ion: 53 Resp: 154057




Ion	Ratio	Lower	Upper
53	100		
52	90.5	74.6	111.8
51	46.5	37.8	56.6



Abundance Ion 53.00 (52.70 to 53.70): VF0

Ion 52.00 (51.70 to 52.70): VF0

Ion 51.00 (50.70 to 51.70): VF0



#16

Acetone

Concen: 449.909 ug/l

RT: 2.23 min Scan# 163


Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27



Tgt Ion: 43 Resp: 213008


Ion	Ratio	Lower	Upper
43	100		
58	17.8	16.6	25.0



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0





#17

Carbon Disulfide

Concen: 109.887 ug/l

RT: 1.77 min Scan# 116

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC100

Tot Ion: 76 Resp: 600048

Ion Ratio Lower Upper

76 100

78 10.2 7.8 11.6



Abundance Ion 76.00 (75.70 to 76.70): VF0

Ion 78.00 (77.70 to 78.70): VF0

1.77

200000



150000

100000

50000

0

Time-->



#18

Methyl Acetate

Concen: 89.663 ug/l

RT: 2.34 min Scan# 174

Delta R.T. 0.00 min

Lab File: VF061289.D


Acq: 8 Jan 2019 16:27

Tgt Ion: 43 Resp: 93940

Ion Ratio Lower Upper

43 100

74 14.5 11.6 17.4



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 74.00 (73.70 to 74.70): VF0

2.34

50000

40000


30000


20000

10000

0

Time-->









#19

Methyl tert-butyl Ether
 Concen: 102.328 ug/l
 RT: 2.43 min Scan# 183
 Delta R.T. 0.00 min
 Lab File: VF061289.D
 Acq: 8 Jan 2019 16:27

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDIC100




Tot Ion: 73 Resp: 382502
 Ion Ratio Lower Upper
 73 100
 57 18.3 15.5 23.3




#20

Methylene Chloride
 Concen: 83.491 ug/l
 RT: 2.19 min Scan# 158
 Delta R.T. 0.00 min
 Lab File: VF061289.D
 Acq: 8 Jan 2019 16:27

Tgt Ion: 84 Resp: 183451
 Ion Ratio Lower Upper
 84 100
 49 147.4 116.6 175.0
 51 45.8 37.7 56.5
 86 62.3 49.8 74.8





#21

trans-1,2-Dichloroethene

Concen: 106.408 ug/l

RT: 2.31 min Scan# 171

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :
MSVOA_F
ClientSampleId :
VSTDIC100





Tot Ion: 96 Resp: 200697

Ion Ratio Lower Upper

96 100

61 167.8 127.0 190.4

98 65.8 51.7 77.5



#22

Diisopropyl ether

Concen: 108.963 ug/l

RT: 2.80 min Scan# 220

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Tgt Ion: 45 Resp: 679261




Ion Ratio Lower Upper


45 100

43 45.2 37.4 56.2

87 17.0 14.2 21.2

59 9.0 8.2 12.2





#23

Vinyl Acetate

Concen: 422.288 ug/l

RT: 3.19 min Scan# 260

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :





VSTDIC100

Tot Ion: 43 Resp: 1205579

Ion Ratio Lower Upper

43 100

86 8.0 6.5 9.7



#24

1,1-Dichloroethane

Concen: 109.026 ug/l

RT: 2.88 min Scan# 228

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27




Tgt Ion: 63 Resp: 398479


Ion Ratio Lower Upper

63 100

98 4.7 2.5 7.6

100 3.1 1.7 5.0





#25

2-Butanone

Concen: 418.967 ug/l

RT: 4.35 min Scan# 377

Delta R.T. 0.01 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :





VSTDIC100

Tot Ion: 43 Resp: 372355

Ion Ratio Lower Upper

43 100

72 18.5 16.5 24.7



#26

2,2-Dichloropropane

Concen: 126.332 ug/l

RT: 3.62 min Scan# 303

Delta R.T. 0.01 min

Lab File: VF061289.D




Acq: 8 Jan 2019 16:27


Tgt Ion: 77 Resp: 265895

Ion Ratio Lower Upper

77 100

97 18.1 8.6 25.9









#27

cis-1,2-Dichloroethene
Concen: 102.861 ug/l
RT: 3.49 min Scan# 290
Delta R.T. 0.01 min
Lab File: VF061289.D
Acq: 8 Jan 2019 16:27

Instrument :
MSVOA_F
ClientSampleId :
VSTDIC100




Tot Ion	Ratio	Lower	Upper
96	100		
61	134.2	0.0	288.2
98	63.1	0.0	132.2




#28

Bromochloromethane
Concen: 90.176 ug/l
RT: 3.73 min Scan# 315
Delta R.T. 0.01 min
Lab File: VF061289.D
Acq: 8 Jan 2019 16:27

Tgt Ion	Ratio	Lower	Upper
49	100		
129	0.0	0.0	0.0
130	80.0	58.6	88.0





#29

Tetrahydrofuran

Concen: 448.537 ug/l

RT: 4.08 min Scan# 350

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC100





Tot Ion: 42 Resp: 154992

Ion Ratio Lower Upper

42 100

72 41.0 32.3 48.5

71 37.5 29.5 44.3



#30

Chloroform

Concen: 95.926 ug/l

RT: 3.87 min Scan# 329

Delta R.T. 0.00 min

Lab File: VF061289.D




Acq: 8 Jan 2019 16:27


Tgt Ion: 83 Resp: 490409

Ion Ratio Lower Upper

83 100

85 69.1 54.4 81.6





#31

Cyclohexane

Concen: 101.637 ug/l

RT: 3.71 min Scan# 313

Delta R.T. 0.01 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC100





Tot Ion: 56 Resp: 393728

Ion Ratio Lower Upper

56 100

69 33.0 24.7 37.1

84 89.3 72.2 108.4



#32

1,1,1-Trichloroethane

Concen: 121.712 ug/l

RT: 4.11 min Scan# 353

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27




Tgt Ion: 97 Resp: 404160


Ion Ratio Lower Upper

97 100

99 69.0 52.7 79.1

61 50.0 40.8 61.2





#33

1,2-Dichloroethane-d4

Concen: 91.298 ug/l

RT: 4.86 min Scan# 429

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC100

Tot Ion: 65 Resp: 215833

Ion Ratio Lower Upper




65 100

67 52.8 0.0 98.2



Abundance Ion 65.00 (64.70 to 65.70): VF0

Ion 67.00 (66.70 to 67.70): VF0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 5.61 min Scan# 505

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27


Tgt Ion: 114 Resp: 328018

Ion Ratio Lower Upper

114 100

63 21.5 0.0 42.8



88 19.8 0.0 40.0




Abundance Ion 114.00 (113.70 to 114.70): V

Ion 63.00 (62.70 to 63.70): VF0

Ion 88.00 (87.70 to 88.70): VF0





#35

Dibromofluoromethane

Concen: 95.237 ug/l

RT: 4.13 min Scan# 355

Delta R.T. 0.01 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC100


Tot Ion:113 Resp: 240340

Ion Ratio Lower Upper

113 100

111 101.4 78.0 117.0




192 13.6 10.6 16.0



Abundance Ion 113.00 (112.70 to 113.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 192.00 (191.70 to 192.70): V



#36

1,1-Dichloropropene

Concen: 92.979 ug/l

RT: 4.30 min Scan# 372

Delta R.T. 0.01 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27


Tgt Ion: 75 Resp: 359196

Ion Ratio Lower Upper

75 100

110 33.1 16.6 49.8



77 32.6 25.1 37.7




Abundance Ion 75.00 (74.70 to 75.70): VF0

Ion 110.00 (109.70 to 110.70): V

Ion 77.00 (76.70 to 77.70): VF0





#37

Ethyl Acetate

Concen: 86.048 ug/l

RT: 4.12 min Scan# 354

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC100


Tot Ion: 43 Resp: 132727

Ion Ratio Lower Upper

43 100

61 134.3 113.3 169.9




70 9.1 9.4 14.0#



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 61.00 (60.70 to 61.70): VF0

Ion 70.00 (69.70 to 70.70): VF0



#38

Carbon Tetrachloride

Concen: 134.507 ug/l

RT: 4.01 min Scan# 343

Delta R.T. 0.01 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27


Tgt Ion: 117 Resp: 416013

Ion Ratio Lower Upper

117 100

119 94.6 78.5 117.7



121 31.8 25.3 37.9




Abundance Ion 117.00 (116.70 to 117.70): V

Ion 119.00 (118.70 to 119.70): V

Ion 121.00 (120.70 to 121.70): V





#39

Methvlcvclohexane

Concen: 108.102 ug/l

RT: 5.47 min Scan# 491

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC100





Tot Ion: 83 Resp: 446417

Ion Ratio Lower Upper

83 100

55 85.3 66.4 99.6

98 48.6 39.1 58.7



#40

Benzene

Concen: 92.640 ug/l

RT: 4.64 min Scan# 407

Delta R.T. 0.00 min

Lab File: VF061289.D




Acq: 8 Jan 2019 16:27


Tgt Ion: 78 Resp: 812226

Ion Ratio Lower Upper

78 100

77 24.8 19.1 28.7





#41

Methacrylonitrile

Concen: 87.097 ug/l

RT: 4.75 min Scan# 418

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC100

Tot Ion: 41 Resp: 76391


Ion Ratio Lower Upper

41 100

39 57.2 52.7 79.1

67 46.2 41.4 62.2

52 43.7 33.0 49.4






Abundance Ion 41.00 (40.70 to 41.70): VF0

Ion 39.00 (38.70 to 39.70): VF0

Ion 67.00 (66.70 to 67.70): VF0

Ion 52.00 (51.70 to 52.70): VF0



#42

1,2-Dichloroethane

Concen: 95.711 ug/l

RT: 4.96 min Scan# 439

Delta R.T. 0.01 min

Lab File: VF061289.D


Acq: 8 Jan 2019 16:27

Tgt Ion: 62 Resp: 280767

Ion Ratio Lower Upper



62 100


98 8.8 0.0 16.8



Abundance Ion 62.00 (61.70 to 62.70): VF0

Ion 98.00 (97.70 to 98.70): VF0









#43

Isopropyl Acetate
 Concen: 96.945 ug/l
 RT: 6.97 min Scan# 643
 Delta R.T. 0.00 min
 Lab File: VF061289.D
 Acq: 8 Jan 2019 16:27

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDIC100




Tot Ion: 43 Resp: 199524
 Ion Ratio Lower Upper
 43 100
 61 32.0 24.3 36.5
 87 0.2 0.2 0.4#




#44

Trichloroethene
 Concen: 87.951 ug/l
 RT: 5.52 min Scan# 496
 Delta R.T. 0.00 min
 Lab File: VF061289.D
 Acq: 8 Jan 2019 16:27

Tgt Ion: 130 Resp: 240027
 Ion Ratio Lower Upper
 130 100
 95 97.1 0.0 184.4





#45

1,2-Dichloropropane

Concen: 104.287 ug/l

RT: 6.25 min Scan# 570

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :





VSTDIC100

Tot Ion: 63 Resp: 233438

Ion Ratio Lower Upper

63 100

65 32.4 25.1 37.7



#46

Dibromomethane

Concen: 99.922 ug/l

RT: 6.10 min Scan# 555

Delta R.T. 0.01 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27




Tgt Ion: 93 Resp: 147175


Ion Ratio Lower Upper

93 100

95 80.6 68.8 103.2

174 88.0 62.9 94.3





#47

Bromodichloromethane

Concen: 96.134 ug/l

RT: 6.40 min Scan# 585

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC100


Tot Ion: 83 Resp: 360435

Ion Ratio Lower Upper

83 100

85 69.6 50.4 75.6




127 8.0 6.4 9.6



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 85.00 (84.70 to 85.70): VF0

Ion 127.00 (126.70 to 127.70): V



#48

Methyl methacrylate

Concen: 97.741 ug/l

RT: 6.72 min Scan# 618

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27


Tgt Ion: 41 Resp: 130261

Ion Ratio Lower Upper

41 100

69 64.6 51.5 77.3



39 59.5 44.1 66.1




Abundance Ion 41.00 (40.70 to 41.70): VF0

Ion 69.00 (68.70 to 69.70): VF0

Ion 39.00 (38.70 to 39.70): VF0









#49
1,4-Dioxane
Concen: 1572.300 ug/l
RT: 6.71 min Scan# 617
Delta R.T. 0.00 min
Lab File: VF061289.D
Acq: 8 Jan 2019 16:27




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC100


Tot Ion: 88 Resp: 14519
Ion Ratio Lower Upper
88 100
43 52.0 38.5 57.7
58 60.8 57.0 85.4



#50
Toluene-d8
Concen: 96.187 ug/l
RT: 7.57 min Scan# 704
Delta R.T. 0.00 min
Lab File: VF061289.D
Acq: 8 Jan 2019 16:27

Tgt Ion: 98 Resp: 720325
Ion Ratio Lower Upper
98 100
100 71.1 55.6 83.4





#51

4-Methyl-2-Pentanone

Concen: 487.915 ug/l

RT: 8.27 min Scan# 775

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC100

Tot Ion: 43 Resp: 697449

Ion Ratio Lower Upper

43 100

58 38.7 29.6 44.4



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0

8.27



300000

200000

100000

0

Time--> 8.10 8.20 8.30 8.40



#52

Toluene

Concen: 92.960 ug/l

RT: 7.64 min Scan# 711

Delta R.T. 0.00 min

Lab File: VF061289.D


Acq: 8 Jan 2019 16:27

Tgt Ion: 92 Resp: 532183

Ion Ratio Lower Upper

92 100

91 161.2 129.2 193.8



Abundance Ion 92.00 (91.70 to 92.70): VF0

Ion 91.00 (90.70 to 91.70): VF0

7.64

400000


300000


200000

100000

0

Time--> 7.50 7.60 7.70 7.80





#53

t-1,3-Dichloropropene

Concen: 91.578 ug/l

RT: 8.29 min Scan# 777

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :





VSTDIC100

Tot Ion: 75 Resp: 279903

Ion Ratio Lower Upper

75 100

77 32.9 24.7 37.1



#54

cis-1,3-Dichloropropene

Concen: 94.890 ug/l

RT: 7.32 min Scan# 678

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27




Tgt Ion: 75 Resp: 368616


Ion Ratio Lower Upper

75 100

77 32.6 27.3 40.9

39 57.8 46.3 69.5





#55

1,1,2-Trichloroethane

Concen: 97.137 ug/l

RT: 8.50 min Scan# 798

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC100

Tot Ion: 97 Resp: 159921


Ion Ratio Lower Upper

97 100

83 86.0 67.8 101.6

85 59.4 45.1 67.7

99 59.5 51.0 76.6






Abundance Ion 97.00 (96.70 to 97.70): VF0

Ion 82.95 (82.65 to 83.65): VF0

Ion 84.95 (84.65 to 85.65): VF0

Ion 98.95 (98.65 to 99.65): VF0



#56

Ethyl methacrylate

Concen: 108.322 ug/l

RT: 8.63 min Scan# 811

Delta R.T. -0.01 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27


Tgt Ion: 69 Resp: 197090

Ion Ratio Lower Upper

69 100

41 76.9 59.0 88.4



39 44.8 35.3 52.9




Abundance Ion 69.00 (68.70 to 69.70): VF0

Ion 41.00 (40.70 to 41.70): VF0

Ion 39.00 (38.70 to 39.70): VF0





#57

1,3-Dichloropropane

Concen: 92.020 ug/l

RT: 8.86 min Scan# 835

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :





VSTDIC100

Tot Ion: 76 Resp: 284780

Ion Ratio Lower Upper

76 100

78 32.8 26.8 40.2



#58

2-Chloroethyl Vinyl ether

Concen: 177.289 ug/l

RT: 7.31 min Scan# 677

Delta R.T. 0.00 min

Lab File: VF061289.D




Acq: 8 Jan 2019 16:27


Tgt Ion: 63 Resp: 39173

Ion Ratio Lower Upper

63 100

106 26.4 16.5 24.7#









#59
2-Hexanone
Concen: 483.356 ug/l
RT: 9.50 min Scan# 900
Delta R.T. 0.00 min
Lab File: VF061289.D
Acq: 8 Jan 2019 16:27




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC100


Tot Ion: 43 Resp: 500205
Ion Ratio Lower Upper
43 100
58 52.7 25.6 76.8



#60
Dibromochloromethane
Concen: 97.146 ug/l
RT: 8.73 min Scan# 821
Delta R.T. 0.00 min
Lab File: VF061289.D
Acq: 8 Jan 2019 16:27

Tgt Ion: 129 Resp: 227178
Ion Ratio Lower Upper
129 100
127 76.5 39.4 118.1





#61

1,2-Dibromoethane

Concen: 98.588 ug/l

RT: 9.00 min Scan# 849

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC100

Tot Ion:107 Resp: 175172

Ion Ratio Lower Upper




107 100

109 100.9 75.1 112.7



Abundance Ion 107.00 (106.70 to 107.70): V

Ion 109.00 (108.70 to 109.70): V



#62

4-Bromofluorobenzene

Concen: 89.986 ug/l

RT: 11.42 min Scan# 1094

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27


Tgt Ion: 95 Resp: 298590

Ion Ratio Lower Upper

95 100

174 63.6 0.0 114.2



176 62.2 0.0 117.8




Abundance Ion 95.00 (94.70 to 95.70): VF0

Ion 174.00 (173.70 to 174.70): V

Ion 176.00 (175.70 to 176.70): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 9.79 min Scan# 929
Delta R.T. 0.00 min
Lab File: VF061289.D
Acq: 8 Jan 2019 16:27

Instrument :
MSVOA_F
ClientSampleId :
VSTDIC100

Tot Ion:117 Resp: 307713
Ion Ratio Lower Upper
117 100
82 55.2 45.3 67.9
119 31.2 25.4 38.0




Abundance Ion 117.00 (116.70 to 117.70): V
Ion 82.00 (81.70 to 82.70): V
Ion 119.00 (118.70 to 119.70): V






#64
Tetrachloroethene
Concen: 92.768 ug/l
RT: 8.16 min Scan# 764
Delta R.T. 0.00 min
Lab File: VF061289.D
Acq: 8 Jan 2019 16:27

Tgt Ion:164 Resp: 218763
Ion Ratio Lower Upper
164 100
166 130.7 100.1 150.1
129 101.3 80.1 120.1
131 103.0 76.7 115.1



Abundance Ion 163.85 (163.55 to 164.55): V
Ion 165.80 (165.50 to 166.50): V
Ion 128.90 (128.60 to 129.60): V
Ion 130.90 (130.60 to 131.60): V









#65
Chlorobenzene
Concen: 95.488 ug/l
RT: 9.81 min Scan# 931
Delta R.T. 0.00 min
Lab File: VF061289.D
Acq: 8 Jan 2019 16:27




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC100


Tot Ion:112 Resp: 597479
Ion Ratio Lower Upper
112 100
114 33.9 25.8 38.8



#66
1,1,1,2-Tetrachloroethane
Concen: 100.102 ug/l
RT: 9.95 min Scan# 945
Delta R.T. 0.00 min
Lab File: VF061289.D
Acq: 8 Jan 2019 16:27

Tgt Ion:131 Resp: 257223
Ion Ratio Lower Upper
131 100
133 92.8 49.9 149.6
119 65.4 33.9 101.6





#67

Ethyl Benzene

Concen: 97.252 ug/l

RT: 9.91 min Scan# 941

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC100

Tot Ion: 91 Resp: 1118625

Ion Ratio Lower Upper




91 100

106 30.4 21.4 32.2



Abundance Ion 91.00 (90.70 to 91.70): VF0

Ion 106.00 (105.70 to 106.70): V



#68

m/p-Xylenes

Concen: 185.877 ug/l

RT: 10.15 min Scan# 965

Delta R.T. 0.00 min

Lab File: VF061289.D


Acq: 8 Jan 2019 16:27

Tgt Ion: 106 Resp: 780250

Ion Ratio Lower Upper



106 100


91 192.0 165.0 247.6



Abundance Ion 106.00 (105.70 to 106.70): V

Ion 91.00 (90.70 to 91.70): VF0









#69
o-Xylene
Concen: 97.085 ug/l
RT: 10.73 min Scan# 1024
Delta R.T. 0.00 min
Lab File: VF061289.D
Acq: 8 Jan 2019 16:27




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC100


Tot Ion:106 Resp: 440856
Ion Ratio Lower Upper
106 100
91 209.6 108.1 324.4



#70
Styrene
Concen: 92.520 ug/l
RT: 10.80 min Scan# 1031
Delta R.T. 0.00 min
Lab File: VF061289.D
Acq: 8 Jan 2019 16:27

Tgt Ion:104 Resp: 578353
Ion Ratio Lower Upper
104 100
78 48.5 38.0 57.0
103 45.3 36.7 55.1





#71

Bromoform

Concen: 92.209 ug/l

RT: 10.79 min Scan# 1030

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC100


Tot Ion:173 Resp: 110311

Ion Ratio Lower Upper

173 100

175 48.7 24.1 72.2




254 8.2 0.0 0.0#



Abundance Ion 173.00 (172.70 to 173.70): V

Ion 175.00 (174.70 to 175.70): V

Ion 254.00 (253.70 to 254.70): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.56 min Scan# 1210

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27


Tgt Ion:152 Resp: 159767

Ion Ratio Lower Upper

152 100

115 54.6 29.6 88.9



150 168.9 0.0 315.4




Abundance Ion 152.00 (151.70 to 152.70): V

Ion 115.00 (114.70 to 115.70): V

Ion 150.00 (149.70 to 150.70): V





#73

Isopropylbenzene

Concen: 89.381 ug/l

RT: 11.13 min Scan# 1065

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :





VSTDIC100

Tot Ion:105 Resp: 1213009

Ion Ratio Lower Upper

105 100

120 27.2 12.9 38.7



#74

N-ethyl acetate

Concen: 89.928 ug/l

RT: 11.38 min Scan# 1090

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Tgt Ion: 43 Resp: 308124




Ion Ratio Lower Upper


43 100

70 34.7 27.4 41.0

55 20.4 14.7 22.1

61 24.7 19.0 28.4





#75

1,1,2,2-Tetrachloroethane

Concen: 92.060 ug/l

RT: 11.70 min Scan# 1123

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC100


Tot Ion: 83 Resp: 218227

Ion Ratio Lower Upper

83 100

131 8.6 3.9 11.7




85 64.5 31.6 94.7



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 131.00 (130.70 to 131.70): VF0

Ion 85.00 (84.70 to 85.70): VF0



#76

1,2,3-Trichloropropane

Concen: 88.805 ug/l

RT: 11.80 min Scan# 1133

Delta R.T. 0.00 min

Lab File: VF061289.D


Acq: 8 Jan 2019 16:27

Tgt Ion: 75 Resp: 156273

Ion Ratio Lower Upper



75 100


77 32.7 17.5 52.5



Abundance Ion 75.00 (74.70 to 75.70): VF0

Ion 77.00 (76.70 to 77.70): VF0





#77

Bromobenzene

Concen: 92.466 ug/l

RT: 11.51 min Scan# 1103

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC100


Tot Ion:156 Resp: 263435

Ion Ratio Lower Upper

156 100

77 159.1 84.3 252.9




158 96.3 47.1 141.3



Abundance Ion 156.00 (155.70 to 156.70): V

Ion 77.00 (76.70 to 77.70): VF0

Ion 158.00 (157.70 to 158.70): V



#78

n-propylbenzene

Concen: 86.847 ug/l

RT: 11.61 min Scan# 1113

Delta R.T. 0.00 min

Lab File: VF061289.D


Acq: 8 Jan 2019 16:27

Tgt Ion: 91 Resp: 1404858

Ion Ratio Lower Upper



91 100


120 23.8 10.7 32.1



Abundance Ion 91.00 (90.70 to 91.70): VF0

Ion 120.00 (119.70 to 120.70): V





#79

2-Chlorotoluene

Concen: 88.986 ug/l

RT: 11.73 min Scan# 1126

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :





VSTDIC100

Tot Ion: 91 Resp: 831330

Ion Ratio Lower Upper

91 100

126 35.5 16.6 49.8



#80

1,3,5-Trimethylbenzene

Concen: 88.399 ug/l

RT: 11.83 min Scan# 1136

Delta R.T. 0.00 min

Lab File: VF061289.D




Acq: 8 Jan 2019 16:27


Tgt Ion:105 Resp: 968509

Ion Ratio Lower Upper

105 100

120 54.1 25.2 75.6





#81

trans-1,4-Dichloro-2-butene

Concen: 116.111 ug/l

RT: 11.88 min Scan# 1141

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC100





Tot Ion: 75 Resp: 93648

Ion Ratio Lower Upper

75 100

53 72.3 62.2 93.2

89 51.1 40.2 60.4



#82

4-Chlorotoluene

Concen: 89.565 ug/l

RT: 11.91 min Scan# 1144

Delta R.T. 0.00 min

Lab File: VF061289.D




Acq: 8 Jan 2019 16:27


Tgt Ion: 91 Resp: 848228

Ion Ratio Lower Upper

91 100

126 34.5 16.4 49.4









#83
tert-Butylbenzene
Concen: 84.668 ug/l
RT: 12.14 min Scan# 1167
Delta R.T. 0.00 min
Lab File: VF061289.D
Acq: 8 Jan 2019 16:27




Instrument :
MSVOA_F
ClientSampleId :
VSTDIC100


Tot Ion:119 Resp: 956442
Ion Ratio Lower Upper
119 100
91 63.5 30.8 92.4
134 27.3 11.9 35.9



#84
1,2,4-Trimethylbenzene
Concen: 91.357 ug/l
RT: 12.22 min Scan# 1175
Delta R.T. 0.00 min
Lab File: VF061289.D
Acq: 8 Jan 2019 16:27

Tgt Ion:105 Resp: 996428
Ion Ratio Lower Upper
105 100
120 50.4 24.7 74.1





#85

sec-Butylbenzene

Concen: 89.638 ug/l

RT: 12.32 min Scan# 1185

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :


VSTDIC100

Tot Ion:105 Resp: 1373527

Ion Ratio Lower Upper




105 100

134 21.5 9.7 28.9



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V



#86

p-Isopropyltoluene

Concen: 87.751 ug/l

RT: 12.46 min Scan# 1200

Delta R.T. -0.01 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27


Tgt Ion:119 Resp: 1107888

Ion Ratio Lower Upper

119 100

134 27.3 13.0 39.0



91 25.3 12.2 36.4




Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): VF0









#87
 1,3-Dichlorobenzene
 Concen: 87.184 ug/l
 RT: 12.48 min Scan# 1202
 Delta R.T. 0.00 min
 Lab File: VF061289.D
 Acq: 8 Jan 2019 16:27




Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDIC100


Tot Ion:146 Resp: 479293
 Ion Ratio Lower Upper
 146 100
 111 41.1 21.1 63.3
 148 66.8 33.1 99.5



#88
 1,4-Dichlorobenzene
 Concen: 93.701 ug/l
 RT: 12.57 min Scan# 1211
 Delta R.T. 0.00 min
 Lab File: VF061289.D
 Acq: 8 Jan 2019 16:27

Tgt Ion:146 Resp: 502603
 Ion Ratio Lower Upper
 146 100
 111 41.2 21.4 64.2
 148 67.1 33.0 98.9









#89
n-Butylbenzene
Concen: 88.661 ug/l
RT: 12.85 min Scan# 1239
Delta R.T. -0.01 min
Lab File: VF061289.D
Acq: 8 Jan 2019 16:27

Instrument :
MSVOA_F
ClientSampleId :
VSTDIC100

Tot Ion: 91 Resp: 1150093
Ion Ratio Lower Upper
91 100
92 68.0 32.1 96.3
134 25.8 13.4 40.1




Abundance Ion 91.00 (90.70 to 91.70): VF0
Ion 92.00 (91.70 to 92.70): VF0
Ion 134.00 (133.70 to 134.70): V






#90
Hexachloroethane
Concen: 90.440 ug/l
RT: 12.93 min Scan# 1247
Delta R.T. 0.00 min
Lab File: VF061289.D
Acq: 8 Jan 2019 16:27

Tgt Ion: 117 Resp: 281707
Ion Ratio Lower Upper
117 100
201 64.4 27.9 83.7



Abundance Ion 117.00 (116.70 to 117.70): V
Ion 201.00 (200.70 to 201.70): V





#91

1,2-Dichlorobenzene

Concen: 89.230 ug/l

RT: 12.95 min Scan# 1249

Delta R.T. -0.01 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC100


Tot Ion:146 Resp: 473871

Ion Ratio Lower Upper

146 100

111 46.2 21.4 64.3




148 66.9 33.1 99.2



Abundance Ion 146.00 (145.70 to 146.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 148.00 (147.70 to 148.70): V



#92

1,2-Dibromo-3-Chloropropane

Concen: 102.995 ug/l

RT: 13.65 min Scan# 1320

Delta R.T. -0.01 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27


Tgt Ion: 75 Resp: 43875

Ion Ratio Lower Upper

75 100

155 72.0 51.3 154.0



157 91.7 60.7 182.0



Abundance Ion 75.00 (74.70 to 75.70): VF0

Ion 155.00 (154.70 to 155.70): V

Ion 157.00 (156.70 to 157.70): V









#93
 1,2,4-Trichlorobenzene
 Concen: 88.712 ug/l
 RT: 14.21 min Scan# 1377
 Delta R.T. 0.00 min
 Lab File: VF061289.D
 Acq: 8 Jan 2019 16:27

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDIC100

Tot Ion:180 Resp: 329896
 Ion Ratio Lower Upper
 180 100
 182 99.0 47.0 141.2
 145 35.6 19.7 59.1




Abundance Ion 180.00 (179.70 to 180.70): V
 Ion 182.00 (181.70 to 182.70): V
 Ion 145.00 (144.70 to 145.70): V






#94
 Hexachlorobutadiene
 Concen: 90.433 ug/l
 RT: 14.20 min Scan# 1376
 Delta R.T. 0.00 min
 Lab File: VF061289.D
 Acq: 8 Jan 2019 16:27

Tgt Ion:225 Resp: 214263
 Ion Ratio Lower Upper
 225 100
 223 60.8 31.8 95.4
 227 63.9 31.3 93.9



Abundance Ion 225.00 (224.70 to 225.70): V
 Ion 223.00 (222.70 to 223.70): V
 Ion 227.00 (226.70 to 227.70): V





#95

Naphthalene

Concen: 102.390 ug/l

RT: 14.47 min Scan# 1403

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27

Instrument :

MSVOA_F

ClientSampleId :

VSTDIC100


Tot Ion:128 Resp: 711713

Ion Ratio Lower Upper

128 100

127 11.9 9.4 14.0




129 11.7 8.8 13.2



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V



#96

1,2,3-Trichlorobenzene

Concen: 90.957 ug/l

RT: 14.61 min Scan# 1418

Delta R.T. 0.00 min

Lab File: VF061289.D

Acq: 8 Jan 2019 16:27


Tgt Ion:180 Resp: 313615

Ion Ratio Lower Upper

180 100

182 101.0 46.5 139.3

145 34.9 17.5 52.6



Abundance Ion 180.00 (179.70 to 180.70): V

Ion 182.00 (181.70 to 182.70): V

Ion 145.00 (144.70 to 145.70): V

