

Data Path : Z:\HPCHEM1\MSVOA_F\DATA\VF011019\
 Data File : VF061359.D
 Acq On : 10 Jan 2019 22:30
 Operator : VA/AP
 Sample : K1099-05
 Misc : 5.21g/5mL/MSVOA-F/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Quant Time: Jan 11 00:33:47 2019
 Quant Method : Z:\HPCHEM1\MSVOA_F\METHODS\82F010819S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 09 07:13:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.87	168	11861	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.60	114	15193	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.77	117	9245	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	12.55	152	2447	50.00	ug/l	0.00


System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.84	65	2833	21.46	ug/l	-0.01
Spiked Amount	50.000		Recovery	=	42.92%	
35) Dibromofluoromethane	4.11	113	4326	37.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	74.42%	
50) Toluene-d8	7.55	98	16541	48.08	ug/l	-0.01
Spiked Amount	50.000		Recovery	=	96.16%	
62) 4-Bromofluorobenzene	11.41	95	3552	23.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	47.30%	


Target Compounds					Qvalue
3) Chloromethane	1.08	50	366	1.877 ug/l #	40
5) Bromomethane	1.27	94	311	4.839 ug/l #	68
8) Diethyl Ether	1.67	74	212	4.891 ug/l	73
12) 1,1-Dichloroethene	1.76	96	136	1.091 ug/l #	1
13) Acrolein	2.00	56	88	Below Cal #	1
14) Allyl chloride	2.06	41	1039	4.561 ug/l #	52
15) Acrylonitrile	2.93	53	652	34.795 ug/l #	12
16) Acetone	2.28	43	187	6.089 ug/l	97
18) Methyl Acetate	2.35	43	240	3.972 ug/l #	65
20) Methylene Chloride	2.18	84	241	1.855 ug/l #	50
21) trans-1,2-Dichloroethene	2.30	96	151	1.143 ug/l #	1
22) Diisopropyl ether	2.80	45	515	1.168 ug/l #	1
23) Vinyl Acetate	3.17	43	491	3.809 ug/l #	77
25) 2-Butanone	4.35	43	441	8.922 ug/l #	56
29) Tetrahydrofuran	4.08	42	206	9.968 ug/l #	36
37) Ethyl Acetate	4.12	43	143	2.213 ug/l #	1
39) Methylcyclohexane	5.59	83	306	1.378 ug/l #	15
41) Methacrylonitrile	4.74	41	263	7.289 ug/l #	44
43) Isopropyl Acetate	6.96	43	122	1.430 ug/l #	45
45) 1,2-Dichloropropane	6.16	63	120	1.139 ug/l #	43
48) Methyl methacrylate	6.72	41	152	2.745 ug/l #	21
51) 4-Methyl-2-Pentanone	8.29	43	357	5.545 ug/l #	38
59) 2-Hexanone	9.48	43	77	1.688 ug/l #	27
74) N-ethyl acetate	11.36	43	60	1.226 ug/l #	48
76) 1,2,3-Trichloropropane	11.65	75	68	2.716 ug/l #	40
84) 1,2,4-Trimethylbenzene	12.20	105	544	3.198 ug/l #	28
85) sec-Butylbenzene	12.20	105	544	2.238 ug/l #	57

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\MSVOA_F\DATA\VF011019\
Data File : VF061359.D
Acq On : 10 Jan 2019 22:30
Operator : VA/AP
Sample : K1099-05
Misc : 5.21g/5mL/MSVOA-F/SOIL
ALS Vial : 23 Sample Multiplier: 1

Quant Time: Jan 11 00:33:47 2019
Quant Method : Z:\HPCHEM1\MSVOA_F\METHODS\82F010819S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 09 07:13:51 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.87 min Scan# 430

Delta R.T. -0.00 min

Lab File: VF061359.D


Acq: 10 Jan 2019 22:30

Tgt Ion:168 Resp: 11861

Ion Ratio Lower Upper




168 100

99 69.2 52.7 79.1



Abundance Ion 168.00 (167.70 to 168.70): VF0

Ion 99.00 (98.70 to 99.70): VF0



#3

Chloromethane

Concen: 1.877 ug/l

RT: 1.08 min Scan# 46

Delta R.T. 0.01 min

Lab File: VF061359.D


Acq: 10 Jan 2019 22:30

Tgt Ion: 50 Resp: 366

Ion Ratio Lower Upper



50 100


52 0.0 28.2 42.2#



Abundance Ion 50.00 (49.70 to 50.70): VF0

Ion 52.00 (51.70 to 52.70): VF0





#5

Bromomethane

Concen: 4.839 ug/l

RT: 1.27 min Scan# 65

Delta R.T. -0.01 min

Lab File: VF061359.D


Acq: 10 Jan 2019 22:30

Tgt Ion: 94 Resp: 311

Ion Ratio Lower Upper




94 100

96 58.3 70.2 105.4#



Abundance Ion 94.00 (93.70 to 94.70): VF0

Ion 96.00 (95.70 to 96.70): VF0



#8

Diethyl Ether

Concen: 4.891 ug/l

RT: 1.67 min Scan# 106

Delta R.T. 0.04 min

Lab File: VF061359.D


Acq: 10 Jan 2019 22:30

Tgt Ion: 74 Resp: 212

Ion Ratio Lower Upper



74 100


45 93.7 62.1 186.3



Abundance Ion 74.00 (73.70 to 74.70): VF0

Ion 45.00 (44.70 to 45.70): VF0





#12

1,1-Dichloroethene

Concen: 1.091 ug/l





RT: 1.76 min Scan# 115

Delta R.T. 0.01 min

Lab File: VF061359.D

Acq: 10 Jan 2019 22:30

Tgt Ion	Ratio	Lower	Upper
96	100		
61	0.0	171.8	257.8#
98	0.0	54.2	81.4#



#13

Acrolein

Concen: Below Cal




RT: 2.00 min Scan# 139


Delta R.T. -0.00 min

Lab File: VF061359.D

Acq: 10 Jan 2019 22:30





Tgt Ion	Ratio	Lower	Upper
56	100		
55	164.4	59.4	89.0#








#14
 Allyl chloride
 Concen: 4.561 ug/l
 RT: 2.06 min Scan# 145
 Delta R.T. -0.03 min
 Lab File: VF061359.D
 Acq: 10 Jan 2019 22:30


Tgt Ion: 41 Resp: 1039
 Ion Ratio Lower Upper
 41 100
 39 57.4 82.6 123.8#
 76 0.0 26.9 40.3#



#15
 Acrylonitrile
 Concen: 34.795 ug/l
 RT: 2.93 min Scan# 233
 Delta R.T. -0.01 min
 Lab File: VF061359.D
 Acq: 10 Jan 2019 22:30

Tgt Ion: 53 Resp: 652
 Ion Ratio Lower Upper
 53 100
 52 0.0 74.6 111.8#
 51 0.0 37.8 56.6#





#16

Acetone

Concen: 6.089 ug/l

RT: 2.28 min Scan# 168

Delta R.T. 0.05 min

Lab File: VF061359.D


Acq: 10 Jan 2019 22:30

Tgt Ion: 43 Resp: 187

Ion Ratio Lower Upper




43 100

58 22.1 16.6 25.0



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0



#18

Methyl Acetate

Concen: 3.972 ug/l

RT: 2.35 min Scan# 175

Delta R.T. 0.01 min

Lab File: VF061359.D


Acq: 10 Jan 2019 22:30

Tgt Ion: 43 Resp: 240

Ion Ratio Lower Upper



43 100


74 0.0 11.6 17.4#



Abundance Ion 43.00 (42.70 to 43.70): VF0





Ion 74.00 (73.70 to 74.70): VF0








#20
Methylene Chloride
Concen: 1.855 ug/l
RT: 2.18 min Scan# 157
Delta R.T. -0.00 min
Lab File: VF061359.D
Acq: 10 Jan 2019 22:30


Tgt Ion: 84 Resp: 241
Ion Ratio Lower Upper
84 100
49 108.9 116.6 175.0#
51 0.0 37.7 56.5#
86 0.0 49.8 74.8#



#21
trans-1,2-Dichloroethene
Concen: 1.143 ug/l
RT: 2.30 min Scan# 170
Delta R.T. -0.00 min
Lab File: VF061359.D
Acq: 10 Jan 2019 22:30

Tgt Ion: 96 Resp: 151
Ion Ratio Lower Upper
96 100
61 0.0 127.0 190.4#
98 0.0 51.7 77.5#





#22

Diisopropyl ether

Concen: 1.168 ug/l

RT: 2.80 min Scan# 220

Delta R.T. 0.01 min

Lab File: VF061359.D

Acq: 10 Jan 2019 22:30

Tgt Ion: 45 Resp: 515


Ion Ratio Lower Upper

45 100

43 281.6 37.4 56.2#

87 0.0 14.2 21.2#

59 0.0 8.2 12.2#






Abundance Ion 45.00 (44.70 to 45.70): VF0

Ion 43.00 (42.70 to 43.70): VF0

Ion 87.00 (86.70 to 87.70): VF0

Ion 59.00 (58.70 to 59.70): VF0



#23

Vinyl Acetate

Concen: 3.809 ug/l

RT: 3.17 min Scan# 258

Delta R.T. -0.01 min

Lab File: VF061359.D


Acq: 10 Jan 2019 22:30

Tgt Ion: 43 Resp: 491

Ion Ratio Lower Upper



43 100


86 0.0 6.5 9.7#



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 86.00 (85.70 to 86.70): VF0





#25

2-Butanone

Concen: 8.922 ug/l

RT: 4.35 min Scan# 377

Delta R.T. 0.01 min

Lab File: VF061359.D





Acq: 10 Jan 2019 22:30

Tgt Ion: 43 Resp: 441

Ion Ratio Lower Upper

43 100

72 0.0 16.5 24.7#



#29

Tetrahydrofuran

Concen: 9.968 ug/l

RT: 4.08 min Scan# 350

Delta R.T. 0.01 min

Lab File: VF061359.D

Acq: 10 Jan 2019 22:30




Tgt Ion: 42 Resp: 206


Ion Ratio Lower Upper

42 100

72 0.0 32.3 48.5#

71 0.0 29.5 44.3#





#33

1,2-Dichloroethane-d4

Concen: 21.459 ug/l

RT: 4.84 min Scan# 427

Delta R.T. -0.01 min

Lab File: VF061359.D


Acq: 10 Jan 2019 22:30

Tgt Ion: 65 Resp: 2833

Ion Ratio Lower Upper

65 100

67 43.2 0.0 98.2



Abundance Ion 65.00 (64.70 to 65.70): VF0

Ion 67.00 (66.70 to 67.70): VF0

4.84

800



600

400

200

0

Time--> 4.75 4.80 4.85 4.90 4.95



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 5.60 min Scan# 504

Delta R.T. -0.00 min

Lab File: VF061359.D

Acq: 10 Jan 2019 22:30


Tgt Ion: 114 Resp: 15193

Ion Ratio Lower Upper

114 100

63 24.5 0.0 42.8

88 20.4 0.0 40.0



Abundance Ion 114.00 (113.70 to 114.70): V

Ion 63.00 (62.70 to 63.70): VF0

Ion 88.00 (87.70 to 88.70): VF0

8000


6000


4000

2000

0

Time--> 5.50 5.55 5.60 5.65 5.70





#35

Dibromofluoromethane

Concen: 37.212 ug/l

RT: 4.11 min Scan# 353

Delta R.T. -0.00 min

Lab File: VF061359.D

Acq: 10 Jan 2019 22:30


Tgt Ion: 113 Resp: 4326

Ion Ratio Lower Upper

113 100

111 111.6 78.0 117.0




192 18.5 10.6 16.0#



Abundance Ion 113.00 (112.70 to 113.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 192.00 (191.70 to 192.70): V



#37

Ethyl Acetate

Concen: 2.213 ug/l

RT: 4.12 min Scan# 354

Delta R.T. 0.01 min

Lab File: VF061359.D

Acq: 10 Jan 2019 22:30


Tgt Ion: 43 Resp: 143

Ion Ratio Lower Upper

43 100

61 0.0 113.3 169.9#



70 0.0 9.4 14.0#




Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 61.00 (60.70 to 61.70): VF0


Ion 70.00 (69.70 to 70.70): VF0








#39
Methylcyclohexane
Concen: 1.378 ug/l
RT: 5.59 min Scan# 503
Delta R.T. 0.12 min
Lab File: VF061359.D
Acq: 10 Jan 2019 22:30

Tgt Ion	Ratio	Lower	Upper
83	100		
55	0.0	66.4	99.6#
98	0.0	39.1	58.7#




Abundance Ion 83.00 (82.70 to 83.70): VF0
Ion 55.00 (54.70 to 55.70): VF0
Ion 98.00 (97.70 to 98.70): VF0






#41
Methacrylonitrile
Concen: 7.289 ug/l
RT: 4.74 min Scan# 417
Delta R.T. -0.00 min
Lab File: VF061359.D
Acq: 10 Jan 2019 22:30

Tgt Ion	Ratio	Lower	Upper
41	100		
39	94.0	52.7	79.1#
67	0.0	41.4	62.2#
52	0.0	33.0	49.4#



Abundance Ion 41.00 (40.70 to 41.70): VF0
Ion 39.00 (38.70 to 39.70): VF0
Ion 67.00 (66.70 to 67.70): VF0
Ion 52.00 (51.70 to 52.70): VF0





#43

Isopropyl Acetate

Concen: 1.430 ug/l


RT: 6.96 min Scan# 642

Delta R.T. -0.00 min




Lab File: VF061359.D

Acq: 10 Jan 2019 22:30

Tgt Ion:	43	Resp:	122
Ion	Ratio	Lower	Upper
43	100		
61	0.0	24.3	36.5#
87	0.0	0.2	0.4#



Abundance Ion 43.00 (42.70 to 43.70): VF0
Ion 61.00 (60.70 to 61.70): VF0
Ion 87.00 (86.70 to 87.70): VF0



#45

1,2-Dichloropropane

Concen: 1.139 ug/l


RT: 6.16 min Scan# 561

Delta R.T. -0.08 min



Lab File: VF061359.D


Acq: 10 Jan 2019 22:30

Tgt Ion:	63	Resp:	120
Ion	Ratio	Lower	Upper
63	100		
65	0.0	25.1	37.7#



Abundance Ion 63.00 (62.70 to 63.70): VF0
Ion 65.00 (64.70 to 65.70): VF0





#48

Methyl methacrylate

Concen: 2.745 ug/l


RT: 6.72 min Scan# 618

Delta R.T. 0.01 min

Lab File: VF061359.D

Acq: 10 Jan 2019 22:30




Tgt Ion:	41	Resp:	152
Ion	Ratio	Lower	Upper
41	100		
69	0.0	51.5	77.3#
39	0.0	44.1	66.1#



Abundance Ion 41.00 (40.70 to 41.70): VF0

Ion 69.00 (68.70 to 69.70): VF0

Ion 39.00 (38.70 to 39.70): VF0



#50

Toluene-d8

Concen: 48.082 ug/l


RT: 7.55 min Scan# 702

Delta R.T. -0.01 min

Lab File: VF061359.D



Acq: 10 Jan 2019 22:30


Tgt Ion:	98	Resp:	16541
Ion	Ratio	Lower	Upper
98	100		
100	65.6	55.6	83.4



Abundance Ion 98.00 (97.70 to 98.70): VF0

Ion 100.00 (99.70 to 100.70): VF0





#51

4-Methyl-2-Pentanone

Concen: 5.545 ug/l

RT: 8.29 min Scan# 777

Delta R.T. 0.02 min

Lab File: VF061359.D


Acq: 10 Jan 2019 22:30

Tgt Ion: 43 Resp: 357

Ion Ratio Lower Upper




43 100

58 0.0 29.6 44.4#



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0



#59

2-Hexanone

Concen: 1.688 ug/l

RT: 9.48 min Scan# 897

Delta R.T. -0.02 min

Lab File: VF061359.D


Acq: 10 Jan 2019 22:30

Tgt Ion: 43 Resp: 77

Ion Ratio Lower Upper



43 100


58 0.0 25.6 76.8#



Abundance Ion 43.00 (42.70 to 43.70): VF0





Ion 58.00 (57.70 to 58.70): VF0








#62
4-Bromofluorobenzene
Concen: 23.646 ug/l
RT: 11.41 min Scan# 1093
Delta R.T. -0.00 min
Lab File: VF061359.D
Acq: 10 Jan 2019 22:30


Tgt Ion: 95 Resp: 3552
Ion Ratio Lower Upper
95 100
174 47.5 0.0 114.2
176 40.6 0.0 117.8



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 9.77 min Scan# 927
Delta R.T. -0.01 min
Lab File: VF061359.D
Acq: 10 Jan 2019 22:30

Tgt Ion: 117 Resp: 9245
Ion Ratio Lower Upper
117 100
82 62.2 45.3 67.9
119 34.2 25.4 38.0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.55 min Scan# 1209

Delta R.T. -0.00 min

Lab File: VF061359.D

Acq: 10 Jan 2019 22:30


Tgt Ion: 152 Resp: 2447

Ion Ratio Lower Upper

152 100

115 58.5 29.6 88.9




150 175.2 0.0 315.4



Abundance Ion 152.00 (151.70 to 152.70): V

Ion 115.00 (114.70 to 115.70): V

Ion 150.00 (149.70 to 150.70): V



#74

N-ethyl acetate

Concen: 1.226 ug/l

RT: 11.36 min Scan# 1088

Delta R.T. -0.01 min

Lab File: VF061359.D

Acq: 10 Jan 2019 22:30

Tgt Ion: 43 Resp: 60


Ion Ratio Lower Upper

43 100

70 0.0 27.4 41.0#

55 0.0 14.7 22.1#

61 0.0 19.0 28.4#






Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 70.00 (69.70 to 70.70): VF0

Ion 55.00 (54.70 to 55.70): VF0

Ion 61.00 (60.70 to 61.70): VF0









#76

1,2,3-Trichloropropane
Concen: 2.716 ug/l
RT: 11.65 min Scan# 1117
Delta R.T. -0.15 min
Lab File: VF061359.D
Acq: 10 Jan 2019 22:30

Tgt Ion: 75 Resp: 68
Ion Ratio Lower Upper
75 100
77 0.0 17.5 52.5#




Abundance Ion 75.00 (74.70 to 75.70): VF0
Ion 77.00 (76.70 to 77.70): VF0





#84


1,2,4-Trimethylbenzene
Concen: 3.198 ug/l
RT: 12.20 min Scan# 1173
Delta R.T. -0.01 min
Lab File: VF061359.D
Acq: 10 Jan 2019 22:30

Tgt Ion: 105 Resp: 544
Ion Ratio Lower Upper
105 100
120 0.0 24.7 74.1#



Abundance Ion 105.00 (104.70 to 105.70): V
Ion 120.00 (119.70 to 120.70): V





#85

sec-Butylbenzene

Concen: 2.238 ug/l

RT: 12.20 min Scan# 1173

Delta R.T. -0.11 min

Lab File: VF061359.D

Acq: 10 Jan 2019 22:30

Tgt Ion: 105 Resp: 544

Ion Ratio Lower Upper

105 100

134 0.0 9.7 28.9#

