

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF011019\
 Data File : VF061345.D
 Acq On : 10 Jan 2019 15:48
 Operator : VA/AP
 Sample : K1006-13RE
 Misc : 5.92g/5mL/MSVOA-F/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 JC-03-010819-GRE

Manual Integrations
 APPROVED

MMDadoda
 1/11/2019 3:52:26 PM

Quant Time: Jan 11 00:49:14 2019
 Quant Method : Z:\HPCHEM1\MSVOA_F\METHODS\82F010819S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 09 07:13:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.88	168	180012	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.61	114	320100	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.78	117	242941	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.55	152	98927	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.85	65	122844	61.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.62%	
35) Dibromofluoromethane	4.12	113	132808	54.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.44%	
50) Toluene-d8	7.56	98	292590	40.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.74%	
62) 4-Bromofluorobenzene	11.42	95	124238	39.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.50%	
Target Compounds						
16) Acetone	2.23	43	29666	63.648	ug/l #	87
20) Methylene Chloride	2.19	84	5379m	2.728	ug/l	
84) 1,2,4-Trimethylbenzene	12.21	105	10263	1.492	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF011019\
Data File : VF061345.D
Acq On : 10 Jan 2019 15:48
Operator : VA/AP
Sample : K1006-13RE
Misc : 5.92g/5mL/MSVOA-F/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
JC-03-010819-GRE

Manual Integrations
APPROVED

MMDadoda
1/11/2019 3:52:26 PM

Quant Time: Jan 11 00:49:14 2019
Quant Method : Z:\HPCHEM1\MSVOA_F\METHODS\82F010819S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 09 07:13:51 2019
Response via : Initial Calibration

