

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF011119\
 Data File : VF061374.D
 Acq On : 11 Jan 2019 18:06
 Operator : VA/AP
 Sample : K1118-05
 Misc : 5.62g/5mL/MSVOA-F/SOIL
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 TESTPIT-SOIL-1

Quant Time: Jan 12 04:20:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F010819S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 09 07:13:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.85	168	62488	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	5.59	114	99781	50.00	ug/l	-0.02
63) Chlorobenzene-d5	9.76	117	76561	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	12.54	152	27713	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.83	65	19465	27.99	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	55.98%	
35) Dibromofluoromethane	4.10	113	29655	38.84	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	77.68%	
50) Toluene-d8	7.54	98	90125	39.89	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	79.78%	
62) 4-Bromofluorobenzene	11.40	95	30468	30.88	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	61.76%	
Target Compounds						
16) Acetone	2.22	43	9335	57.696	ug/l	97
20) Methylene Chloride	2.17	84	4005	5.851	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF011119\
Data File : VF061374.D
Acq On : 11 Jan 2019 18:06
Operator : VA/AP
Sample : K1118-05
Misc : 5.62g/5mL/MSVOA-F/SOIL
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
TESTPIT-SOIL-1

Quant Time: Jan 12 04:20:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F010819S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 09 07:13:51 2019
Response via : Initial Calibration

