

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF011619\
 Data File : VF061415.D
 Acq On : 16 Jan 2019 16:49
 Operator : VA/AP
 Sample : VF0116SBS01
 Misc : 5.00µg/5mL/MSVOA-F/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VF0116SBS01

Quant Time: Jan 17 02:54:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 04:51:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.86	168	216223	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.58	114	370515	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.77	117	327085	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.55	152	167071	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	4.83	65	118356	44.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.68%	
35) Dibromofluoromethane	4.10	113	141219	48.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.60%	
50) Toluene-d8	7.55	98	396167	49.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.92%	
62) 4-Bromofluorobenzene	11.40	95	168116	44.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.84%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	0.96	85	74834	17.614	ug/l	88
3) Chloromethane	1.09	50	57971	18.609	ug/l	97
4) Vinyl Chloride	1.13	62	53176	19.190	ug/l	96
5) Bromomethane	1.30	94	35912	19.874	ug/l	87
6) Chloroethane	1.37	64	22777	18.550	ug/l	89
7) Trichlorofluoromethane	1.47	101	89259	16.569	ug/l	94
8) Diethyl Ether	1.63	74	12210	17.689	ug/l	92
9) 1,1,2-Trichlorotrifluoroet	1.82	101	50177	19.385	ug/l	98
10) Methyl Iodide	1.88	142	51108	12.736	ug/l	100
11) Tert butyl alcohol	2.57	59	11054	97.825	ug/l #	89
12) 1,1-Dichloroethene	1.79	96	38160	19.843	ug/l	90
13) Acrolein	2.00	56	7922	88.365	ug/l	97
14) Allyl chloride	2.09	41	63908	29.913	ug/l	87
15) Acrylonitrile	2.93	53	28284	97.719	ug/l	99
16) Acetone	2.24	43	60423	112.826	ug/l #	91
17) Carbon Disulfide	1.80	76	116584	19.443	ug/l #	92
18) Methyl Acetate	2.34	43	18319	17.248	ug/l	95
19) Methyl tert-butyl Ether	2.43	73	74068	17.181	ug/l	96
20) Methylene Chloride	2.18	84	16791	7.751	ug/l	88
21) trans-1,2-Dichloroethene	2.31	96	39235	18.543	ug/l	95
22) Diisopropyl ether	2.79	45	137951	20.158	ug/l	94
23) Vinyl Acetate	3.18	43	342832	115.631	ug/l	99
24) 1,1-Dichloroethane	2.87	63	82883	19.297	ug/l	98
25) 2-Butanone	4.32	43	98122	110.177	ug/l	94
26) 2,2-Dichloropropane	3.59	77	50869	19.276	ug/l	96
27) cis-1,2-Dichloroethene	3.47	96	59912	19.182	ug/l	96
28) Bromochloromethane	3.71	49	35606	18.442	ug/l #	97
29) Tetrahydrofuran	4.07	42	38662	100.819	ug/l	97
30) Chloroform	3.86	83	109961	18.970	ug/l	96
31) Cyclohexane	3.69	56	92541	21.569	ug/l	92
32) 1,1,1-Trichloroethane	4.08	97	83560	19.660	ug/l	93
36) 1,1-Dichloropropene	4.27	75	85600	19.926	ug/l	98
37) Ethyl Acetate	4.10	43	35112	20.447	ug/l #	74
38) Carbon Tetrachloride	3.99	117	71106	18.933	ug/l	92

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF011619\
 Data File : VF061415.D
 Acq On : 16 Jan 2019 16:49
 Operator : VA/AP
 Sample : VF0116SBS01
 Misc : 5.00µ/5mL/MSVOA-F/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VF0116SBS01

Quant Time: Jan 17 02:54:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 04:51:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	5.45	83	94187	20.231	ug/l	92
40) Benzene	4.62	78	206106	21.561	ug/l	97
41) Methacrylonitrile	4.74	41	18881	20.663	ug/l #	85
42) 1,2-Dichloroethane	4.93	62	58872	17.283	ug/l	96
43) Isopropyl Acetate	6.95	43	41443	19.229	ug/l	92
44) Trichloroethene	5.49	130	59469	20.280	ug/l	92
45) 1,2-Dichloropropane	6.22	63	52648	20.985	ug/l	97
46) Dibromomethane	6.08	93	31373	18.806	ug/l	92
47) Bromodichloromethane	6.38	83	76739	18.592	ug/l	94
48) Methyl methacrylate	6.71	41	28836	19.925	ug/l #	89
49) 1,4-Dioxane	6.71	88	4508	404.927	ug/l #	86
51) 4-Methyl-2-Pentanone	8.25	43	165385	104.381	ug/l	98
52) Toluene	7.62	92	121290	19.267	ug/l	94
53) t-1,3-Dichloropropene	8.27	75	67824	20.694	ug/l	97
54) cis-1,3-Dichloropropene	7.29	75	87368	21.326	ug/l	93
55) 1,1,2-Trichloroethane	8.48	97	36352	20.335	ug/l	91
56) Ethyl methacrylate	8.61	69	41421	19.916	ug/l	91
57) 1,3-Dichloropropane	8.85	76	63829	19.050	ug/l	98
58) 2-Chloroethyl Vinyl ether	7.29	63	21539	115.409	ug/l	90
59) 2-Hexanone	9.49	43	126472	114.469	ug/l	99
60) Dibromochloromethane	8.71	129	50158	19.193	ug/l	100
61) 1,2-Dibromoethane	8.98	107	37956	18.547	ug/l	84
64) Tetrachloroethene	8.15	164	49489	20.120	ug/l	96
65) Chlorobenzene	9.79	112	136770	20.562	ug/l	99
66) 1,1,1,2-Tetrachloroethane	9.92	131	56082	21.165	ug/l	98
67) Ethyl Benzene	9.89	91	259691	21.369	ug/l	98
68) m/p-Xylenes	10.12	106	186102	42.577	ug/l	97
69) o-Xylene	10.70	106	98588	20.811	ug/l	97
70) Styrene	10.78	104	138008	21.283	ug/l	97
71) Bromoform	10.76	173	25227	21.189	ug/l	96
73) Isopropylbenzene	11.12	105	302352	22.251	ug/l	100
74) N-amyl acetate	11.36	43	73902	23.000	ug/l	95
75) 1,1,2,2-Tetrachloroethane	11.69	83	51707	21.360	ug/l	98
76) 1,2,3-Trichloropropane	11.79	75	36681	21.088	ug/l	100
77) Bromobenzene	11.49	156	61614	21.471	ug/l	97
78) n-propylbenzene	11.59	91	347575	19.617	ug/l	99
79) 2-Chlorotoluene	11.71	91	202324	21.460	ug/l	99
80) 1,3,5-Trimethylbenzene	11.82	105	241842	22.054	ug/l	99
81) trans-1,4-Dichloro-2-buten	11.87	75	23164	30.635	ug/l	95
82) 4-Chlorotoluene	11.90	91	204118	21.455	ug/l	97
83) tert-Butylbenzene	12.12	119	237450	20.921	ug/l	99
84) 1,2,4-Trimethylbenzene	12.20	105	239997	21.247	ug/l	96
85) sec-Butylbenzene	12.30	105	340159	22.069	ug/l	95
86) p-Isopropyltoluene	12.46	119	280908	22.201	ug/l	98
87) 1,3-Dichlorobenzene	12.47	146	120563	19.995	ug/l	98
88) 1,4-Dichlorobenzene	12.56	146	113947	21.013	ug/l	97
89) n-Butylbenzene	12.84	91	287112	22.003	ug/l	97
90) Hexachloroethane	12.91	117	63183	21.184	ug/l	97
91) 1,2-Dichlorobenzene	12.94	146	111798	20.826	ug/l	97
92) 1,2-Dibromo-3-Chloropropan	13.63	75	8290	19.308	ug/l	64

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF011619\
Data File : VF061415.D
Acq On : 16 Jan 2019 16:49
Operator : VA/AP
Sample : VF0116SBS01
Misc : 5.00µ/5mL/MSVOA-F/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
VF0116SBS01

Quant Time: Jan 17 02:54:28 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 04:51:59 2019
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	14.20	180	78303	20.822	ug/l	97
94) Hexachlorobutadiene	14.19	225	47266	20.255	ug/l	98
95) Naphthalene	14.45	128	133660	18.707	ug/l	97
96) 1,2,3-Trichlorobenzene	14.60	180	68780	19.757	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF011619\
 Data File : VF061415.D
 Acq On : 16 Jan 2019 16:49
 Operator : VA/AP
 Sample : VF0116SBS01
 Misc : 5.00µ/5mL/MSVOA-F/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VF0116SBS01

Quant Time: Jan 17 02:54:28 2019
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 04:51:59 2019
 Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.86 min Scan# 429

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :


VF0116SBS01

Tot Ion:168 Resp: 216223

Ion Ratio Lower Upper

168 100

99 63.6 53.5 80.3



Abundance Ion 168.00 (167.70 to 168.70): VF0

Ion 99.00 (98.70 to 99.70): VF0

4.86



60000

40000

20000

0

Time--> 4.70 4.80 4.90 5.00 5.10



#2

Dichlorodifluoromethane

Concen: 17.614 ug/l

RT: 0.96 min Scan# 33

Delta R.T. 0.00 min

Lab File: VF061415.D


Acq: 16 Jan 2019 16:49

Tgt Ion: 85 Resp: 74834

Ion Ratio Lower Upper

85 100

87 29.6 18.4 55.0



Abundance Ion 85.00 (84.70 to 85.70): VF0

Ion 87.00 (86.70 to 87.70): VF0

0.96

25000

20000


15000


10000

5000

0

Time--> 0.80 0.90 1.00 1.10





#3

Chloromethane

Concen: 18.609 ug/l

RT: 1.09 min Scan# 46

Delta R.T. -0.01 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :


VF0116SBS01

Tot Ion: 50 Resp: 57971

Ion Ratio Lower Upper




50 100

52 32.1 24.6 36.8



Abundance Ion 50.00 (49.70 to 50.70): VF0

Ion 52.00 (51.70 to 52.70): VF0



#4

Vinyl Chloride

Concen: 19.190 ug/l

RT: 1.13 min Scan# 51

Delta R.T. 0.00 min

Lab File: VF061415.D


Acq: 16 Jan 2019 16:49

Tgt Ion: 62 Resp: 53176

Ion Ratio Lower Upper



62 100


64 34.1 25.6 38.4



Abundance Ion 62.00 (61.70 to 62.70): VF0

Ion 64.00 (63.70 to 64.70): VF0





#5

Bromomethane

Concen: 19.874 ug/l

RT: 1.30 min Scan# 68

Delta R.T. -0.03 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :


VF0116SBS01

Tot Ion: 94 Resp: 35912

Ion Ratio Lower Upper




94 100

96 82.6 76.0 114.0



Abundance Ion 94.00 (93.70 to 94.70): VF0

Ion 96.00 (95.70 to 96.70): VF0



#6

Chloroethane

Concen: 18.550 ug/l

RT: 1.37 min Scan# 75

Delta R.T. -0.02 min

Lab File: VF061415.D


Acq: 16 Jan 2019 16:49

Tgt Ion: 64 Resp: 22777

Ion Ratio Lower Upper



64 100


66 26.8 26.6 40.0



Abundance Ion 64.00 (63.70 to 64.70): VF0

Ion 66.00 (65.70 to 66.70): VF0









#7

Trichlorofluoromethane
 Concen: 16.569 ug/l
 RT: 1.47 min Scan# 85
 Delta R.T. -0.01 min
 Lab File: VF061415.D
 Acq: 16 Jan 2019 16:49

Instrument :
 MSVOA_F
 ClientSampleId :
 VF0116SBS01




Tot Ion:101 Resp: 89259
 Ion Ratio Lower Upper
 101 100
 103 66.6 49.7 74.5




#8

Diethyl Ether
 Concen: 17.689 ug/l
 RT: 1.63 min Scan# 101
 Delta R.T. 0.00 min
 Lab File: VF061415.D
 Acq: 16 Jan 2019 16:49

Tgt Ion: 74 Resp: 12210
 Ion Ratio Lower Upper
 74 100
 45 140.7 65.8 197.5





#9

1,1,2-Trichlorotrifluoroethane

Concen: 19.385 ug/l

RT: 1.82 min Scan# 120

Delta R.T. -0.02 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :

VF0116SBS01


Tot Ion:101 Resp: 50177

Ion Ratio Lower Upper

101 100

85 50.1 41.9 62.9




151 69.3 55.1 82.7



Abundance Ion 101.00 (100.70 to 101.70): V

Ion 85.00 (84.70 to 85.70): VF0

Ion 151.00 (150.70 to 151.70): V



#10

Methyl Iodide

Concen: 12.736 ug/l

RT: 1.88 min Scan# 127

Delta R.T. -0.03 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49


Tgt Ion:142 Resp: 51108

Ion Ratio Lower Upper

142 100

127 63.5 50.7 76.1



141 15.0 11.8 17.6




Abundance Ion 142.00 (141.70 to 142.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 141.00 (140.70 to 141.70): V









#11

Tert butyl alcohol
 Concen: 97.825 ug/l
 RT: 2.57 min Scan# 196
 Delta R.T. 0.01 min
 Lab File: VF061415.D
 Acq: 16 Jan 2019 16:49

Instrument :
 MSVOA_F
 ClientSampleId :
 VF0116SBS01




Tot Ion: 59 Resp: 11054
 Ion Ratio Lower Upper
 59 100
 57 10.5 11.9 17.9#




#12

1,1-Dichloroethene
 Concen: 19.843 ug/l
 RT: 1.79 min Scan# 117
 Delta R.T. -0.01 min
 Lab File: VF061415.D
 Acq: 16 Jan 2019 16:49

Tgt Ion: 96 Resp: 38160
 Ion Ratio Lower Upper
 96 100
 61 212.9 186.9 280.3
 98 66.7 53.0 79.4









#13
Acrolein
Concen: 88.365 ug/l
RT: 2.00 min Scan# 139
Delta R.T. 0.01 min
Lab File: VF061415.D
Acq: 16 Jan 2019 16:49




Instrument :
MSVOA_F
ClientSampleId :
VF0116SBS01


Tot Ion: 56 Resp: 7922
Ion Ratio Lower Upper
56 100
55 69.5 57.5 86.3



#14
Allyl chloride
Concen: 29.913 ug/l
RT: 2.09 min Scan# 148
Delta R.T. 0.00 min
Lab File: VF061415.D
Acq: 16 Jan 2019 16:49

Tgt Ion: 41 Resp: 63908
Ion Ratio Lower Upper
41 100
39 102.0 94.6 141.8
76 28.0 25.9 38.9









#15
 Acrylonitrile
 Concen: 97.719 ug/l
 RT: 2.93 min Scan# 233
 Delta R.T. 0.00 min
 Lab File: VF061415.D
 Acq: 16 Jan 2019 16:49




Instrument :
 MSVOA_F
 ClientSampleId :
 VF0116SBS01


Tot Ion: 53 Resp: 28284
 Ion Ratio Lower Upper
 53 100
 52 86.1 69.6 104.4
 51 47.5 36.8 55.2



#16
 Acetone
 Concen: 112.826 ug/l
 RT: 2.24 min Scan# 163
 Delta R.T. 0.01 min
 Lab File: VF061415.D
 Acq: 16 Jan 2019 16:49

Tgt Ion: 43 Resp: 60423
 Ion Ratio Lower Upper
 43 100
 58 20.5 13.3 19.9#









#17
Carbon Disulfide
Concen: 19.443 ug/l
RT: 1.80 min Scan# 118
Delta R.T. -0.03 min
Lab File: VF061415.D
Acq: 16 Jan 2019 16:49




Instrument :
MSVOA_F
ClientSampleId :
VF0116SBS01


Tot Ion: 76 Resp: 116584
Ion Ratio Lower Upper
76 100
78 8.8 9.6 14.4#



#18
Methyl Acetate
Concen: 17.248 ug/l
RT: 2.34 min Scan# 173
Delta R.T. 0.00 min
Lab File: VF061415.D
Acq: 16 Jan 2019 16:49

Tgt Ion: 43 Resp: 18319
Ion Ratio Lower Upper
43 100
74 18.8 13.4 20.0









#19
Methyl tert-butyl Ether
Concen: 17.181 ug/l
RT: 2.43 min Scan# 182
Delta R.T. 0.00 min
Lab File: VF061415.D
Acq: 16 Jan 2019 16:49




Instrument :
MSVOA_F
ClientSampleId :
VF0116SBS01


Tot Ion: 73 Resp: 74068
Ion Ratio Lower Upper
73 100
57 20.5 14.8 22.2



#20
Methylene Chloride
Concen: 7.751 ug/l
RT: 2.18 min Scan# 157
Delta R.T. 0.01 min
Lab File: VF061415.D
Acq: 16 Jan 2019 16:49

Tgt Ion: 84 Resp: 16791
Ion Ratio Lower Upper
84 100
49 168.0 120.0 180.0
51 41.7 38.6 58.0
86 68.5 61.0 91.6





#21

trans-1,2-Dichloroethene

Concen: 18.543 ug/l

RT: 2.31 min Scan# 170

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :

VF0116SBS01





Tot Ion: 96 Resp: 39235

Ion Ratio Lower Upper

96 100

61 167.7 131.8 197.8

98 69.7 47.4 71.0



#22

Diisopropyl ether

Concen: 20.158 ug/l

RT: 2.79 min Scan# 219

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Tgt Ion: 45 Resp: 137951




Ion Ratio Lower Upper


45 100

43 53.7 39.0 58.6

87 20.3 13.9 20.9

59 9.4 7.8 11.6





#23

Vinyl Acetate

Concen: 115.631 ug/l

RT: 3.18 min Scan# 258

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :





VF0116SBS01

Tot Ion: 43 Resp: 342832

Ion Ratio Lower Upper

43 100

86 8.0 6.2 9.4



#24

1,1-Dichloroethane

Concen: 19.297 ug/l

RT: 2.87 min Scan# 227

Delta R.T. 0.01 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49




Tgt Ion: 63 Resp: 82883


Ion Ratio Lower Upper

63 100

98 5.7 2.9 8.8

100 1.9 1.6 4.8





#25

2-Butanone

Concen: 110.177 ug/l

RT: 4.32 min Scan# 374

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :





VF0116SBS01

Tot Ion: 43 Resp: 98122

Ion Ratio Lower Upper

43 100

72 17.1 15.8 23.6



#26

2,2-Dichloropropane

Concen: 19.276 ug/l

RT: 3.59 min Scan# 300

Delta R.T. 0.00 min

Lab File: VF061415.D




Acq: 16 Jan 2019 16:49


Tgt Ion: 77 Resp: 50869

Ion Ratio Lower Upper

77 100

97 18.6 8.5 25.5









#27

cis-1,2-Dichloroethene
Concen: 19.182 ug/l
RT: 3.47 min Scan# 288
Delta R.T. 0.01 min
Lab File: VF061415.D
Acq: 16 Jan 2019 16:49

Instrument :
MSVOA_F
ClientSampleId :
VF0116SBS01




Tot Ion	Ratio	Lower	Upper
96	100		
61	140.3	0.0	270.2
98	64.8	0.0	124.0




#28

Bromochloromethane
Concen: 18.442 ug/l
RT: 3.71 min Scan# 312
Delta R.T. 0.00 min
Lab File: VF061415.D
Acq: 16 Jan 2019 16:49

Tgt Ion	Ratio	Lower	Upper
49	100		
129	2.3	0.0	0.0#
130	79.5	65.4	98.2





#29

Tetrahydrofuran

Concen: 100.819 ug/l

RT: 4.07 min Scan# 349

Delta R.T. 0.01 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :

VF0116SBS01


Tot Ion: 42 Resp: 38662

Ion Ratio Lower Upper

42 100

72 39.3 28.9 43.3




71 37.8 29.6 44.4



Abundance Ion 42.00 (41.70 to 42.70): VF0

Ion 72.00 (71.70 to 72.70): VF0

Ion 71.00 (70.70 to 71.70): VF0



#30

Chloroform

Concen: 18.970 ug/l

RT: 3.86 min Scan# 327

Delta R.T. 0.01 min

Lab File: VF061415.D


Acq: 16 Jan 2019 16:49

Tgt Ion: 83 Resp: 109961

Ion Ratio Lower Upper



83 100


85 62.7 52.7 79.1



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 85.00 (84.70 to 85.70): VF0





#31

Cyclohexane

Concen: 21.569 ug/l

RT: 3.69 min Scan# 310

Delta R.T. 0.01 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :

VF0116SBS01


Tot Ion: 56 Resp: 92541

Ion Ratio Lower Upper




56 100

69 31.1 24.6 36.8

84 82.3 74.2 111.2



Abundance Ion 56.00 (55.70 to 56.70): VF0
Ion 69.00 (68.70 to 69.70): VF0
Ion 84.00 (83.70 to 84.70): VF0



#32

1,1,1-Trichloroethane

Concen: 19.660 ug/l

RT: 4.08 min Scan# 350

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49


Tgt Ion: 97 Resp: 83560

Ion Ratio Lower Upper



97 100


99 57.6 50.1 75.1

61 47.6 43.1 64.7



Abundance Ion 97.00 (96.70 to 97.70): VF0
Ion 99.00 (98.70 to 99.70): VF0
Ion 61.00 (60.70 to 61.70): VF0





#33

1,2-Dichloroethane-d4

Concen: 44.336 ug/l

RT: 4.83 min Scan# 426

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :


VF0116SBS01

Tot Ion: 65 Resp: 118356

Ion Ratio Lower Upper




65 100

67 50.9 0.0 99.8



Abundance Ion 65.00 (64.70 to 65.70): VF0

Ion 67.00 (66.70 to 67.70): VF0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 5.58 min Scan# 502

Delta R.T. -0.01 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49


Tgt Ion: 114 Resp: 370515

Ion Ratio Lower Upper

114 100

63 19.8 0.0 39.0



88 19.0 0.0 37.4




Abundance Ion 114.00 (113.70 to 114.70): V

Ion 63.00 (62.70 to 63.70): VF0

Ion 88.00 (87.70 to 88.70): VF0





#35

Dibromofluoromethane

Concen: 48.302 ug/l

RT: 4.10 min Scan# 352

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :

VF0116SBS01


Tot Ion:113 Resp: 141219

Ion Ratio Lower Upper

113 100

111 98.0 79.4 119.0




192 15.8 12.4 18.6



Abundance Ion 113.00 (112.70 to 113.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 192.00 (191.70 to 192.70): V



#36

1,1-Dichloropropene

Concen: 19.926 ug/l

RT: 4.27 min Scan# 369

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49


Tgt Ion: 75 Resp: 85600

Ion Ratio Lower Upper

75 100

110 32.8 15.8 47.3



77 31.0 24.0 36.0




Abundance Ion 75.00 (74.70 to 75.70): VF0

Ion 110.00 (109.70 to 110.70): V

Ion 77.00 (76.70 to 77.70): VF0





#37

Ethyl Acetate

Concen: 20.447 ug/l

RT: 4.10 min Scan# 352

Delta R.T. 0.01 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :

VF0116SBS01





Tot Ion: 43 Resp: 35112

Ion Ratio Lower Upper

43 100

61 74.5 81.7 122.5#

70 9.7 6.2 9.2#



#38

Carbon Tetrachloride

Concen: 18.933 ug/l

RT: 3.99 min Scan# 340

Delta R.T. 0.01 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49




Tgt Ion: 117 Resp: 71106

Ion Ratio Lower Upper

117 100

119 106.3 79.0 118.6

121 39.3 27.0 40.6









#39
Methvlcvclohexane
Concen: 20.231 ug/l
RT: 5.45 min Scan# 488
Delta R.T. 0.00 min
Lab File: VF061415.D
Acq: 16 Jan 2019 16:49

Instrument :
MSVOA_F
ClientSampleId :
VF0116SBS01

Tot Ion: 83 Resp: 94187
Ion Ratio Lower Upper
83 100
55 92.3 67.4 101.2
98 50.4 36.7 55.1




Abundance Ion 83.00 (82.70 to 83.70): VF0
Ion 55.00 (54.70 to 55.70): VF0
Ion 98.00 (97.70 to 98.70): VF0






#40
Benzene
Concen: 21.561 ug/l
RT: 4.62 min Scan# 404
Delta R.T. -0.01 min
Lab File: VF061415.D
Acq: 16 Jan 2019 16:49

Tgt Ion: 78 Resp: 206106
Ion Ratio Lower Upper
78 100
77 25.1 19.0 28.6



Abundance Ion 78.10 (77.80 to 78.80): VF0
Ion 77.00 (76.70 to 77.70): VF0





#41

Methacrylonitrile

Concen: 20.663 ug/l

RT: 4.74 min Scan# 416

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :

VF0116SBS01

Tot Ion: 41 Resp: 18881


Ion Ratio Lower Upper

41 100

39 72.7 43.1 64.7#

67 50.8 38.8 58.2

52 50.0 33.2 49.8#






Abundance Ion 41.00 (40.70 to 41.70): VF0

Ion 39.00 (38.70 to 39.70): VF0

Ion 67.00 (66.70 to 67.70): VF0

Ion 52.00 (51.70 to 52.70): VF0



#42

1,2-Dichloroethane

Concen: 17.283 ug/l

RT: 4.93 min Scan# 436

Delta R.T. 0.00 min

Lab File: VF061415.D


Acq: 16 Jan 2019 16:49

Tgt Ion: 62 Resp: 58872

Ion Ratio Lower Upper



62 100


98 7.1 0.0 17.0



Abundance Ion 62.00 (61.70 to 62.70): VF0

Ion 98.00 (97.70 to 98.70): VF0









#43

Isopropyl Acetate
 Concen: 19.229 ug/l
 RT: 6.95 min Scan# 641
 Delta R.T. 0.00 min
 Lab File: VF061415.D
 Acq: 16 Jan 2019 16:49

Instrument :
 MSVOA_F
 ClientSampleId :
 VF0116SBS01




Tot Ion: 43 Resp: 41443
 Ion Ratio Lower Upper
 43 100
 61 29.3 20.3 30.5
 87 0.7 0.5 0.7




#44

Trichloroethene
 Concen: 20.280 ug/l
 RT: 5.49 min Scan# 493
 Delta R.T. 0.00 min
 Lab File: VF061415.D
 Acq: 16 Jan 2019 16:49

Tgt Ion: 130 Resp: 59469
 Ion Ratio Lower Upper
 130 100
 95 93.8 0.0 202.8





#45

1,2-Dichloropropane

Concen: 20.985 ug/l

RT: 6.22 min Scan# 567

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :





VF0116SBS01

Tot Ion: 63 Resp: 52648

Ion Ratio Lower Upper

63 100

65 32.1 24.4 36.6



#46

Dibromomethane

Concen: 18.806 ug/l

RT: 6.08 min Scan# 552

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49




Tgt Ion: 93 Resp: 31373

Ion Ratio Lower Upper

93 100

95 76.0 62.1 93.1

174 81.1 74.6 111.8





#47

Bromodichloromethane

Concen: 18.592 ug/l

RT: 6.38 min Scan# 583

Delta R.T. 0.01 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :

VF0116SBS01


Tot Ion: 83 Resp: 76739

Ion Ratio Lower Upper

83 100

85 61.2 53.0 79.4




127 8.3 6.9 10.3



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 85.00 (84.70 to 85.70): VF0

Ion 127.00 (126.70 to 127.70): V



#48

Methyl methacrylate

Concen: 19.925 ug/l

RT: 6.71 min Scan# 616

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49


Tgt Ion: 41 Resp: 28836

Ion Ratio Lower Upper

41 100

69 78.2 51.7 77.5#



39 57.9 49.2 73.8




Abundance Ion 41.00 (40.70 to 41.70): VF0

Ion 69.00 (68.70 to 69.70): VF0

Ion 39.00 (38.70 to 39.70): VF0









#49
1,4-Dioxane
Concen: 404.927 ug/l
RT: 6.71 min Scan# 616
Delta R.T. 0.01 min
Lab File: VF061415.D
Acq: 16 Jan 2019 16:49




Instrument :
MSVOA_F
ClientSampleId :
VF0116SBS01


Tot Ion: 88 Resp: 4508
Ion Ratio Lower Upper
88 100
43 63.9 55.5 83.3
58 57.6 59.9 89.9#



#50
Toluene-d8
Concen: 49.457 ug/l
RT: 7.55 min Scan# 701
Delta R.T. 0.00 min
Lab File: VF061415.D
Acq: 16 Jan 2019 16:49

Tgt Ion: 98 Resp: 396167
Ion Ratio Lower Upper
98 100
100 71.0 55.6 83.4





#51

4-Methyl-2-Pentanone

Concen: 104.381 ug/l

RT: 8.25 min Scan# 772

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :


VF0116SBS01

Tot Ion: 43 Resp: 165385

Ion Ratio Lower Upper




43 100

58 34.7 28.8 43.2



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0



#52

Toluene

Concen: 19.267 ug/l

RT: 7.62 min Scan# 708

Delta R.T. 0.00 min

Lab File: VF061415.D


Acq: 16 Jan 2019 16:49

Tgt Ion: 92 Resp: 121290

Ion Ratio Lower Upper



92 100


91 172.6 131.2 196.8



Abundance Ion 92.00 (91.70 to 92.70): VF0

Ion 91.00 (90.70 to 91.70): VF0





#53

t-1,3-Dichloropropene

Concen: 20.694 ug/l

RT: 8.27 min Scan# 774

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :





VF0116SBS01

Tot Ion: 75 Resp: 67824

Ion Ratio Lower Upper

75 100

77 32.2 24.4 36.6



#54

cis-1,3-Dichloropropene

Concen: 21.326 ug/l

RT: 7.29 min Scan# 675

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49




Tgt Ion: 75 Resp: 87368


Ion Ratio Lower Upper

75 100

77 27.9 25.4 38.0

39 54.5 47.8 71.8





#55

1,1,2-Trichloroethane

Concen: 20.335 ug/l

RT: 8.48 min Scan# 796

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :

VF0116SBS01

Tot Ion: 97 Resp: 36352


Ion Ratio Lower Upper

97 100

83 93.0 70.4 105.6

85 64.3 44.7 67.1

99 72.7 50.8 76.2






Abundance Ion 97.00 (96.70 to 97.70): VF0

Ion 82.95 (82.65 to 83.65): VF0

Ion 84.95 (84.65 to 85.65): VF0

Ion 98.95 (98.65 to 99.65): VF0



#56

Ethyl methacrylate

Concen: 19.916 ug/l

RT: 8.61 min Scan# 809

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49


Tgt Ion: 69 Resp: 41421

Ion Ratio Lower Upper

69 100

41 80.9 57.5 86.3



39 43.9 32.0 48.0




Abundance Ion 69.00 (68.70 to 69.70): VF0

Ion 41.00 (40.70 to 41.70): VF0

Ion 39.00 (38.70 to 39.70): VF0









#57
 1,3-Dichloropropane
 Concen: 19.050 ug/l
 RT: 8.85 min Scan# 833
 Delta R.T. 0.00 min
 Lab File: VF061415.D
 Acq: 16 Jan 2019 16:49




Instrument :
 MSVOA_F
 ClientSampleId :
 VF0116SBS01


Tot Ion: 76 Resp: 63829
 Ion Ratio Lower Upper
 76 100
 78 33.0 25.3 37.9



#58
 2-Chloroethyl Vinyl ether
 Concen: 115.409 ug/l
 RT: 7.29 min Scan# 675
 Delta R.T. 0.00 min
 Lab File: VF061415.D
 Acq: 16 Jan 2019 16:49

Tgt Ion: 63 Resp: 21539
 Ion Ratio Lower Upper
 63 100
 106 32.4 21.9 32.9









#59
2-Hexanone
Concen: 114.469 ug/l
RT: 9.49 min Scan# 898
Delta R.T. 0.01 min
Lab File: VF061415.D
Acq: 16 Jan 2019 16:49




Instrument :
MSVOA_F
ClientSampleId :
VF0116SBS01


Tot Ion: 43 Resp: 126472
Ion Ratio Lower Upper
43 100
58 52.7 25.9 77.6



#60
Dibromochloromethane
Concen: 19.193 ug/l
RT: 8.71 min Scan# 819
Delta R.T. 0.00 min
Lab File: VF061415.D
Acq: 16 Jan 2019 16:49

Tgt Ion: 129 Resp: 50158
Ion Ratio Lower Upper
129 100
127 77.9 39.1 117.2





#61

1,2-Dibromoethane

Concen: 18.547 ug/l

RT: 8.98 min Scan# 846

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :


VF0116SBS01

Tot Ion:107 Resp: 37956

Ion Ratio Lower Upper




107 100

109 103.1 70.8 106.2



Abundance Ion 107.00 (106.70 to 107.70): V

Ion 109.00 (108.70 to 109.70): V



#62

4-Bromofluorobenzene

Concen: 44.919 ug/l

RT: 11.40 min Scan# 1092

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49


Tgt Ion: 95 Resp: 168116

Ion Ratio Lower Upper

95 100

174 68.9 0.0 130.4



176 68.3 0.0 126.2



Abundance Ion 95.00 (94.70 to 95.70): VF0

Ion 174.00 (173.70 to 174.70): V

Ion 176.00 (175.70 to 176.70): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 9.77 min Scan# 926
Delta R.T. 0.00 min
Lab File: VF061415.D
Acq: 16 Jan 2019 16:49




Instrument :
MSVOA_F
ClientSampleId :
VF0116SBS01


Tot Ion:117 Resp: 327085
Ion Ratio Lower Upper
117 100
82 57.4 46.4 69.6
119 31.0 27.2 40.8



#64
Tetrachloroethene
Concen: 20.120 ug/l
RT: 8.15 min Scan# 762
Delta R.T. 0.00 min
Lab File: VF061415.D
Acq: 16 Jan 2019 16:49

Tgt Ion:164 Resp: 49489
Ion Ratio Lower Upper
164 100
166 122.6 99.6 149.4
129 89.5 76.4 114.6
131 92.9 77.9 116.9









#65
Chlorobenzene
Concen: 20.562 ug/l
RT: 9.79 min Scan# 928
Delta R.T. 0.00 min
Lab File: VF061415.D
Acq: 16 Jan 2019 16:49




Instrument :
MSVOA_F
ClientSampleId :
VF0116SBS01


Tot Ion:112 Resp: 136770
Ion Ratio Lower Upper
112 100
114 31.4 25.4 38.0



#66
1,1,1,2-Tetrachloroethane
Concen: 21.165 ug/l
RT: 9.92 min Scan# 942
Delta R.T. 0.00 min
Lab File: VF061415.D
Acq: 16 Jan 2019 16:49

Tgt Ion:131 Resp: 56082
Ion Ratio Lower Upper
131 100
133 98.4 47.4 142.2
119 65.8 33.1 99.3





#67

Ethyl Benzene

Concen: 21.369 ug/l

RT: 9.89 min Scan# 939

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :





VF0116SBS01

Tot Ion: 91 Resp: 259691

Ion Ratio Lower Upper

91 100

106 30.0 23.1 34.7



#68

m/p-Xylenes

Concen: 42.577 ug/l

RT: 10.12 min Scan# 962

Delta R.T. 0.00 min

Lab File: VF061415.D




Acq: 16 Jan 2019 16:49


Tgt Ion: 106 Resp: 186102

Ion Ratio Lower Upper

106 100

91 213.1 166.2 249.4







#69
o-Xylene
Concen: 20.811 ug/l
RT: 10.70 min Scan# 1021
Delta R.T. 0.00 min
Lab File: VF061415.D
Acq: 16 Jan 2019 16:49

Instrument :
MSVOA_F
ClientSampleId :
VF0116SBS01



Tot Ion:106 Resp: 98588
Ion Ratio Lower Upper
106 100
91 232.2 113.3 339.8



Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): VF0




Time-->




#70
Styrene
Concen: 21.283 ug/l
RT: 10.78 min Scan# 1029
Delta R.T. 0.00 min
Lab File: VF061415.D
Acq: 16 Jan 2019 16:49


Tgt Ion:104 Resp: 138008
Ion Ratio Lower Upper
104 100
78 46.7 38.8 58.2
103 44.5 37.0 55.6




Abundance Ion 104.00 (103.70 to 104.70): V
Ion 78.00 (77.70 to 78.70): VF0
Ion 103.00 (102.70 to 103.70): V



Time-->





#71

Bromoform

Concen: 21.189 ug/l

RT: 10.76 min Scan# 1027

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :

VF0116SBS01


Tot Ion:173 Resp: 25227

Ion Ratio Lower Upper

173 100

175 51.0 24.1 72.4




254 0.0 0.0 0.0



Abundance Ion 173.00 (172.70 to 173.70): V

Ion 175.00 (174.70 to 175.70): V

Ion 254.00 (253.70 to 254.70): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.55 min Scan# 1208

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49


Tgt Ion:152 Resp: 167071

Ion Ratio Lower Upper

152 100

115 57.3 30.0 90.0



150 157.4 0.0 321.4




Abundance Ion 152.00 (151.70 to 152.70): V

Ion 115.00 (114.70 to 115.70): V

Ion 150.00 (149.70 to 150.70): V





#73

Isopropylbenzene

Concen: 22.251 ug/l

RT: 11.12 min Scan# 1063

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :


VF0116SBS01

Tot Ion:105 Resp: 302352

Ion Ratio Lower Upper

105 100

120 25.4 12.6 37.8



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V

11.12

200000

150000



100000

50000

0

Time-->

11.00 11.10 11.20 11.30



#74

N-ethyl acetate

Concen: 23.000 ug/l

RT: 11.36 min Scan# 1088

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Tgt Ion: 43 Resp: 73902


Ion Ratio Lower Upper

43 100

70 31.6 27.6 41.4

55 19.8 15.4 23.2

61 23.5 21.8 32.6



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 70.00 (69.70 to 70.70): VF0

Ion 55.00 (54.70 to 55.70): VF0

Ion 61.00 (60.70 to 61.70): VF0

11.36

60000


40000


20000

0

Time-->

11.30 11.40 11.50





#75

1,1,2,2-Tetrachloroethane

Concen: 21.360 ug/l

RT: 11.69 min Scan# 1121

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :

VF0116SBS01


Tot Ion: 83 Resp: 51707

Ion Ratio Lower Upper

83 100

131 8.3 4.0 12.2




85 66.8 32.7 98.1



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 131.00 (130.70 to 131.70): VF0

Ion 85.00 (84.70 to 85.70): VF0



#76

1,2,3-Trichloropropane

Concen: 21.088 ug/l

RT: 11.79 min Scan# 1131

Delta R.T. 0.00 min

Lab File: VF061415.D


Acq: 16 Jan 2019 16:49

Tgt Ion: 75 Resp: 36681

Ion Ratio Lower Upper



75 100


77 32.1 16.1 48.3



Abundance Ion 75.00 (74.70 to 75.70): VF0

Ion 77.00 (76.70 to 77.70): VF0





#77

Bromobenzene

Concen: 21.471 ug/l

RT: 11.49 min Scan# 1101

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :

VF0116SBS01


Tot Ion:156 Resp: 61614

Ion Ratio Lower Upper

156 100

77 160.0 78.3 234.9




158 94.7 48.9 146.7



Abundance Ion 156.00 (155.70 to 156.70): V

Ion 77.00 (76.70 to 77.70): VF0

Ion 158.00 (157.70 to 158.70): V



#78

n-propylbenzene

Concen: 19.617 ug/l

RT: 11.59 min Scan# 1111

Delta R.T. 0.00 min

Lab File: VF061415.D


Acq: 16 Jan 2019 16:49

Tgt Ion: 91 Resp: 347575

Ion Ratio Lower Upper



91 100


120 21.0 10.9 32.6



Abundance Ion 91.00 (90.70 to 91.70): VF0

Ion 120.00 (119.70 to 120.70): V





#79

2-Chlorotoluene

Concen: 21.460 ug/l

RT: 11.71 min Scan# 1123

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :





VF0116SBS01

Tot Ion: 91 Resp: 202324

Ion Ratio Lower Upper

91 100

126 30.6 15.0 45.1



#80

1,3,5-Trimethylbenzene

Concen: 22.054 ug/l

RT: 11.82 min Scan# 1134

Delta R.T. 0.00 min

Lab File: VF061415.D




Acq: 16 Jan 2019 16:49


Tgt Ion:105 Resp: 241842

Ion Ratio Lower Upper

105 100

120 49.0 24.7 74.1





#81

trans-1,4-Dichloro-2-butene

Concen: 30.635 ug/l

RT: 11.87 min Scan# 1139

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :

VF0116SBS01





Tot Ion: 75 Resp: 23164

Ion Ratio Lower Upper

75 100

53 79.3 57.8 86.6

89 50.8 40.7 61.1



#82

4-Chlorotoluene

Concen: 21.455 ug/l

RT: 11.90 min Scan# 1142

Delta R.T. 0.00 min

Lab File: VF061415.D




Acq: 16 Jan 2019 16:49


Tgt Ion: 91 Resp: 204118

Ion Ratio Lower Upper

91 100

126 33.6 16.0 47.9





#83

tert-Butylbenzene

Concen: 20.921 ug/l

RT: 12.12 min Scan# 1165

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :

VF0116SBS01


Tot Ion:119 Resp: 237450

Ion Ratio Lower Upper

119 100

91 62.1 31.3 93.8




134 24.0 12.4 37.0



Abundance Ion 119.00 (118.70 to 119.70): V

Ion 91.00 (90.70 to 91.70): VF0

Ion 134.00 (133.70 to 134.70): V



#84

1,2,4-Trimethylbenzene

Concen: 21.247 ug/l

RT: 12.20 min Scan# 1173

Delta R.T. 0.00 min

Lab File: VF061415.D


Acq: 16 Jan 2019 16:49

Tgt Ion:105 Resp: 239997

Ion Ratio Lower Upper



105 100

120 46.9 24.9 74.6



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V









#85
 sec-Butylbenzene
 Concen: 22.069 ug/l
 RT: 12.30 min Scan# 1183
 Delta R.T. 0.00 min
 Lab File: VF061415.D
 Acq: 16 Jan 2019 16:49




Instrument :
 MSVOA_F
 ClientSampleId :
 VF0116SBS01


Tot Ion:105 Resp: 340159
 Ion Ratio Lower Upper
 105 100
 134 17.8 10.0 30.0



#86
 p-Isopropyltoluene
 Concen: 22.201 ug/l
 RT: 12.46 min Scan# 1199
 Delta R.T. 0.00 min
 Lab File: VF061415.D
 Acq: 16 Jan 2019 16:49

Tgt Ion:119 Resp: 280908
 Ion Ratio Lower Upper
 119 100
 134 25.1 13.4 40.2
 91 22.6 11.4 34.2





#87

1,3-Dichlorobenzene

Concen: 19.995 ug/l

RT: 12.47 min Scan# 1200

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :

VF0116SBS01


Tot Ion:146 Resp: 120563

Ion Ratio Lower Upper

146 100

111 42.2 21.1 63.4




148 63.5 32.9 98.6



Abundance Ion 146.00 (145.70 to 146.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 148.00 (147.70 to 148.70): V



#88

1,4-Dichlorobenzene

Concen: 21.013 ug/l

RT: 12.56 min Scan# 1209

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49


Tgt Ion:146 Resp: 113947

Ion Ratio Lower Upper

146 100

111 39.7 20.8 62.2



148 67.1 32.1 96.5




Abundance Ion 146.00 (145.70 to 146.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 148.00 (147.70 to 148.70): V





#89

n-Butylbenzene

Concen: 22.003 ug/l

RT: 12.84 min Scan# 1238

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :

VF0116SBS01


Tot Ion: 91 Resp: 287112

Ion Ratio Lower Upper

91 100

92 60.9 31.6 94.7




134 25.6 13.6 40.8



Abundance Ion 91.00 (90.70 to 91.70): VF0

Ion 92.00 (91.70 to 92.70): VF0

Ion 134.00 (133.70 to 134.70): V



#90

Hexachloroethane

Concen: 21.184 ug/l

RT: 12.91 min Scan# 1245

Delta R.T. 0.00 min

Lab File: VF061415.D


Acq: 16 Jan 2019 16:49

Tgt Ion: 117 Resp: 63183

Ion Ratio Lower Upper



117 100


201 56.1 29.3 87.8



Abundance Ion 117.00 (116.70 to 117.70): V

Ion 201.00 (200.70 to 201.70): V





#91

1,2-Dichlorobenzene

Concen: 20.826 ug/l

RT: 12.94 min Scan# 1248

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :

MSVOA_F

ClientSampleId :

VF0116SBS01


Tot Ion:146 Resp: 111798

Ion Ratio Lower Upper

146 100

111 39.7 21.1 63.3




148 65.8 31.9 95.7



Abundance Ion 146.00 (145.70 to 146.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 148.00 (147.70 to 148.70): V



#92

1,2-Dibromo-3-Chloropropane

Concen: 19.308 ug/l

RT: 13.63 min Scan# 1318

Delta R.T. -0.01 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49


Tgt Ion: 75 Resp: 8290

Ion Ratio Lower Upper

75 100

155 67.6 49.7 149.1



157 78.3 61.4 184.1



Abundance Ion 75.00 (74.70 to 75.70): VF0

Ion 155.00 (154.70 to 155.70): V

Ion 157.00 (156.70 to 157.70): V









#93
 1,2,4-Trichlorobenzene
 Concen: 20.822 ug/l
 RT: 14.20 min Scan# 1375
 Delta R.T. 0.00 min
 Lab File: VF061415.D
 Acq: 16 Jan 2019 16:49

Instrument :
 MSVOA_F
 ClientSampleId :
 VF0116SBS01

Tot Ion:180 Resp: 78303
 Ion Ratio Lower Upper
 180 100
 182 94.6 48.5 145.6
 145 33.1 17.5 52.6




Abundance Ion 180.00 (179.70 to 180.70): V
 Ion 182.00 (181.70 to 182.70): V
 Ion 145.00 (144.70 to 145.70): V






#94
 Hexachlorobutadiene
 Concen: 20.255 ug/l
 RT: 14.19 min Scan# 1374
 Delta R.T. 0.00 min
 Lab File: VF061415.D
 Acq: 16 Jan 2019 16:49

Tgt Ion:225 Resp: 47266
 Ion Ratio Lower Upper
 225 100
 223 61.9 31.1 93.2
 227 67.1 32.2 96.6



Abundance Ion 225.00 (224.70 to 225.70): V
 Ion 223.00 (222.70 to 223.70): V
 Ion 227.00 (226.70 to 227.70): V





#95

Naphthalene

Concen: 18.707 ug/l

RT: 14.45 min Scan# 1401

Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Instrument :


MSVOA_F

ClientSampleId :

VF0116SBS01

Tot Ion:128 Resp: 133660

Ion	Ratio	Lower	Upper
128	100		
127	13.6	9.5	14.3
129	10.8	8.4	12.6






Abundance

Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V



#96

1,2,3-Trichlorobenzene

Concen: 19.757 ug/l

RT: 14.60 min Scan# 1416


Delta R.T. 0.00 min

Lab File: VF061415.D

Acq: 16 Jan 2019 16:49

Tgt Ion:180 Resp: 68780

Ion	Ratio	Lower	Upper
180	100		
182	94.2	47.0	141.2
145	32.2	17.1	51.2



Abundance

Ion 180.00 (179.70 to 180.70): V

Ion 182.00 (181.70 to 182.70): V

Ion 145.00 (144.70 to 145.70): V

